



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Ormond Beach Police Department Instrument Serial Number: n/a

Department Inspection: Agency Inspection/Agency Contact: Records Audit: Registration, Agency Inspection Notification, The following records were audited: Agency Inspection Reports, Breath Test Instrument Repair/Maintenance Records, Instrument Registration, Dry Gas Standard Certificate(s) of Analysis. Date of Notification/Contact: 8/31/2021. Agency Inspector: Michael Garner / Rob Pearson.

Instrument/Area: Equipment: Supplies: Clean/Dry, Secure, Limited Access, Proper Number of Simulators, Checked Simulators for Air Leak Resistant Seal and Proper Temperature, Class A Glassware, Distilled/Deionized Water, Mouth Alcohol Solution, Acetone Stock Solution, Alcohol Reference Solution, Dry Gas Standard, Mouth Pieces.

Action: Instrument Complies with Chapter 11D-8, FAC, Instrument Does Not Comply with Chapter 11D-8, FAC, Remain in Evidentiary Use, Return to/Place into Evidentiary Use, Remove from Evidentiary Use, Remain Out of Evidentiary Use, Conduct an Agency Inspection Before Evidentiary Use. Comments: Recommended clearly marking expired ARS/DGS to avoid accidental use. Recommended purchasing more DGS soon (~250 psi remaining) on final cylinder. Compliant with Chapter 11D-8, FAC.

Taylor Gutschow

Digitally signed by Taylor Gutschow Date: 2021.09.03 09:46:13 -04'00'

9/3/2021

Signature of Alcohol Testing Program Staff Member

Date

Original - FDLE

Copy - Agency