

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.  
Time of Inspection: 14:11

Date of Inspection: 12/17/2022

Serial Number: 80-000812  
Software: 8100.27

| Check or Test                               | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202112A<br>Exp: 12/07/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202108D<br>Exp: 08/17/2023 | 0.20g/210L Test (g/210L)<br>Lot#:202206C<br>Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:15922080A5<br>Exp: 08/05/2024 |
|----------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-----------------------------------------------------------------------------|
| 0.000                      | 0.049                                                       | 0.078                                                       | 0.198                                                       | 0.082                                                                       |
| 0.000                      | 0.049                                                       | 0.078                                                       | 0.198                                                       | 0.082                                                                       |
| 0.000                      | 0.049                                                       | 0.079                                                       | 0.198                                                       | 0.082                                                                       |

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BC 1193

BRAD R COLLITO

Signature and Printed Name

12/17/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO  
Time of Inspection: 10:22

Date of Inspection: 12/18/2022

Serial Number: 80-005249  
Software: 8100.27

| Check or Test                               | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202112A<br>Exp: 12/07/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202108D<br>Exp: 08/17/2023 | 0.20g/210L Test (g/210L)<br>Lot#:202206C<br>Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:15922080A5<br>Exp: 08/05/2024 |
|----------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-----------------------------------------------------------------------------|
| 0.000                      | 0.047                                                       | 0.078                                                       | 0.196                                                       | 0.081                                                                       |
| 0.000                      | 0.048                                                       | 0.079                                                       | 0.199                                                       | 0.081                                                                       |
| 0.000                      | 0.047                                                       | 0.078                                                       | 0.198                                                       | 0.081                                                                       |

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

B. COLLITO

BRAD R COLLITO

Signature and Printed Name

12/18/2022  
Date

# Florida Department of Law Enforcement

## Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO  
Time of Inspection: 14:50

Date of Inspection: 12/17/2022

Serial Number: 80-005249  
Software: 8100.27

| Check or Test                               | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      |     | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202112A<br>Exp: 12/07/2023 | 0.08g/210L Test (g/210L)<br>Lot#:<br>Exp: | 0.20g/210L Test (g/210L)<br>Lot#:<br>Exp: | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:<br>Exp: |
|----------------------------|-------------------------------------------------------------|-------------------------------------------|-------------------------------------------|--------------------------------------------------------|
| 0.000 / 0.000              | 0.048 / 0.048                                               |                                           |                                           |                                                        |
| INT / 0.000                | 0.049 / 0.048                                               |                                           |                                           |                                                        |
| / 0.000                    | INT / INT                                                   |                                           |                                           |                                                        |

Number of Simulators Used: 5

**Remarks:**

00: Interferent Detect. 05: Interferent Detect, Interferent Detect. Non-compliance: FAILED INT TWO TI MES.

*Simulator w/acetone removed from service. POSSIBLE LEAK. WAITED OVERNIGHT, UTILIZED DIFFERENT SIMULATOR. CONDUCTED 10 STABILITY CHECKS PER SIMULATOR WITH NO DISCREPANCIES. BC1193 12/18/2022*

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*BC1193*

Signature and Printed Name

BRAD R COLLITO

12/17/2022  
Date

HERNANDO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005249  
12/18/2022  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 07:41 |
| Control Test       | 0.000  | 07:42 |
| Air Blank          | 0.000  | 07:42 |
| Control Test       | 0.000  | 07:43 |
| Air Blank          | 0.000  | 07:43 |
| Control Test       | 0.000  | 07:44 |
| Air Blank          | 0.000  | 07:45 |
| Control Test       | 0.000  | 07:45 |
| Air Blank          | 0.000  | 07:46 |
| Control Test       | 0.000  | 07:46 |
| Air Blank          | 0.000  | 07:47 |
| Control Test       | 0.000  | 07:48 |
| Air Blank          | 0.000  | 07:48 |
| Control Test       | 0.000  | 07:49 |
| Air Blank          | 0.000  | 07:49 |
| Control Test       | 0.000  | 07:50 |
| Air Blank          | 0.000  | 07:50 |
| Control Test       | 0.000  | 07:51 |
| Air Blank          | 0.000  | 07:52 |
| Control Test       | 0.000  | 07:52 |
| Air Blank          | 0.000  | 07:53 |
| Control Test Stats |        |       |
| Average            | 0.0000 |       |
| Std Dev            | 0.0000 |       |
| Rel Std Dev(%)     | 0.0000 |       |

B C 1193

---

Operator's Signature

HERNANDO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005249  
12/18/2022  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 07:54 |
| Control Test       | INT*   | 07:55 |
| Air Blank          | 0.000  | 07:55 |
| Control Test       | INT*   | 07:56 |
| Air Blank          | 0.000  | 07:57 |
| Control Test       | INT*   | 07:57 |
| Air Blank          | 0.000  | 07:58 |
| Control Test       | INT*   | 07:58 |
| Air Blank          | 0.000  | 07:59 |
| Control Test       | INT*   | 08:00 |
| Air Blank          | 0.000  | 08:00 |
| Control Test       | INT*   | 08:01 |
| Air Blank          | 0.000  | 08:01 |
| Control Test       | INT*   | 08:02 |
| Air Blank          | 0.000  | 08:03 |
| Control Test       | INT*   | 08:03 |
| Air Blank          | 0.000  | 08:04 |
| Control Test       | INT*   | 08:04 |
| Air Blank          | 0.000  | 08:05 |
| Control Test       | INT*   | 08:06 |
| Air Blank          | 0.000  | 08:06 |
| Control Test Stats |        |       |
| Average            | 0.0000 |       |
| Std Dev            | 0.0000 |       |
| Rel Std Dev(%)     | 0.0000 |       |

\*Interferent Detect

BC-1193

Operator's Signature

HERNANDO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005249  
12/18/2022  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 08:09 |
| Control Test       | 0.048  | 08:10 |
| Air Blank          | 0.000  | 08:10 |
| Control Test       | 0.047  | 08:11 |
| Air Blank          | 0.000  | 08:12 |
| Control Test       | 0.047  | 08:12 |
| Air Blank          | 0.000  | 08:13 |
| Control Test       | 0.047  | 08:13 |
| Air Blank          | 0.000  | 08:14 |
| Control Test       | 0.046  | 08:15 |
| Air Blank          | 0.000  | 08:15 |
| Control Test       | 0.047  | 08:16 |
| Air Blank          | 0.000  | 08:16 |
| Control Test       | 0.047  | 08:17 |
| Air Blank          | 0.000  | 08:18 |
| Control Test       | 0.047  | 08:18 |
| Air Blank          | 0.000  | 08:19 |
| Control Test       | 0.047  | 08:19 |
| Air Blank          | 0.000  | 08:20 |
| Control Test       | 0.048  | 08:21 |
| Air Blank          | 0.000  | 08:21 |
| Control Test Stats |        |       |
| Average            | 0.0471 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 1.2052 |       |

BC1193

Operator's Signature

HERNANDO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005249  
12/18/2022  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 08:27 |
| Control Test       | 0.076  | 08:27 |
| Air Blank          | 0.000  | 08:28 |
| Control Test       | 0.077  | 08:29 |
| Air Blank          | 0.000  | 08:29 |
| Control Test       | 0.077  | 08:30 |
| Air Blank          | 0.000  | 08:30 |
| Control Test       | 0.078  | 08:31 |
| Air Blank          | 0.000  | 08:32 |
| Control Test       | 0.078  | 08:32 |
| Air Blank          | 0.000  | 08:33 |
| Control Test       | 0.077  | 08:33 |
| Air Blank          | 0.000  | 08:34 |
| Control Test       | 0.077  | 08:35 |
| Air Blank          | 0.000  | 08:35 |
| Control Test       | 0.077  | 08:36 |
| Air Blank          | 0.000  | 08:36 |
| Control Test       | 0.077  | 08:37 |
| Air Blank          | 0.000  | 08:38 |
| Control Test       | 0.076  | 08:38 |
| Air Blank          | 0.000  | 08:39 |
| Control Test Stats |        |       |
| Average            | 0.0770 |       |
| Std Dev            | 0.0007 |       |
| Rel Std Dev(%)     | 0.8658 |       |

BC-1193

Operator's Signature

HERNANDO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005249  
12/18/2022  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 09:14 |
| Control Test       | 0.196  | 09:14 |
| Air Blank          | 0.000  | 09:15 |
| Control Test       | 0.196  | 09:16 |
| Air Blank          | 0.000  | 09:16 |
| Control Test       | 0.196  | 09:17 |
| Air Blank          | 0.000  | 09:17 |
| Control Test       | 0.195  | 09:18 |
| Air Blank          | 0.000  | 09:18 |
| Control Test       | 0.194  | 09:19 |
| Air Blank          | 0.000  | 09:20 |
| Control Test       | 0.194  | 09:20 |
| Air Blank          | 0.000  | 09:21 |
| Control Test       | 0.194  | 09:22 |
| Air Blank          | 0.000  | 09:22 |
| Control Test       | 0.194  | 09:23 |
| Air Blank          | 0.000  | 09:23 |
| Control Test       | 0.193  | 09:24 |
| Air Blank          | 0.000  | 09:25 |
| Control Test       | 0.193  | 09:25 |
| Air Blank          | 0.000  | 09:26 |
| Control Test Stats |        |       |
| Average            | 0.1945 |       |
| Std Dev            | 0.0012 |       |
| Rel Std Dev(%)     | 0.6059 |       |

BC 1193

Operator's Signature

HERNANDO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005249  
12/18/2022  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 09:27 |
| Control Test       | 0.080  | 09:28 |
| Air Blank          | 0.000  | 09:28 |
| Control Test       | 0.080  | 09:28 |
| Air Blank          | 0.000  | 09:29 |
| Control Test       | 0.081  | 09:29 |
| Air Blank          | 0.000  | 09:30 |
| Control Test       | 0.081  | 09:30 |
| Air Blank          | 0.000  | 09:31 |
| Control Test       | 0.080  | 09:31 |
| Air Blank          | 0.000  | 09:31 |
| Control Test       | 0.080  | 09:32 |
| Air Blank          | 0.000  | 09:32 |
| Control Test       | 0.080  | 09:33 |
| Air Blank          | 0.000  | 09:33 |
| Control Test       | 0.081  | 09:33 |
| Air Blank          | 0.000  | 09:34 |
| Control Test       | 0.080  | 09:34 |
| Air Blank          | 0.000  | 09:35 |
| Control Test       | 0.081  | 09:35 |
| Air Blank          | 0.000  | 09:36 |
| Control Test Stats |        |       |
| Average            | 0.0804 |       |
| Std Dev            | 0.0005 |       |
| Rel Std Dev(%)     | 0.6423 |       |

B. C-1193

Operator's Signature