

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIALEAH PD

Time of Inspection: 09:41

Date of Inspection: 07/28/2022

Serial Number: 80-000876

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1224982 Exp: 10/31/2022
0.000	0.050	0.080	0.205	0.075 / 0.081
0.000	0.051	0.080	0.204	0.075 / 0.081
0.000	0.050	0.081	0.204	0.074 / 0.081

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. NEW DRY GAS CYLINDER INSTALLED

LOT # 402370890

EXPIRATION 05-09-2025

H #1614

07-28-2022

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

H #1614

CESAR CORREA

Signature and Printed Name

07/28/2022
Date