

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.

Time of Inspection: 07:59

Date of Inspection: 09/25/2022

Serial Number: 80-001004

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:25721080A2 Exp: 11/05/2023
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PTO Rly #121 (WAGNER)

ROBERT F WAGNER

Signature and Printed Name

09/25/2022
Date

From: [Wagner, Robert](#)
To: [Soto, Israel](#)
Subject: RE: <EXTERNAL EMAIL>Instrument 80-001004 and 80-006233 Agency Inspections
Date: Monday, October 17, 2022 11:01:08 AM

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

Talk about fat finger the keys. 5 TY

From: Soto, Israel [mailto:IsraelSoto@fdle.state.fl.us]
Sent: Monday, October 17, 2022 10:21 AM
To: Wagner, Robert <robert.wagner@bcso.us>
Subject: <EXTERNAL EMAIL>Instrument 80-001004 and 80-006233 Agency Inspections

WARNING: This email originated outside of the Brevard County Sheriff's Office. **DO NOT CLICK** links or attachments unless you recognize the sender and know the content is safe. If you believe this is SPAM please forward to SPAM@bcso.us.

Good Morning

The October inspections for Intoxilyzers 80-001004 and 80-006233 show that 55 simulators were used (see attached).

Can you confirm how many simulators were used for the inspections?



Israel Soto

Department Inspector
Alcohol Testing Program
Florida Department of Law Enforcement
israelsoto@fdle.state.fl.us
Office: (850) 617-1286
Cell: (850)-274-2757

Please note: Florida has a very broad public records law. Most written communications to or from state officials regarding state business are public records available to the public and media upon request. Your email communications may therefore be subject to public disclosure.

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.

Serial Number: 80-001004

Time of Inspection: 11:10

Date of Inspection: 10/11/2022

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#: 202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#: 202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 25721080A2 Exp: 11/05/2023
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.080	0.198	0.079

Number of Simulators Used: 55

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ROBERT F WAGNER

Signature and Printed Name

10/11/2022

Date