

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.
Time of Inspection: 08:52

Date of Inspection: 09/25/2022

Serial Number: 80-001010
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:25721080A2 Exp: 11/05/2023
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.081	0.203	0.079
0.000	0.050	0.081	0.204	0.079

Number of Simulators Used: 5

Remarks:

*AFTER DIAGNOSTIC FAIL ON 1ST INSPECTION
ATTEMPT RAN 5 DIAGNOSTIC V'S FROM INSPECTION MENU.
ALL "OK" RAN THIS INSPECTION.
IN COMPLIANCE!*

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PTD [Signature]

ROBERT F WAGNER

Signature and Printed Name

09/25/2022
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.
Time of Inspection: 08:06

Date of Inspection: 09/25/2022

Serial Number: 80-001010
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:25721080A2 Exp: 11/05/2023
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.081	0.202	0.080
0.000	0.050	0.081	0.203	0.079

Number of Simulators Used: 5

Remarks:

DIAGNOSTIC FAIL UNK REASONNon-compliance:DF UNKNOWN REASON.

*RAN 5 DIAGNOSTIC V's FROM INSPECTOR MENU.
ALL RESULTS "OK"
RAN ANOTHER MONTHLY INSPECTION.*

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

FTO [Signature] 11/21/2022

ROBERT F WAGNER

Signature and Printed Name

09/25/2022
Date