

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 09:32

Date of Inspection: 11/16/2022

Serial Number: 80-001018
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:29821080A4 Exp: 12/05/2023
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.195	0.080
0.000	0.049	0.078	0.195	0.081

Number of Simulators Used: 5

Remarks:

REMOVED FROM EVIDENCE AT PTDF ON 11-16-2022
ANALYZED BY R. LIGHTSEY
11-16-2022
AT 0932

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ROBERT D LIGHTSEY

Signature and Printed Name

11/16/2022
Date