

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO

Time of Inspection: 07:56

Date of Inspection: 10/31/2022

Serial Number: 80-001077

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#: 202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#: 202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 402477283 Exp: 06/24/2025
0.000	0.045	0.077	0.189 / 0.196	0.080
0.000	0.047	0.078	0.192 / 0.196	0.080
0.000	0.047	0.078	0.194 / 0.198	0.079

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance. *Everything was connected. The temperature might have lowered when I moved it to connect it to the Intoxilyzer*

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GARY W CARTWRIGHT JR

Signature and Printed Name

10/31/2022

Date