

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FLAGLER COUNTY SO
Time of Inspection: 09:09

Date of Inspection: 07/25/2022

Serial Number: 80-001138
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06022080A3 Exp: 04/05/2024
0.000	0.050	0.080	0.169 / 0.199	0.081
0.000	0.050	0.080	0.169 / 0.198	0.081
0.000	0.051	0.081	0.169 / 0.199	0.081

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance. After Low result it was determined that there was a small hole in the reference solution bottle. The reference solution was disposed of and a new reference solution with same Lot and expiration date was opened. After the new reference solution was ready, the test was repeated, which gave a valid result.

Amended 8/2/22
Master Deputy Edward Faircloth #53

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWARD W FAIRCLOTH
Signature and Printed Name

07/25/2022
Date