

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SHERIFF OFFICE
 Time of Inspection: 19:30

Date of Inspection: 10-20-2022 Serial Number: 80-001367
 Software: 8100.27

| CHECK OR TEST | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | |
| Diagnostic Check (Pre-Inspection): OK | | |
| Alcohol Free Subject Test: 0.000 | | |
| Mouth Alcohol Test: Slope Not Met | | |
| Interferent Detect Test: Interferent Detect | | |
| Diagnostic Check (Post-Inspection): OK | | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.08g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.20g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____ |
|----------------------------|---|---|---|--|
| | | | | |
| | | | | |
| | | | | |

Number of Simulators Used: 1

Remarks: AMENDED; INSTRUMENT WOULD NOT ALLOW ME TO PERFORM AN INSPECTION! (DSP FAIL) I TRIED TWICE WITH OTHER POWER OPTIONS!

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M. Deane

M. DEANE

Signature and Printed Name

11-2-2022
Date