

Return Material Authorization

Ship to:



CMI, Inc.



Enforcement Electronics

Shipment to repair facility authorized by: TJ Graham on 4/19/2022

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001651

Bill To Address:

FDLE FMROC

4700 Terminal Drive

Fort Myers, FL 33991

Ship to Address:

FDLE FMROC

4700 Terminal Drive Suite 1

Ft. Myers, FL 33907

Reason for Return:

DVM monitor shows breath hose temperature and Cell temperature at .0 FDLE is the POC for the repairs.

Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$_____.
- ☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: David Reyes-Rivera

Phone #: 239-335-7563 Email: davidreyes@fdle.state.fl.us

ATP Contact Name: David Reyes ATP Email: davidreyes@fdle.state.fl.us

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: Collier County Sheriff's Office

Serial Number: 80-001651

Time of Inspection:

Date of Inspection: 12/15/2022

Software:

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume Check: OK			Barometric Pressure Sensor Check: OK		
Alcohol Free Subject Test: 0.000			Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect			Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks: Instrument is out at CMI for repair.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
David E. Reyes-Rivera
Signature and Printed Name

12/15/2022
Date