Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL Time of Inspection: 10:58 Serial Number: 80-003410

Date of Inspection: 09/27/2022

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:212654 Exp: 01/05/2024
/ 0.000	0.048	0.078	0.195	0.081
/ 0.000	0.048	0.079	0.197	0.081
/ 0.000	0.048	0.079	0.197	0.081

Number	of	Simula	tors	Used:	5

Remarks:

A F / M A: RFI Detect. 00: Ambient Fail.

RFI Detect: Area checked for electronics, OK second test.

Ambient Fail: Ventilation increased, area cleared of possible contaminants, OK second test.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

5GT. (1)

JOSHUA M BUDAY

Signature and Printed Name

09/27/2022 Date



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Florida Highway Patrol Instrument Serial Number: 80-003410 **AGENCY INSPECTION DATA REVIEW** Time of Inspection: 10:58 Agency Inspector: Joshua Buday Date of Inspection: 09-27-2022 **Agency Inspection Discrepancy:** Incomplete Untimely/Not Received **Erroneous Information** Procedural П Other Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded. Lot Number □ Expiration Date for _____g/ 210L □ Alcohol Reference Solution □ Dry Gas Standard is □ Incorrect □ Expired. FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The ☐REASON for repeating the following test(s); OR the Possible Cause and Corrective Action Taken on the following test(s) was not recorded: Mouth Alcohol Test 🗵 Alcohol Free Subject Test Alcohol Free Test П Interferent Detect Test 0.05 g/210L Test 0.08 g/210L Test ☐ 0.20 g/210L Test □ 0.08 g/210L Dry Gas Standard Test FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the П requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use. The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC. Other: OTHER ELECTRONIC DATA REVIEW Comments: **Login Records** Date: **Cylinder Change Records Control Test Records** Date: **Diagnostic Check Records** Date: **CORRECTIVE ACTION** Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by __ Provide a written explanation regarding the referenced item(s) to the Department Inspector by (Date). Upload the Agency Inspection(s). Remove the instrument from evidentiary use until otherwise directed by the Department. No action required Other:

Israel Soto

Digitally signed by Israel Soto Date: 2022.10.07 09:46:52 -04'00'

Signature of Alcohol Testing Program Staff Member

10/7/2022 Date

Soto, Israel

From: Buday, Joshua <JoshuaBuday@flhsmv.gov>

Sent: Monday, October 31, 2022 1:44 PM

To: Soto, Israel **Subject:** AI - 80-003410

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

Israel,

Good afternoon. Just an FYI for you sir, I had two issues when attempting to complete an AI on this instrument today:

Had an MA/AF issue on the first attempt, and then a power surge on the second attempt which caused the instrument to shut off during the test. I will be conducting another AI attempt later today for October, but just you are in the know when you see two AIs for the month uploaded.

Respectfully,

Sergeant Joshua Buday District Supervisor Troop D – Deland / Volusia District

Florida Highway Patrol
1551 East International Speedway Boulevard
Deland, Florida 32724
Office: (386) 624-0592
JoshuaBuday@flhsmv.gov
www.flhsmv.gov/FHP
www.BeATrooper.com

This email originated from a Florida Department of Highway Safety and Motor Vehicles email address.

Always use caution when clicking links or opening attachments unless you recognize the sender and know the content is safe.

Please Note: Florida has very broad public records laws. Unless a statutory exemption applies, emails are subject to public disclosure.

This email has been scanned by the Symantec Email Security.cloud service.

1

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL Time of Inspection: 13:25

Date of Inspection: 10/31/2022

Serial Number: 80-003410

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000				

Number	of	Simulator	S	US	ed:	5		
Remarks	5:							

00: Sequence Aborted. Non-compliance: .

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.	
I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Perm performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.	it and that I
567 JOSHUA M BUDAY Signature and Printed Name	

10/31/2022 Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL Time of Inspection: 15:58

Date of Inspection: 10/31/2022

Serial Number: 80-003410

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:212654 Exp: 01/05/2024
0.000	0.046	0.077	0.194	0.078
0.000	0.046	0.078	0.195	0.078
0.000	0.047	0.078	0.196	0.078

Number	of	Simulators	Used:	5	

Remarks:

UNIT CLEAN / FUNCTIONAL AT TIME OF INSPECTION.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

10/31/2022 Date