

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Fort Myers PD

Serial Number: 80-005655

Time of Inspection:

Date of Inspection: September 2022

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks: On Tuesday, September 27, 2022, Officer Daniel Aguilera #550 was scheduled to complete the Agency Inspections on the agency assigned intoxilyzer's. On that day, Tuesday 9/27/2022, Fort Myers PD activated all sworn personnel to an activated Hurricane Shift status and all officers were mandated to report into work, in preparedness of Hurricane Ian. Hurricane Ian impacted our city and area on Wednesday, 9/28/2022, causing all officers to work a mandated shift for the purposes of Hurricane only related duties until the city could return to normal operations. Officer Aguilera was on this mandated shift until October 11, 2022 and was not able to complete the agency inspections for September 2022.

The above instrument complies (____) does not comply (____) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Daniel J Aguilera

Signature and Printed Name

10/26/2022
Date