



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Manatee CSO

Instrument Serial Number: 80-006559

AGENCY INSPECTION DISCREPANCY

Agency Inspector: William Coleman

Inspection Date: 11/14/2022

- Agency inspection discrepancy options including test repeats, calendar month, expired solutions, and lot numbers.

REQUESTED CORRECTIVE ACTION

- Requested corrective action options including hand-written amendments, written explanations, and instrument removal.

Taylor Gutschow

Digitally signed by Taylor Gutschow Date: 2023.05.17 09:15:50 -04'00'

05/17/2023

Signature of Alcohol Testing Program Staff Member

Date

Ammended
 Florida Department of Law Enforcement
 Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
 Time of Inspection: 10:24

Date of Inspection: 11/14/2022

Serial Number: 80-006559
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202101C <i>wdc</i> Exp: 08/17/2023 01/12/2023	0.08g/210L Test (g/210L) Lot#: 202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#: 202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 402419647 Exp: 04/14/2025
0.000	0.049	0.079	0.199	0.081
0.000	0.050	0.079	0.199	0.081
0.000	0.049	0.078	0.199	0.081

Number of Simulators Used: 5

Remarks: Wrong .05 Reference Solution Entered. *(wdc)*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WILLIAM COLEMAN

Signature and Printed Name

11/14/2022

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 16:57

Date of Inspection: 03/23/2022

Serial Number: 80-006559
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202100B Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1303107 Exp: 03/28/2023
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.199	0.081
0.000	0.050	0.080	0.200	0.080


Number of Simulators Used: 5

Remarks:

.20 Reference Solution entered incorrectly; should read 202005A

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WILLIAM D COLEMAN

Signature and Printed Name

03/23/2022
Date