## Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Broward County SO

Serial Number: 80-007008

Time of Inspection:

Date of Inspection: January 2022

Software: 8100.27

| CHECK OR TEST                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     |    |
| Diagnostic Check (Pre-Inspection): OK       |     |    |
| Alcohol Free Subject Test: 0.000            |     |    |
| Mouth Alcohol Test: Slope Not Met           |     |    |
| Interferent Detect Test: Interferent Detect |     |    |
| Diagnostic Check (Post-Inspection): OK      |     | -  |

| Alcohol Free<br>Test<br>(g/210L) | 0.05g/210L Test<br>(g/210L)<br>Lot#:<br>Exp: | 0.08g/210L Test<br>(g/210L)<br>Lot#:<br>Exp: | 0.20g/210L Test<br>(g/210L)<br>Lot#:<br>Exp: | 0.08 g/210L<br>Dry Gas Std Test<br>(g/210L)<br>Lot#:<br>Exp: |
|----------------------------------|--|--|--|--|
|                                  |  |  |  |  |

| Number of | Simulators | Use | ed: |           |    |     |            |     |      |    |      |     |     |          |
|-----------|------------|-----|-----|-----------|----|-----|------------|-----|------|----|------|-----|-----|----------|
| Remarks:  | Inspection | was | not | conducted | as | the | instrument | was | sent | to | CMI. | Inc | for | repairs. |

| The above           | instrument complies () with Chapter 11D-8, FAC.   |
|---------------------|---|
| I certify performed | that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I this inspection in accordance with the provisions of Chapter 11D-8, FAC. |
|                     | Signature and Printed Name  01/05/2022 Date   |

## Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Broward County SO

Serial Number: 80-007008

Time of Inspection:

Number of Simulators Used:

Date of Inspection: March 2022

Software: 8100.27

| CHECK OR TEST                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     |    |
| Diagnostic Check (Pre-Inspection): OK       |     |    |
| Alcohol Free Subject Test: 0.000            |     |    |
| Mouth Alcohol Test: Slope Not Met           |     |    |
| Interferent Detect Test: Interferent Detect |     |    |
| Diagnostic Check (Post-Inspection): OK      |     |    |

| Alcohol Free<br>Test<br>(g/210L) | 0.05g/210L Test<br>(g/210L)<br>Lot#:<br>Exp: | 0.08g/210L Test<br>(g/210L)<br>Lot#:<br>Exp: | 0.20g/210L Test<br>(g/210L)<br>Lot#:<br>Exp: | 0.08 g/210L<br>Dry Gas Std Test<br>(g/210L)<br>Lot#:<br>Exp: |
|----------------------------------|--|--|--|--|
|                                  |  |  |  |  |

|          |            |     | -   |           |    |     |            |     |      |     |     |     |     |         |  |
|----------|------------|-----|-----|-----------|----|-----|------------|-----|------|-----|-----|-----|-----|---------|--|
| Remarks: | Inspection | wag | not | conducted | 20 | the | instrument | WSO | een+ | to. | CMT | Tnc | for | ropaire |  |

| The above instrument complies () does not comply () with Chapter 11D-8, FAC.  |
|---|
| I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC. |
| I a la la   |
| Signature and Printed Name  |
| 03/02/2022  |
| Date  |