

**** AMENDED ****

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 09:32

Date of Inspection: 11/30/2022

Serial Number: 80-007443

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402310313 Exp: 01/15/2025
0.000	0.047	0.077	0.196	0.081
0.000	0.047	0.077	0.196	0.081
0.000	0.047	0.077	0.196	0.081

Number of Simulators Used: 5

Remarks:

IN COMPLIANCE/ CHECKED CASE NO CRACKS

11/30/22 When Completing first Agency Inspection, Instrument Froze on Final Analytical Check saying OK. Instrument had to be turned off, upon power up Instrument had not saved Inspection which was compliant. Instrument run through Second Agency Inspection and was again Compliant.
11/30/2022 Sgt. Anthony Dobosiewicz

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]
Signature and Printed Name

ANTHONY N DOBOSIEWICZ

11/30/2022
Date

**** AMENDED ****