

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: FDLE (On loan to Zephyrhills PD)

Instrument Serial Number: 80-000206

AGENCY INSPECTION DATA REVIEW											
Agency Inspector: Ailisa Starnes							f Inspection: 05	6/26/2022	Time of Inspection: 10:48:15		
							timely/Not Received Erroneous Information				
				Procedural	<mark>⊠ C</mark>	other (Mis	sing Required	Remarks)			
	□ Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.										
	Lot Number Expiration Date forg/ 210L Alcohol Reference Solution Dry Gas Standard is Incorrect Expired.										
\boxtimes											
	Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The 🛛 REASON for repeating the following										
		s); OR the 🛛 Possible (
		Alcohol Free Subject							Interferent Detect Test		
		0.05 g/210L Test		□ 0.08 g/21	0L Test		0.20 g/210L Te	est 🗆	0.08 g/210L Dry Gas Standard Test		
	FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the										
	instrument from service and notify the Department Inspector."										
	The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.										
	□ The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the										
	requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.										
	The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.										
		Chapter 11D-8, FAC a	nd the	instrument was	s not rem	loved fro	m evidentiary u	se.			
	The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.										
	Other	:									
AT											
OTHER ELECTRONIC DATA REVIEW											
		Records		Commen	ISI						

OTHER ELECTRONIC DATA REVIEW					
Login Records Date:	Comments: The Alcohol Free Subject / Mouth Alcohol Test was repeated due to an Improper Sample.				
Cylinder Change Records Date:	Please include a remark describing the corrective action taken before repeating the test.				
Control Test Records Date:					
Diagnostic Check Records Date:					

CORRECTIVE ACTION

Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by <u>9/4/2022</u> (Date).

Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).

- □ Upload the Agency Inspection(s).
- Remove the instrument from evidentiary use until otherwise directed by the Department.
- No action required
- Other: _____

Jaylor Jutschon

Signature of Alcohol Testing Program Staff Member

8/4/2022 Date

FDLE/ATP Form 42 October 2017 Issuing Authority: Alcohol Testing Program PRINTED COPIES UNCONTROLLED For Internal ATP Use Only