From: Gutschow, Taylor
To: "Lt. Will Cartwright"
Cc: Graham, Thomas
Subject: Missed March AI on 80-001002
Date: Thursday, June 2, 2022 1:08:00 PM
Attachments: Affected Breath Test on 80-001002 for March 2022.pdf
Form 40- Blank AI.pdf

Please have the AI who missed the March inspection fill out the attached blank Form 40 indicating that there was no Agency Inspection conducted for the month of March. Then have them send the form back to me.

Because the inspection got missed, the instrument won't be compliant with Chapter 11D-8 FAC that month. I've attached the one affected breath test. Either you or the AI who missed the inspection will need to contact your local State Attorney's Office and let them know about the non-compliance. If they have any questions on what that means, they can contact me or David Reyes (the other Department Inspector in Fort Myers) to discuss. Our contact info is in the link below.

http://www.fdle.state.fl.us/Alcohol-Testing-Program/Staff

Taylor Gutschow
Department Inspector for the Alcohol Testing Program
Florida Department of Law Enforcement
Fort Myers Regional Operations Center
4700 Terminal Drive, Suite 3
Fort Myers, FL 33907

Desk #239-335-7500
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

From: Lt. Will Cartwright <wcartwright@hardeeso.onmicrosoft.com>
Sent: Thursday, June 2, 2022 12:58 PM
To: Gutschow, Taylor <TaylorGutschow@fdle.state.fl.us>
Subject: Re: Missed March AI on 1002

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

Yes the guy that's was over it at that time failed to do it that month so I did it on April 8

Get Outlook for iOS

From: Gutschow, Taylor <TaylorGutschow@fdle.state.fl.us>
Sent: Thursday, June 2, 2022 12:57:04 PM
To: Lt. Will Cartwright <wcartwright@hardeeso.onmicrosoft.com>
Subject: Missed March AI on 1002

Back in March, instrument 1002 had a breath test run on it (see attached PDF) but it did not get inspected that month. Do you know why?

Taylor Gutschow
Department Inspector for the Alcohol Testing Program
Florida Department of Law Enforcement
Fort Myers Regional Operations Center
4700 Terminal Drive, Suite 3
Fort Myers, FL 33907

Desk #239-335-7500
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.