

## Gutschow, Taylor

---

**From:** Michael Feltovic <Michael.Feltovic@sarasotasheriff.org>  
**Sent:** Monday, January 23, 2023 12:01 PM  
**To:** Gutschow, Taylor  
**Subject:** Re: Intox 1344  
**Attachments:** 80-001344 0522 AI OUT.pdf; 80-001344 0622 AI OUT.pdf; 80-001344 0722 AI OUT.pdf

**CAUTION:** This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

Taylor,

After the April 2022 inspection, it was brought to my attention that the breath tube on that instrument had a leak. I took the instrument out of service until it could be sent to FDLE or repair. As it turns out, when I sent it to FDLE for DI, an o ring was replaced and the breath tube was not leaking anymore (I think I mentioned it to you). I didn't perform AI on that instrument for May, June or July 2022, but I did complete Form 40's for each month just to document the lack of inspections and attached them. If you have any other questions, let me know. Thanks!



Michael Feltovic  
Law Enforcement Deputy  
Special Operations Bureau  
6010 Cattleridge Blvd. Sarasota, Florida 34232  
Phone Number: (941) 861-4076  
[Michael.Feltovic@sarasotasheriff.org](mailto:Michael.Feltovic@sarasotasheriff.org)  
[www.SarasotaSheriff.org](http://www.SarasotaSheriff.org)



---

**From:** Gutschow, Taylor <TaylorGutschow@fdle.state.fl.us>  
**Sent:** Friday, January 20, 2023 2:25 PM  
**To:** Michael Feltovic <Michael.Feltovic@sarasotasheriff.org>  
**Subject:** Intox 1344

**CAUTION:** This email originated from outside of the Sheriff's Office. Do not click links or open attachments unless you know the sender is safe.

Do you have any Agency Inspection reports for instrument 1344 for May, June, or July 2022? If so, can you email them to me? Thanks!

Taylor Gutschow

Department Inspector for the Alcohol Testing Program  
Florida Department of Law Enforcement  
Fort Myers Regional Operations Center  
4700 Terminal Drive, Suite 3  
Fort Myers, FL 33907

Desk #239-335-7500  
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

---

"Please Note: All mail to and from the Sarasota County Sheriff's Office is subject to the Public Records Law of Florida."

## Gutschow, Taylor

---

**From:** Michael Feltovic <Michael.Feltovic@sarasotasheriff.org>  
**Sent:** Saturday, January 28, 2023 3:44 PM  
**To:** Gutschow, Taylor  
**Subject:** Inspections  
**Attachments:** 80-001344 0123 AI 1.pdf; 80-001346 0123 AI 1.pdf

**CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.**

Taylor,

I had 2 instruments unable to complete inspections on first attempts. I had a simulator issue. On my .08 simulator, I think it was vapor locked, and the sample wasn't being introduced into the instrument. Since I was doing multiple instruments at the same time, it affected 2 inspections. I aborted the inspections when I determined it was the simulator, was able to clear the simulator, and continue with the inspections on all our instruments without any further issues. I have attached the 2 inspection reports with hand written notes explaining the issue. Thanks!



Michael Feltovic  
Law Enforcement Deputy  
Special Operations Bureau  
6010 Cattleridge Blvd. Sarasota, Florida 34232  
Phone Number: (941) 861-4076  
[Michael.Feltovic@sarasotasheriff.org](mailto:Michael.Feltovic@sarasotasheriff.org)  
[www.SarasotaSheriff.org](http://www.SarasotaSheriff.org)



---

"Please Note: All mail to and from the Sarasota County Sheriff's Office is subject to the Public Records Law of Florida."