

Gutschow, Taylor

From: Gutschow, Taylor
Sent: Wednesday, September 7, 2022 10:27 AM
To: Anthony Dobosiewicz
Subject: Failed July AI on 7440
Attachments: Failed July AI.pdf; 80-007440 Affected Breath Test.pdf

Good morning,

The attachment marked "Failed July AI" is the inspection we were talking about last night. It read triple 0's on both of the 0.08 ARS Tests. The inspection didn't pass and it wasn't repeated, so there was no valid AI for July. It looks like the instrument may have been taken out of service right after that because there is no August AI or further breath tests on it. It also looks like another instrument you were inspecting at the time, instrument 6637, was having similar issues but you got it to read 0.08 values on the retest. (I emailed you about that inspection and it needs a remark added). This instrument, 7440, is the one that had the sticky note marked EEPROM Fail when you brought it to me last week for its annual inspection. However, the July AI passed its initial Diagnostic Check and I don't see any EEPROM issues during that inspection.

The original paperwork for the failed July inspection will need a remark added to it describing what happened and why a passing AI couldn't be obtained. I'll need you to scan it back to me once the comment is added. I will also need you to distribute the attachment called "Affected Breath Test" to your local State Attorney's Office and let them know it was run on an instrument that wasn't compliant with 11D-8 for the month of July (since there was no passing AI for the month). If they have any questions, you can provide my contact info and have them give me a call/email.

Taylor Gutschow

Department Inspector for the Alcohol Testing Program
Florida Department of Law Enforcement
Fort Myers Regional Operations Center
4700 Terminal Drive, Suite 3
Fort Myers, FL 33907

Desk #239-335-7500
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.