



Highlands County Sheriff's Office

Sheriff Paul Blackman

400 S. Eucalyptus Street
Sebring, Florida 33870
863-402-7200
www.highlandssheriff.org



Dually Accredited

January , 2022

Taylor Gutschow
Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Fort Myers, FL 33907

REF: 01-2022 Inspection

Dear Mr. Gutschow,

Agency Inspection results for the month of January, 2022 inspection.
Instruments 80-007505 passed inspection without any issues. Instrument 80-007507 also passed without issue other than I had to repeat the Interferent test due to a small seal leak in the simulator which I retightened and proceeded without issue.

Sgt G,Glasco
Agency Inspector

Cc: File



Highlands County Sheriff's Office

Sheriff Paul Blackman

400 S. Eucalyptus Street
Sebring, Florida 33870
863-402-7200
www.highlandssheriff.org



Dually Accredited

February , 2022

Taylor Gutschow
Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Fort Myers, FL 33907

REF: 02-2022 Inspection

Dear Mr. Gutschow,

Agency Inspection results for the month of February, 2022 inspection.
Instruments 80-007505 passed inspection without any issues. Instrument 80-007507 also passed without issue

Sgt G, Glasco
Agency Inspector

Cc: File



Highlands County Sheriff's Office

Sheriff Paul Blackman

400 S. Eucalyptus Street
Sebring, Florida 33870
863-402-7200
www.highlandssheriff.org



Dually Accredited

March 2022

Taylor Gutschow
Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Fort Myers, FL 33907

REF: 03-2022 Inspection

Dear Mr. Gutschow,

Agency Inspection results for the month of March, 2022 inspection. Instruments 80-007505 passed inspection without any issues. Instrument 80-007507 also passed without issue

Sgt G, Glasco
Agency Inspector

Cc: File



Florida Department of
Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Manatee County Sheriff's Office Instrument Serial Number: _____

Agency Notification	Records Audit:
Agency Inspector/Contact: <u>Sgt. William Coleman</u>	The following records were audited: <input type="checkbox"/> Instrument Registration <input type="checkbox"/> Agency Inspection Reports <input type="checkbox"/> Breath Test Instrument Repair/Maintenance Records <input type="checkbox"/> Dry Gas Standard Certificate(s) of Analysis <input type="checkbox"/> Other: <u>Prior field notes</u>
Date of Notification: <u>5/9/2022</u>	
Comments: 	

Instrument/Area:	Equipment:	Supplies:
<input type="checkbox"/> Clean/Dry <input type="checkbox"/> Secure <input type="checkbox"/> Limited Access <input type="checkbox"/> Other: <u>Well ventilated</u>	<input type="checkbox"/> Proper Number of Simulators <input type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input type="checkbox"/> Class A Glassware	<input type="checkbox"/> Distilled/Deionized Water <input type="checkbox"/> Mouth Alcohol Solution <input type="checkbox"/> Acetone Stock Solution <input type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard <input type="checkbox"/> Mouth Pieces
Comments: Observed Agency Inspection process. Trooper Ledford from FHP (Troop F) was also present and conducting his own inspections. Took several Manatee CSO and FHP instruments back to FDLE with me for their annual Department Inspections. Sent refill of acetone stock solution. 		

Taylor Gutschow

Signature of Alcohol Testing Program Staff Member

Digitally signed by Taylor Gutschow
Date: 2022.05.10 11:57:15 -04'00'

5/10/2022

Date