## FLORIDA DEPARTMENT OF LAW ENFORCEMENT ALCOHOL TESTING PROGRAM BREATH ALCOHOL TEST AFFIDAVIT

Instrument Type: Intoxilyzer 8000
Instrument Registered To: HILLSBOROUGH CO SO
Instrument Serial Number: 80-000830 Software: 8100.27

Date of Test: 09/21/2022

Date of Last Agency Inspection: 09/01/2022

Observation Period Began: 20:53
Subject's Name: RAFAEL R RODRIGUEZ

DOB: 12/25/1985 Sex: M

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time	
	Diagnostics Check	OK	21:13	
	Air Blank	0.000	21:14	
	Control Test	0.083	21:14	
	Air Blank	0.000	21:15	
	Subject Sample #1	0.000	21:16	
	Air Blank	0.000	21:16	
	Air Blank	0.000	21:18	
	Subject Sample #2	0.000	21:19	
	Air Blank	0.000	21:20	
	Control Test	0.081	21:20	
	Air Blank	0.000	21:20	
	Diagnostics Check	OK	21:21	

Cylinder Lot: 06021080A2 Exp: 04/05/2023

Signature of Notary Public-State of Florida Printed Name of

Printed Name of Notary Public-State of Florida

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.