



Agency Florida Department of Law Enforcement

S/N 80-000223

Date In 03-22-2022 DI Completion Date 03-23-2022

☐ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By _____

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service: Printer jammed during 0.050
stability check, repeated test for a clean print.
FDLE loaner will return to evidentiary use when sent
to an agency. IS

Department Inspection

By IS

Attachments

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☐ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use
- ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Tech Review / Date

Admin Review 04'06' Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 13:33

Date of Inspection: 03/23/2022

Serial Number: 80-000223

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.050	0.079	0.200	0.081
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.079	0.198	0.081
0.000	0.048	0.081	0.198	0.080
0.000	0.047	0.080	0.197	0.081
0.000	0.048	0.080	0.198	0.081

Standard Deviations	0.0008	0.0006	0.0011	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISRAEL SOTO

Signature and Printed Name

03/23/2022

Date

Stability Checks

DLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000223
3/23/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:32
Control Test	0.049	11:33
Air Blank	0.000	11:33
Control Test	0.049	11:34
Air Blank	0.000	11:35
Control Test	0.049	11:35
Air Blank	0.000	11:36
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.050
Paper jammed

[Signature]

DLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000223
3/23/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:32
Control Test	0.049	11:33
Air Blank	0.000	11:33
Control Test	0.049	11:34
Air Blank	0.000	11:35
Control Test	0.049	11:35
Air Blank	0.000	11:36
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.050 repeat

[Signature]

Operator's Signature

DLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000223
3/23/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:37
Control Test	0.078	11:38
Air Blank	0.000	11:38
Control Test	0.078	11:39
Air Blank	0.000	11:39
Control Test	0.078	11:40
Air Blank	0.000	11:41
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet

[Signature]

Operator's Signature

DLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000223
3/23/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:43
Control Test	0.200	11:43
Air Blank	0.000	11:44
Control Test	0.199	11:45
Air Blank	0.000	11:45
Control Test	0.198	11:46
Air Blank	0.000	11:46
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

[Signature]

Operator's Signature

DLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000223
3/23/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:48
Control Test	0.079	11:48
Air Blank	0.000	11:49
Control Test	0.080	11:49
Air Blank	0.000	11:50
Control Test	0.079	11:50
Air Blank	0.000	11:51
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Dry

[Signature]

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 10:57

Date of Inspection: 03/23/2022

Serial Number: 80-000223

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

N/A Compliance not determined

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

03/23/2022

Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000223, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000223</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/23/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:33</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel Soto
Date: 2022.03.23 14:25:06 -04'00'

03/23/2022

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Florida Department of Law Enforcement

S/N 80-000223

Florida Department of
Law Enforcement

Date In 08-09-2022 DI Completion Date _____

☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By	IS	Quality Checks	By	Date	Flow Calibration	By	Date																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____			☐ Breath Tube Screen ☐ Replace External O-Rings ☐ Instrument Set Up Verified ☐ R-Value _____ ☐ Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) ☐ Barometric Pressure Check Gauge ID # _____ ☐ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																										
			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A															
Simulator	Serial #	Lot #/Exp																														
0.050																																
0.080																																
0.200																																
0.080 DGS	N/A																															
			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Maintenance</th> <th>By</th> </tr> </thead> <tbody> <tr> <td colspan="2">☐ Battery Replacement</td> <td></td> </tr> <tr> <td colspan="2">☐ Dry Gas Regulator Replacement</td> <td></td> </tr> <tr> <td colspan="2">☐ Breath Tube Replacement</td> <td></td> </tr> <tr> <td colspan="2">☐ Other _____</td> <td></td> </tr> <tr> <td colspan="2"> </td> <td></td> </tr> <tr> <td colspan="2"> </td> <td></td> </tr> <tr> <td colspan="2"> </td> <td></td> </tr> </tbody> </table>						Maintenance		By	☐ Battery Replacement			☐ Dry Gas Regulator Replacement			☐ Breath Tube Replacement			☐ Other _____											
Maintenance		By																														
☐ Battery Replacement																																
☐ Dry Gas Regulator Replacement																																
☐ Breath Tube Replacement																																
☐ Other _____																																

Calibration Adjustment	Department Inspection																																																																						
By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			By _____ Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.050</td> <td></td> </tr> <tr> <td>0.080</td> <td></td> </tr> <tr> <td>0.200</td> <td></td> </tr> </tbody> </table> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Attachments</th> </tr> </thead> <tbody> <tr> <td>☐ Form 41</td> <td>☐ Post-Stability Checks</td> </tr> <tr> <td>☐ Stability Checks</td> <td>☐ Flow Calibration</td> </tr> <tr> <td>☐ Calibration Certificate</td> <td>☐ Form 40</td> </tr> <tr> <td>☐ Calibration Adjustment</td> <td>☒ Other <u>Form 51</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200		Attachments		☐ Form 41	☐ Post-Stability Checks	☐ Stability Checks	☐ Flow Calibration	☐ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☒ Other <u>Form 51</u>
Simulator	Serial #	Lot #	Expiration																																																																				
0.000		N/A	N/A																																																																				
0.040																																																																							
0.100																																																																							
0.200																																																																							
0.300																																																																							
0.080 DGS	N/A																																																																						
Simulator	Serial #	Lot #	Expiration																																																																				
0.050																																																																							
0.080																																																																							
0.200																																																																							
0.080 DGS	N/A																																																																						
Simulator	Serial Number																																																																						
0.000																																																																							
Interferent																																																																							
0.050																																																																							
0.080																																																																							
0.200																																																																							
Attachments																																																																							
☐ Form 41	☐ Post-Stability Checks																																																																						
☐ Stability Checks	☐ Flow Calibration																																																																						
☐ Calibration Certificate	☐ Form 40																																																																						
☐ Calibration Adjustment	☒ Other <u>Form 51</u>																																																																						
Notes/Suggested Service: <u>Instrument does not power on.</u> <u>Sending instrument to repair. Compliance with 11D-8</u> <u>unable to be determined. IS</u> _____ _____ _____ _____ _____	☐ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date _____ Admin Review / Date _____																																																																						

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Israel Soto on 08-17-2022

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000223

Bill To Address:

Alcohol Testing Program - FDLE

Ship to Address:

FDLE Off-Site Mail Facility

Alcohol Testing Program

813 B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

Instrument does not power on.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Thomas Graham

Phone #: 850-631-1351 Email: ThomasGraham@fdle.state.fl.us

ATP Contact Name: Israel Soto ATP Email: israelsoto@fdle.state.fl.us