



Agency Florida Department of Law Enforcement

S/N 80-000224

Date In 09-22-2022 DI Completion Date 09-23-2022

☐ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By _____

Barometric Pressure Gauge	ID #
---------------------------	------

☐ Post Calibration Adjustment Stability ChecksNotes/Suggested Service:

Instrument will remain at ATP until sent to an agency for use. -IS

Department Inspection

By IS

Barometric Pressure ID# 28427

Gauge 1010	<u>Instrument 1012</u>
------------	------------------------

Mouth Alcohol Solution Lot # 2022-A

Acetone Stock Solution Lot # 2022-B

Attachments

- | | |
|---|--|
| ☒ Form 41 | ☐ Post-Stability Checks |
| ☒ Stability Checks | ☐ Flow Calibration |
| ☒ Calibration Certificate | ☐ Form 40 |
| ☐ Calibration Adjustment | ☐ Other |

- ☒ Instrument Complies with Chapter 11D-8, FAC
 - ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
 - ☒ Return to/Place into Evidentiary Use
 - ☐ Remain Out of Evidentiary Use
 - ☒ Conduct an Agency Inspection Before Evidentiary Use

Phil Nicodemo Digitally signed by Phil Nicodemo
Date: 2022.09.23 11:57:09 -0400

Tech Review / Date

13:05:15
-04'00'
Admin Review / Date

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 11:17

Date of Inspection: 09/23/2022

Serial Number: 80-000224

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08121080A1 Exp: 05/05/2023
0.000	0.050	0.078	0.202	0.080
0.000	0.049	0.079	0.203	0.081
0.000	0.049	0.079	0.203	0.080
0.000	0.049	0.079	0.203	0.080
0.000	0.049	0.079	0.203	0.080
0.000	0.049	0.079	0.204	0.080
0.000	0.049	0.079	0.203	0.081
0.000	0.050	0.079	0.203	0.080
0.000	0.049	0.080	0.204	0.079
0.000	0.049	0.080	0.204	0.079

Standard Deviations	0.0004	0.0005	0.0006	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

ISRAEL SOTO

09/23/2022
 Date

stability checks

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000224
09/23/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:24
Control Test	0.051	08:25
Air Blank	0.000	08:25
Control Test	0.051	08:26
Air Blank	0.000	08:27
Control Test	0.051	08:27
Air Blank	0.000	08:28
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000224
09/23/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:29
Control Test	0.081	08:30
Air Blank	0.000	08:30
Control Test	0.079	08:31
Air Blank	0.000	08:32
Control Test	0.080	08:32
Air Blank	0.000	08:33
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

wet



Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000224
09/23/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:34
Control Test	0.206	08:35
Air Blank	0.000	08:35
Control Test	0.204	08:36
Air Blank	0.000	08:37
Control Test	0.204	08:37
Air Blank	0.000	08:38
Control Test Stats		
Average	0.2047	
Std Dev	0.0012	
Rel Std Dev(%)	0.5642	



Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000224
09/23/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:41
Control Test	0.081	08:41
Air Blank	0.000	08:41
Control Test	0.081	08:42
Air Blank	0.000	08:42
Control Test	0.080	08:43
Air Blank	0.000	08:43
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Dry



Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000224, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000224</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>09/23/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:17</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel Soto
Date: 2022.09.23 11:28:31 -04'00'

09/23/2022

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality