



INSTRUMENT PROCESSING SHEET

Agency Florida Department of Law Enforcement

S/N 80-000229

Florida Department of
Law Enforcement

Date In 5/19/2022

DI Completion Date 5/26/2022

☐ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>PN</u>	Quality Checks	By <u>JO</u>	Date <u>5/26/2022</u>	Flow Calibration	By	Date															
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>155</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.488</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks			Flow Column # ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value ☐ Post Calibration Verification (L/s) Flow Column # 32 mm (.139 - .169) 36 mm (.156 - .190) 53 mm (.228 - .278) 103 mm (.447 - .547)																	
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG113403 05/14/2023</td> </tr> </tbody> </table>			Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01/11/2024	0.080	MP6292	202201D 01/18/2024	0.200	MP6293	202201E 01/18/2024	0.080 DGS	N/A	AG113403 05/14/2023	Maintenance By ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other		
Simulator	Serial #	Lot #/Exp																				
0.050	MP6291	202201C 01/11/2024																				
0.080	MP6292	202201D 01/18/2024																				
0.200	MP6293	202201E 01/18/2024																				
0.080 DGS	N/A	AG113403 05/14/2023																				

Calibration Adjustment	By	Department Inspection	By <u>JO</u>																																																												
Barometric Pressure Gauge ID # <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A				Barometric Pressure ID# <u>28662</u> Gauge <u>1013</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP6291</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293	
Simulator	Serial #	Lot #	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial #	Lot #	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	MP6289																																																														
Interferent	MP6290																																																														
0.050	MP6291																																																														
0.080	MP6292																																																														
0.200	MP6293																																																														
Notes/Suggested Service: 0.200 g/210L stability check repeated due to printer issue. See print out. <u>JO 5/26/22</u> Instrument will remain at FDLE until requested by an agency for evidentiary use. <u>JO 5/26/22</u> Post Department Inspection - I entered the Department Inspection a second time. See print out. <u>JO 5/26/22</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Israel Soto Digitally signed by Israel Soto Date: 2022.05.26 11:43:00 -04'00' Tech Review / Date <u>5/26/22</u> Admin Review / Date <u>5/26/22</u>																																																													

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Serial Number: 80-000229

Time of Inspection: 10:45

Date of Inspection: 05/26/2022

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.049	0.078	0.200	0.079
0.000	0.048	0.078	0.200	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.048	0.078	0.200	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079


Standard Deviations	0.0005	0.0004	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



THOMAS J GRAHAM

Signature and Printed Name

05/26/2022
Date

80-000229 FDL E

Stability
Checks

5/26/22 JD

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000
05/26/2022
Software: 8100.27

SN 80-000229

Test	g/210L	Time
Air Blank	0.000	08:22
Control Test	0.050	08:23
Air Blank	0.000	08:23
Control Test	0.049	08:24
Air Blank	0.000	08:24
Control Test	0.049	08:25
Air Blank	0.000	08:26
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

FDLE
Intoxilyzer - Alcohol
Model 8000
05/26/2022
Software: 8100.27

SN 80-000229

Test	g/2	Time
Air Blank	0.01	08:33
Control Test	0.21	08:34
Air Blank	0.01	08:35
Control Test	0.21	08:35
Air Blank	0.01	08:36
Control Test	0.21	08:37
Air Blank	0.01	08:37
Control Test Stats		
Average	0.21	
Std Dev	0.01	
Rel Std Dev(%)	0.21	

Repeated due to
printer issue. JD

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000
05/26/2022
Software: 8100.27

SN 80-000229

Test	g/210L	Time
Air Blank	0.000	08:38
Control Test	0.201	08:39
Air Blank	0.000	08:39
Control Test	0.200	08:40
Air Blank	0.000	08:41
Control Test	0.201	08:41
Air Blank	0.000	08:42
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000
05/26/2022
Software: 8100.27

SN 80-000229

Test	g/210L	Time
Air Blank	0.000	08:43
Control Test	0.078	08:43
Air Blank	0.000	08:44
Control Test	0.078	08:44
Air Blank	0.000	08:45
Control Test	0.079	08:45
Air Blank	0.000	08:45
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

DGS

JD

Operator's Signature

JD

Operator's Signature

JD

Operator's Signature

JD

Operator's Signature

JD

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000229 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000229</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/26/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:45</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/26/2022 Date

THOMAS J GRAHAM,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Serial Number: 80-000229

Time of Inspection: 10:49

Date of Inspection: 05/26/2022

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK		No	Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK		No	Barometric Pressure Sensor Check: OK		No
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

ENTER THE DEPARTMENT INSPECTION BY ACCIDENT

Not a compliance
Check 5/26/22
JG

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JG

THOMAS J GRAHAM

Signature and Printed Name

05/26/2022
Date



Agency Florida Department of Law Enforcement

S/N 80-000229

Date In 7/6/2022DI Completion Date 07-07-2022☐Ship ☐P/U ☐H/D ☐CMI ☐EE

Calibration Adjustment

By _____

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service: FDLE Loaner Instrument.
Will remain at ATP until requested for by an agency. IS

Department Inspection

By IS

Attachments

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☒ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use
- ☒ Conduct an Agency Inspection Before Evidentiary Use

Tech Review / DateAdmin Review / Date 04'00'

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 12:25

Date of Inspection: 07/07/2022

Serial Number: 80-000229

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.200	0.079
0.000	0.048	0.078	0.200	0.078
0.000	0.048	0.079	0.200	0.079
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.200	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.078


Standard Deviations	0.0000	0.0003	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISRAEL SOTO

Signature and Printed Name

07/07/2022

Date

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	80-000229	Florida Department of Law Enforcement	7/7/2022	PN


0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
FDLE Intoxilyzer - Alcohol Analyzer Model: 8000 07/07/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:55 ✓ Control Test 0.048 09:56 ✓ Air Blank 0.000 09:56 ✓ Control Test 0.047 09:57 ✓ Air Blank 0.000 09:58 ✓ Control Test 0.047 09:58 ✓ Air Blank 0.000 09:59 Control Test Stats Average 0.0473 Std Dev 0.0006 Rel Std Dev(%) 1.2198	FDLE Intoxilyzer - Alcohol Analyzer Model: 8000 07/07/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:00 Control Test 0.078 10:01 Air Blank 0.000 10:02 Control Test 0.078 10:02 Air Blank 0.000 10:03 Control Test 0.078 10:03 Air Blank 0.000 10:04 Control Test Stats Average 0.078 Std Dev 0.01 Rel Std Dev(%) 0.01 P.N 7/7/2022	FDLE Intoxilyzer - Alcohol Analyzer Model: 8000 07/07/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:06 Control Test 0.200 10:07 Air Blank 0.000 10:07 Control Test 0.200 10:08 Air Blank 0.000 10:09 Control Test 0.200 10:09 Air Blank 0.000 10:10 Control Test Stats Average 0.2000 Std Dev 0.0000 Rel Std Dev(%) 0.00	FDLE Intoxilyzer - Alcohol Analyzer Model: 8000 07/07/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:17 Control Test 0.000 10:17 Air Blank 0.000 10:18 Control Test 0.000 10:18 Air Blank 0.000 10:18 Control Test 0.000 10:18 Air Blank 0.000 10:19 Control Test 0.000 10:19 Control Test Stats Average 0.0000 Std Dev 0.0000 Rel Std Dev(%) 0.00 DGS
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

Comments: Instrument Printer produced faded printouts during some stability tests. Tests (0.20 & 0.08 DGS) were repeated and printed on external printer (i.e., see attachments).

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000229
07/07/2022
Software: 8100.27

Test	g/210L	Time

Air Blank	0.000	10:11
Control Test	0.200 *	10:12
Air Blank	0.000	10:13
Control Test	0.200 ,	10:13
Air Blank	0.000	10:14
Control Test	0.200 ,	10:15
Air Blank	0.000	10:15
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	




Operator's Signature


FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000229
07/07/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:20
Control Test	0.079	10:21
Air Blank	0.000	10:21
Control Test	0.079	10:22
Air Blank	0.000	10:22
Control Test	0.079	10:22
Air Blank	0.000	10:23
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

D65



Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000229, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000229</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/07/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:25</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/07/2022

Date

Israel Soto

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 09:21

Date of Inspection: 07/07/2022

Serial Number: 80-000229

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Compliance NOT Determined P.N 7/7/2022

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO

Signature and Printed Name

07/07/2022

Date