



Agency Florida Highway Patrol Jackson County Jail

S/N 80-000759

Date In 1/12/2022

DI Completion Date 01-12-2022

Sh

P/U

☐ H/D

Intake		By PN	
☐ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☐ Handle ☒ Keyboard ☐ Dry Gas Shelf ☒ Feet ☐ Breath Tube ☒ Ports ☐ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:			
Quality Checks		By PN	
		Date 1/12/2022	
☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value 190 ☒ Flow Verification (L/s) Flow Column # ATP105 32 mm .175 (.139 - .169) 36 mm .191 (.156 - .190) 53 mm .257 (.228 - .278) 103 mm .535 (.447 - .547)			
☒ Barometric Pressure Check Gauge ID # 28421 ☒ Stability Checks			
Simulator	Serial #	Lot #/Exp	
0.050	MP5088	202010A 10/05/2022	
0.080	MP5089	202010B 10/05/2022	
0.200	MP5090	202010D 10/06/2022	
0.080 DGS	N/A	08121080A1 5/5/2023	
Flow Calibration		By IS	
		Date 1/12/2022	
Flow Column # ATP102 ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value 192 ☒ Post Calibration Verification (L/s) Flow Column # ATP105 32 mm .148 (.139 - .169) 36 mm .160 (.156 - .190) 53 mm .242 (.228 - .278) 103 mm .531 (.447 - .547)			
Maintenance		By	
☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other			

Calibration Adjustment By <u>IS</u>				Department Inspection By <u>IS</u>			
Barometric Pressure Gauge <u>1025</u> ID # <u>30793</u>				Barometric Pressure ID# <u>28421</u>			
Simulator	Serial #	Lot #	Expiration	Gauge <u>1023</u> Instrument <u>1022</u>			
0.000	MP5091	N/A	N/A	Mouth Alcohol Solution Lot # <u>2021-D</u>			
0.040	MP5082	21070	3/1/2023	Acetone Stock Solution Lot # <u>2021-C</u>			
0.100	MP5083	21080	3/8/2023				
0.200	MP5084	20510	12/03/2022				
0.300	MP5085	21030	2/2/2023				
0.080 DGS	N/A	AG113403	05-14-2023				
☒ Post Calibration Adjustment Stability Checks							
Simulator	Serial #	Lot #	Expiration	Attachments			
0.050	MP5088	202010A	10-05-2022	☒ Form 41			☒ Post-Stability Checks
0.080	MP5089	202010B	10-05-2022	☒ Stability Checks			☒ Flow Calibration
0.200	MP5090	202010D	10-06-2022	☒ Calibration Certificate			☐ Form 40
0.080 DGS	N/A	08121080A1	05-05-2023	☒ Calibration Adjustment			☐ Other _____
Notes/Suggested Service: _____				☒ Instrument Complies with Chapter 11D-8, FAC			
Average between wet and dry 0.080 g/210L stability checks more than 0.003, performed Optical Bench Calibration Adjust. IS 01-12-2022				☐ Instrument Does Not Comply with Chapter 11D-8, FAC			
Tech Review: Corrected dry gas expiration date under Post Calibration Adjustment Stability Checks IS 01-12-2022				☒ Return to/Place into Evidentiary Use			
				☐ Remain Out of Evidentiary Use			
				☒ Conduct an Agency Inspection Before Evidentiary Use			
				2022.01.			
				Richard A Williams			
				13			
				10:58:12			
				Tech Review / Date			
				Admin Review / Date			

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 14:02

Date of Inspection: 01/12/2022

Serial Number: 80-000759
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08121080A1 Exp: 05/05/2023
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.079


Standard Deviations	0.0004	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 ISRAEL SOTO
Signature and Printed Name

01/12/2022
Date

STABILITY CHECKS

86-000759

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
01/12/2022
Software: 8100.27

SN 80-000759
Model 8000
01/12/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:16
Control Test	0.048	09:17
Air Blank	0.000	09:18
Control Test	0.048	09:19
Air Blank	0.000	09:19
Control Test	0.048	09:20
Air Blank	0.000	09:20
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
01/12/2022
Software: 8100.27

SN 80-000759
Model 8000
01/12/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:28
Control Test	0.197	09:29
Air Blank	0.000	09:29
Control Test	0.197	09:30
Air Blank	0.000	09:30
Control Test	0.198	09:31
Air Blank	0.000	09:32
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	



Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
01/12/2022
Software: 8100.27

SN 80-000759
Model 8000
01/12/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:57
Control Test	0.082	09:58
Air Blank	0.000	09:58
Control Test	0.082	09:59
Air Blank	0.000	09:59
Control Test	0.081	09:59
Air Blank	0.000	09:59
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

D65


Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000759, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000759</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/12/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:02</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. Digitally signed by Israel Soto
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program. Date: 2022.01.12 15:19:10 -05'00'

Israel Soto

01/12/2022

Date

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 January 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Optical Bench Cal. Adj.

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN BU-000755
01/12/2022 10:35:20

Auto Calibration
Max Power Res Value = 73
Auto Range Res Value = 52

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12515, 9um Io = 13150

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.1530 (-0.0240)
Sample #2 = 0.1170 (-0.0020)
Sample #3 = 0.0970 (-0.0300)
Sample #4 = 0.0610 (-0.0740)
Avg % Abs = 0.0917 (-0.0353)
STD DEV = 0.0004 (-0.0000)
REL STD DEV = 30.958 (102.722)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 0.1200 (-0.0290)
Sample #2 = 0.1100 (-0.0150)
Sample #3 = 0.1290 (-0.0210)
Sample #4 = 0.1010 (-0.0080)
Avg % Abs = 0.1133 (-0.0147)
STD DEV = 0.0143 (-0.0065)
REL STD DEV = 12.613 (44.362)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12503, 9um Io = 13150

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.8250 (-0.0260)
Sample #2 = 0.7970 (-0.0180)
Sample #3 = 0.8130 (-0.0100)
Sample #4 = 0.8020 (-0.0450)
Avg % Abs = 0.8040 (-0.0243)
STD DEV = 0.0082 (-0.0183)
REL STD DEV = 1.018 (75.367)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.4810 (-0.0010)
Sample #2 = 1.4360 (-0.0280)
Sample #3 = 1.4740 (-0.0080)
Sample #4 = 1.4830 (-0.0150)
Avg % Abs = 1.4643 (-0.0170)
STD DEV = 0.0249 (-0.0101)
REL STD DEV = 1.704 (59.699)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12499, 9um Io = 13147

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 1.8900 (-0.0180)
Sample #2 = 1.8830 (-0.0030)
Sample #3 = 1.8850 (-0.0110)
Sample #4 = 1.8850 (-0.0400)
Avg % Abs = 1.8843 (-0.0087)
STD DEV = 0.0012 (-0.0274)
REL STD DEV = 0.061 (316.485)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 3.4210 (-0.0010)
Sample #2 = 3.4280 (-0.0020)
Sample #3 = 3.4370 (-0.0070)
Sample #4 = 3.4230 (-0.0220)
Avg % Abs = 3.4293 (-0.0057)
STD DEV = 0.0071 (-0.0148)
REL STD DEV = 0.207 (251.946)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12492, 9um Io = 13143

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 3.6510 (-0.0250)
Sample #2 = 3.5920 (-0.0040)
Sample #3 = 3.6330 (-0.0310)
Sample #4 = 3.6220 (-0.0280)
Avg % Abs = 3.6157 (-0.0183)
STD DEV = 0.0212 (-0.0194)
REL STD DEV = 0.587 (105.814)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 6.6120 (-0.0230)
Sample #2 = 6.5770 (-0.0040)
Sample #3 = 6.6010 (-0.0200)
Sample #4 = 6.5830 (-0.0230)
Avg % Abs = 6.5870 (-0.0157)
STD DEV = 0.0125 (-0.0102)
REL STD DEV = 0.190 (65.198)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12493, 9um Io = 13145

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 5.3470 (-0.0170)
Sample #2 = 5.2770 (-0.0150)
Sample #3 = 5.2920 (-0.0460)
Sample #4 = 5.2960 (-0.0520)
Avg % Abs = 5.2883 (-0.0377)
STD DEV = 0.0100 (-0.0199)
REL STD DEV = 0.189 (52.720)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 9.6380 (-0.0230)
Sample #2 = 9.5840 (-0.0010)
Sample #3 = 9.5780 (-0.0240)
Sample #4 = 9.5870 (-0.0360)
Avg % Abs = 9.5830 (-0.0203)
STD DEV = 0.0046 (-0.0178)
REL STD DEV = 0.048 (87.471)

***** AUTO CAL DATA *****

Channel 1
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.092
Std Dev = 0.03 Rel Std Dev = 30.96
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.804
Std Dev = 0.01 Rel Std Dev = 1.02
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.884
Std Dev = 0.00 Rel Std Dev = 0.06
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.616
Std Dev = 0.02 Rel Std Dev = 0.59
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.288
Std Dev = 0.01 Rel Std Dev = 0.19
Zero Order Coef = -225.07
First Order Coef = 2598.78
Second Order Coef = 27.34
Standard Deviation = 13.950491

Channel 2
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.113
Std Dev = 0.01 Rel Std Dev = 12.61
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.464
Std Dev = 0.02 Rel Std Dev = 1.70
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.429
Std Dev = 0.01 Rel Std Dev = 0.21
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.587
Std Dev = 0.01 Rel Std Dev = 0.19
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.583
Std Dev = 0.00 Rel Std Dev = 0.05
Zero Order Coef = -156.52
First Order Coef = 1390.76
Second Order Coef = 12.11
Standard Deviation = 4.729115

Optical Bench Cal. Adj.

Post Stability Checks

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.040	0.0005
0.100	0.100	-0.0001
0.200	0.200	-0.0001
0.300	0.300	0.0001

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0000
0.040	0.040	-0.0000
0.100	0.100	0.0001
0.200	0.200	-0.0001
0.300	0.300	0.0001

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 ng/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1
 Sample #1 = 3179.00
 Sample #2 = 3307.00
 Sample #3 = 3228.00
 Sample #4 = 3338.00
 Average Result = 3291.0000
 STD DEV = 56.7166
 REL STD DEV = 1.723

 ***** CHANNEL 2
 Sample #1 = 3337.00
 Sample #2 = 3336.00
 Sample #3 = 3345.00
 Sample #4 = 3387.00
 Average Result = 3356.0000
 STD DEV = 27.2213
 REL STD DEV = 0.811

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1024
 3 un H2O Adjust (ng/l*10,000) = 518
 9 un H2O Adjust (ng/l*10,000) = 453
 **** AUTO CAL PASS

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000759
 01/12/2022
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:06
Control Test	0.050	12:06
Air Blank	0.000	12:07
Control Test	0.050	12:08
Air Blank	0.000	12:08
Control Test	0.049	12:09
Air Blank	0.000	12:09
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000759
 01/12/2022
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:15
Control Test	0.198	12:15
Air Blank	0.000	12:16
Control Test	0.198	12:16
Air Blank	0.000	12:17
Control Test	0.198	12:18
Air Blank	0.000	12:18
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000759
 01/12/2022
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:10
Control Test	0.078	12:11
Air Blank	0.000	12:11
Control Test	0.079	12:12
Air Blank	0.000	12:12
Control Test	0.079	12:13
Air Blank	0.000	12:14
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000759
 01/12/2022
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:19
Control Test	0.080	12:19
Air Blank	0.000	12:20
Control Test	0.080	12:20
Air Blank	0.000	12:21
Control Test	0.080	12:21
Air Blank	0.000	12:21
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

wet

Dry

FLOW CALIBRATION

80-000759

INTOXILYZER 8000
Instrument Initialization
07:49 01/12/2022

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000759
01/12/2022
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
SORT(Diff) = 7.348
2: Rate (Liters/min) = 15
SORT(Diff) = 11.703
3: Rate (Liters/min) = 30
SORT(Diff) = 20.660
Dependent Data Scale Factor = 10000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 722
Rounded Intercept = -779030
Correlation = 0.99664

pt dx 1-12-2022



INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-000759Florida Department of
Law EnforcementDate In 04-19-2022 DI Completion Date 04-21-2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By IS	Quality Checks	By IS	Date <u>04-21-2022</u>	Flow Calibration	By	Date																																						
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Agency Inspector</u> <u>reported that instrument lost</u> <u>forms.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>190</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.253</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks			☐ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																								
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C 01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202201D 01-18-2024</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E 01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG113403 05-14-2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202201D 01-18-2024	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	AG113403 05-14-2023																												
Simulator	Serial #	Lot #/Exp																																											
0.050	MP6291	202201C 01-11-2024																																											
0.080	MP6292	202201D 01-18-2024																																											
0.200	MP6293	202201E 01-18-2024																																											
0.080 DGS	N/A	AG113403 05-14-2023																																											
Calibration Adjustment		By _____		Department Inspection By IS _____																																									
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>26932</u> Gauge <u>1024</u> Instrument <u>1026</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>																																									
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	MP6289																																												
Interferent	MP6290																																												
0.050	MP6291																																												
0.080	MP6292																																												
0.200	MP6293																																												
☐ Post Calibration Adjustment Stability Checks				Attachments																																									
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____																					
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Notes/Suggested Service: <u>Forms uploaded prior to</u> <u>Quality Checks and Department Inspection. IS</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																									
				Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2022.04.21 12:55:04 -0400																																									
				22 14:34:45 04/00																																									
				Tech Review / Date _____ Admin Review / Date _____																																									

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 10:19

Date of Inspection: 04/21/2022

Serial Number: 80-000759
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.080
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.078	0.200	0.079
0.000	0.050	0.079	0.200	0.079

Standard Deviations	0.0004	0.0003	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

04/21/2022
Date

Stability Checks

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000759
04/21/2022
Software: 8100.27

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000759
04/21/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:30
Control Test	0.049	07:30
Air Blank	0.000	07:31
Control Test	0.048	07:32
Air Blank	0.000	07:32
Control Test	0.049	07:33
Air Blank	0.000	07:33
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Test	g/210L	Time
Air Blank	0.000	07:35
Control Test	0.078	07:35
Air Blank	0.000	07:36
Control Test	0.079	07:37
Air Blank	0.000	07:37
Control Test	0.079	07:38
Air Blank	0.000	07:38
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

wet

External printer was
left connected

Operator's Signature

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000759
04/21/2022
Software: 8100.27

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000759
04/21/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:41
Control Test	0.200	07:41
Air Blank	0.000	07:42
Control Test	0.200	07:43
Air Blank	0.000	07:43
Control Test	0.200	07:44
Air Blank	0.000	07:44
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	07:51
Control Test	0.080	07:51
Air Blank	0.000	07:51
Control Test	0.080	07:52
Air Blank	0.000	07:52
Control Test	0.079	07:53
Air Blank	0.000	07:53
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Dry

Operator's Signature

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000759, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000759</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>04/21/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>10:19</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel
Date: 2022.04.21 12:16:18
-04'00'

04/21/2022

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 07:04

Date of Inspection: 04/21/2022

Serial Number: 80-000759
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

CHECKING FORMS UPLOAD

N/A Compliance not determined

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

04/21/2022

Date



INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-000759Florida Department of
Law EnforcementDate In 11-16-2022 DI Completion Date 11-17-2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By IS	Quality Checks	By IS	Date <u>11-17-2022</u>	Flow Calibration	By	Date																																							
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Agency Inspector</u> <u>indicated that instrument lost</u> <u>forms.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>177</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.535</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks			☐ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C 01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202201D 01-18-2024</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E 01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG113403 05-14-2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202201D 01-18-2024	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	AG113403 05-14-2023																													
Simulator	Serial #	Lot #/Exp																																												
0.050	MP6291	202201C 01-11-2024																																												
0.080	MP6292	202201D 01-18-2024																																												
0.200	MP6293	202201E 01-18-2024																																												
0.080 DGS	N/A	AG113403 05-14-2023																																												
Calibration Adjustment		By _____		Department Inspection			By IS _____																																							
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28427</u> Gauge <u>1021</u> Instrument <u>1024</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP6289																																													
Interferent	MP6290																																													
0.050	MP6291																																													
0.080	MP6292																																													
0.200	MP6293																																													
☐ Post Calibration Adjustment Stability Checks				Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					<table border="1"><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></tbody></table>			☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other _____												
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
☒ Form 41	☐ Post-Stability Checks																																													
☒ Stability Checks	☐ Flow Calibration																																													
☒ Calibration Certificate	☐ Form 40																																													
☐ Calibration Adjustment	☐ Other _____																																													
Notes/Suggested Service: <u>Forms re-uploaded by CMI</u> <u>prior to Quality Checks and Department Inspection. IS</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																										
				Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2022.11.17 11:19:12 -0500																																										
				18 15:42:49 Tech Review / Date Admin Review / Date																																										

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 08:42

Date of Inspection: 11/17/2022

Serial Number: 80-000759
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.079	0.200	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.079	0.201	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.200	0.080


Standard Deviations	0.0004	0.0000	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


ISRAEL SOTO
Signature and Printed Name

11/17/2022
Date

Stability Checks

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000759
11/17/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:58
Control Test	0.049	06:59
Air Blank	0.000	07:00
Control Test	0.049	07:00
Air Blank	0.000	07:01
Control Test	0.050	07:01
Air Blank	0.000	07:02
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000759
11/17/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:03
Control Test	0.078	07:03
Air Blank	0.000	07:04
Control Test	0.079	07:05
Air Blank	0.000	07:05
Control Test	0.079	07:06
Air Blank	0.000	07:06
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

wet

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000759
11/17/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:07
Control Test	0.199	07:08
Air Blank	0.000	07:08
Control Test	0.199	07:09
Air Blank	0.000	07:10
Control Test	0.200	07:10
Air Blank	0.000	07:11
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000759
11/17/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:12
Control Test	0.078	07:12
Air Blank	0.000	07:12
Control Test	0.079	07:13
Air Blank	0.000	07:13
Control Test	0.080	07:13
Air Blank	0.000	07:14
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

Dry

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000759, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000759</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>11/17/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>08:42</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Date: 2022.11.17 10:11:39
-05'00'

Date

ISRAEL SOTO,

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Department Inspector

Service • Integrity • Respect • Quality