



INSTRUMENT PROCESSING SHEET

Agency Columbia County Sheriff's Office

S/N 80-000777

Florida Department of Law Enforcement

Date In 12/19/2022

DI Completion Date 12/19/2022

Ship P/U H/D CMI EE

Intake, Quality Checks, Flow Calibration, Maintenance, and Notes sections with various checkboxes and data entry fields.

Calibration Adjustment and Department Inspection sections with tables for simulator data and inspection results.

Notes/Suggested Service, compliance checkboxes, and digital signatures of David Eliezer Reyes-Rivera and Israel Soto.

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	80-000777	Columbia County Sheriff's Office	12/19/2022	PN

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤0.003 of Wet
-------------------------------------	-------------------------------------	-------------------------------------	--

COLUMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000777
12/19/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:16
Control Test	0.050	09:17
Air Blank	0.000	09:17
Control Test	0.050	09:18
Air Blank	0.000	09:18
Control Test	0.051	09:19
Air Blank	0.000	09:20
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	


Operator's Signature

COLUMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000777
12/19/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:29
Control Test	0.078	09:30
Air Blank	0.000	09:30
Control Test	0.078	09:31
Air Blank	0.000	09:32
Control Test	0.077	09:32
Air Blank	0.000	09:33
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	


Operator's Signature

COLUMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000777
12/19/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:34
Control Test	0.200	09:35
Air Blank	0.000	09:35
Control Test	0.200	09:36
Air Blank	0.000	09:36
Control Test	0.200	09:37
Air Blank	0.000	09:38
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

COLUMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000777
12/19/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:40
Control Test	0.079	09:41
Air Blank	0.000	09:41
Control Test	0.078	09:41
Air Blank	0.000	09:42
Control Test	0.079	09:42
Air Blank	0.000	09:43
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	


Operator's Signature

Comments: N/A

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLUMBIA COUNTY SO
Time of Inspection: 11:28

Date of Inspection: 12/19/2022

Serial Number: 80-000777
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.078	0.201	0.079
0.000	0.050	0.078	0.202	0.079
0.000	0.050	0.078	0.201	0.079
0.000	0.050	0.078	0.201	0.079
0.000	0.049	0.078	0.201	0.080
0.000	0.049	0.078	0.202	0.080
0.000	0.050	0.078	0.201	0.079
0.000	0.049	0.078	0.201	0.080
0.000	0.050	0.078	0.200	0.079

Standard Deviations	0.0005	0.0000	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO

Signature and Printed Name

12/19/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000777 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000777</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>COLUMBIA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/19/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:28</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

12/19/2022

Date

Phil Nicodemo

PHIL NICODEMO,
Department Inspector

Digitally signed by Phil
Nicodemo
Date: 2022.12.19 11:56:39 -05'00'

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality