



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000794 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000794</u>		UNCERTAINTY* ±
Owning Agency:	<u>ST LUCIE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/01/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:40</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/01/2022

Date



DAVID E REYES-RIVERA,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ST LUCIE COUNTY SO
Time of Inspection: 08:40

Date of Inspection: 11/01/2022

Serial Number: 80-000794
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.049	0.079	0.203	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.080	0.203	0.080
0.000	0.050	0.081	0.202	0.080


Standard Deviations	0.0003	0.0006	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

DAVID E REYES-RIVERA

11/01/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000794	St. Lucie County Sheriff's Office	11/01/2022	DERR <i>[Signature]</i>

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
ST LUCIE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000794 11/01/2022 Software: 8100.27	ST LUCIE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000794 11/01/2022 Software: 8100.27	ST LUCIE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000794 11/01/2022 Software: 8100.27	ST LUCIE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000794 11/01/2022 Software: 8100.27																																																																																																																																																
<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:46</td></tr> <tr><td>Control Test</td><td>0.051</td><td>06:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:47</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:49</td></tr> <tr><td>Control Test</td><td>0.048</td><td>06:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:50</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>3.0963</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:46	Control Test	0.051	06:47	Air Blank	0.000	06:47	Control Test	0.049	06:48	Air Blank	0.000	06:49	Control Test	0.048	06:49	Air Blank	0.000	06:50	Control Test Stats			Average	0.0493		Std Dev	0.0015		Rel Std Dev(%)	3.0963		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:51</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:52</td></tr> <tr><td>Control Test</td><td>0.078</td><td>06:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:54</td></tr> <tr><td>Control Test</td><td>0.078</td><td>06:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:55</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:51	Control Test	0.079	06:52	Air Blank	0.000	06:52	Control Test	0.078	06:53	Air Blank	0.000	06:54	Control Test	0.078	06:54	Air Blank	0.000	06:55	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:56</td></tr> <tr><td>Control Test</td><td>0.203</td><td>06:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:57</td></tr> <tr><td>Control Test</td><td>0.200</td><td>06:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:59</td></tr> <tr><td>Control Test</td><td>0.201</td><td>06:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:00</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.2013</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7587</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:56	Control Test	0.203	06:57	Air Blank	0.000	06:57	Control Test	0.200	06:58	Air Blank	0.000	06:59	Control Test	0.201	06:59	Air Blank	0.000	07:00	Control Test Stats			Average	0.2013		Std Dev	0.0015		Rel Std Dev(%)	0.7587		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:01</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:02</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:03</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:03</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	07:01	Control Test	0.079	07:01	Air Blank	0.000	07:02	Control Test	0.080	07:02	Air Blank	0.000	07:03	Control Test	0.080	07:03	Air Blank	0.000	07:03	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:46																																																																																																																																																	
Control Test	0.051	06:47																																																																																																																																																	
Air Blank	0.000	06:47																																																																																																																																																	
Control Test	0.049	06:48																																																																																																																																																	
Air Blank	0.000	06:49																																																																																																																																																	
Control Test	0.048	06:49																																																																																																																																																	
Air Blank	0.000	06:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	3.0963																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:51																																																																																																																																																	
Control Test	0.079	06:52																																																																																																																																																	
Air Blank	0.000	06:52																																																																																																																																																	
Control Test	0.078	06:53																																																																																																																																																	
Air Blank	0.000	06:54																																																																																																																																																	
Control Test	0.078	06:54																																																																																																																																																	
Air Blank	0.000	06:55																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:56																																																																																																																																																	
Control Test	0.203	06:57																																																																																																																																																	
Air Blank	0.000	06:57																																																																																																																																																	
Control Test	0.200	06:58																																																																																																																																																	
Air Blank	0.000	06:59																																																																																																																																																	
Control Test	0.201	06:59																																																																																																																																																	
Air Blank	0.000	07:00																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2013																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	0.7587																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:01																																																																																																																																																	
Control Test	0.079	07:01																																																																																																																																																	
Air Blank	0.000	07:02																																																																																																																																																	
Control Test	0.080	07:02																																																																																																																																																	
Air Blank	0.000	07:03																																																																																																																																																	
Control Test	0.080	07:03																																																																																																																																																	
Air Blank	0.000	07:03																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature																																																																																																																																																