



## INSTRUMENT PROCESSING SHEET

Agency Dixie County Sheriff's OfficeS/N 80-000810Florida Department of  
Law EnforcementDate In 12/06/2022 DI Completion Date 12/08/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By PN	Quality Checks	By PN	Date <u>12/7/2022</u>	Flow Calibration	By	Date																																								
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: <u>Missing a single foot on the back</u> <u>left side of the dry gas shelf.</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>142</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>.156</u> (.139 - .169) 36 mm <u>.164</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28662</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5089</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5084</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803 10/25/2024</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202201C 01/11/2024	0.080	MP5089	202201D 01/18/2024	0.200	MP5084	202201E 01/18/2024	0.080 DGS	N/A	AG229803 10/25/2024	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By</th></tr></thead><tbody><tr><td><input type="checkbox"/> Battery Replacement</td><td></td><td></td></tr><tr><td><input type="checkbox"/> Dry Gas Regulator Replacement</td><td></td><td></td></tr><tr><td><input type="checkbox"/> Breath Tube Replacement</td><td></td><td></td></tr><tr><td><input type="checkbox"/> Other _____</td><td></td><td></td></tr></tbody></table>			Maintenance		By	<input type="checkbox"/> Battery Replacement			<input type="checkbox"/> Dry Gas Regulator Replacement			<input type="checkbox"/> Breath Tube Replacement			<input type="checkbox"/> Other _____												
Simulator	Serial #	Lot #/Exp																																													
0.050	MP5088	202201C 01/11/2024																																													
0.080	MP5089	202201D 01/18/2024																																													
0.200	MP5084	202201E 01/18/2024																																													
0.080 DGS	N/A	AG229803 10/25/2024																																													
Maintenance		By																																													
<input type="checkbox"/> Battery Replacement																																															
<input type="checkbox"/> Dry Gas Regulator Replacement																																															
<input type="checkbox"/> Breath Tube Replacement																																															
<input type="checkbox"/> Other _____																																															
<b>Calibration Adjustment</b>		<b>By PN</b>			<b>Department Inspection</b>																																										
Barometric Pressure Gauge <u>1020</u> ID # <u>28427</u>					Barometric Pressure ID# <u>28662</u> Gauge <u>1018</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>21410</td><td>09/30/2023</td></tr><tr><td>0.100</td><td>MP6296</td><td>22310</td><td>08/11/2024</td></tr><tr><td>0.200</td><td>MP6297</td><td>22050</td><td>02/07/2024</td></tr><tr><td>0.300</td><td>MP5085</td><td>22220</td><td>06/15/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>08121080A1</td><td>05/05/2023</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	21410	09/30/2023	0.100	MP6296	22310	08/11/2024	0.200	MP6297	22050	02/07/2024	0.300	MP5085	22220	06/15/2024	0.080 DGS	N/A	08121080A1	05/05/2023				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5084</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5084
Simulator	Serial #	Lot #	Expiration																																												
0.000	MP6294	N/A	N/A																																												
0.040	MP6295	21410	09/30/2023																																												
0.100	MP6296	22310	08/11/2024																																												
0.200	MP6297	22050	02/07/2024																																												
0.300	MP5085	22220	06/15/2024																																												
0.080 DGS	N/A	08121080A1	05/05/2023																																												
Simulator	Serial Number																																														
0.000	MP5086																																														
Interferent	MP5087																																														
0.050	MP5088																																														
0.080	MP5089																																														
0.200	MP5084																																														
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks					<b>Attachments</b>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202201C</td><td>01/11/2024</td></tr><tr><td>0.080</td><td>MP5089</td><td>202201D</td><td>01/18/2024</td></tr><tr><td>0.200</td><td>MP5084</td><td>202201E</td><td>01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803</td><td>10/25/2024</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202201C	01/11/2024	0.080	MP5089	202201D	01/18/2024	0.200	MP5084	202201E	01/18/2024	0.080 DGS	N/A	AG229803	10/25/2024				<table border="1"><tbody><tr><td><input checked="" type="checkbox"/> Form 41</td><td><input checked="" type="checkbox"/> Post-Stability Checks</td></tr><tr><td><input checked="" type="checkbox"/> Stability Checks</td><td><input type="checkbox"/> Flow Calibration</td></tr><tr><td><input checked="" type="checkbox"/> Calibration Certificate</td><td><input type="checkbox"/> Form 40</td></tr><tr><td><input checked="" type="checkbox"/> Calibration Adjustment</td><td><input type="checkbox"/> Other _____</td></tr></tbody></table>			<input checked="" type="checkbox"/> Form 41	<input checked="" type="checkbox"/> Post-Stability Checks	<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration	<input checked="" type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40	<input checked="" type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____												
Simulator	Serial #	Lot #	Expiration																																												
0.050	MP5088	202201C	01/11/2024																																												
0.080	MP5089	202201D	01/18/2024																																												
0.200	MP5084	202201E	01/18/2024																																												
0.080 DGS	N/A	AG229803	10/25/2024																																												
<input checked="" type="checkbox"/> Form 41	<input checked="" type="checkbox"/> Post-Stability Checks																																														
<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration																																														
<input checked="" type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40																																														
<input checked="" type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____																																														
Notes/Suggested Service: _____					<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																										
					<table border="1"><tbody><tr><td>Taylor Gutschow</td><td>Digitally signed by Taylor Gutschow Date: 2022.12.09 09:29:48 -05'00'</td><td>Israel Soto</td><td>Digitally signed by Israel Soto Date: 2022.12.09 13:53:45 -05'00'</td></tr></tbody></table>			Taylor Gutschow	Digitally signed by Taylor Gutschow Date: 2022.12.09 09:29:48 -05'00'	Israel Soto	Digitally signed by Israel Soto Date: 2022.12.09 13:53:45 -05'00'																																				
Taylor Gutschow	Digitally signed by Taylor Gutschow Date: 2022.12.09 09:29:48 -05'00'	Israel Soto	Digitally signed by Israel Soto Date: 2022.12.09 13:53:45 -05'00'																																												
					Tech Review / Date _____ Admin Review / Date _____																																										

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DIXIE COUNTY SO  
Time of Inspection: 13:55

Date of Inspection: 12/08/2022

Serial Number: 80-000810  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.048	0.078	0.201	0.079
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.078	0.202	0.079
0.000	0.049	0.078	0.202	0.079
0.000	0.049	0.078	0.201	0.078
0.000	0.050	0.078	0.201	0.079
0.000	0.050	0.079	0.200	0.078
0.000	0.050	0.078	0.201	0.078

Standard Deviations	0.0006	0.0003	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO

Signature and Printed Name

12/08/2022  
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	<del>80-00010</del>	Dixie County Sheriff's Office	12/07/2022	PN

80-000810 PN 12/9/2022

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 ≤0.003 of Wet
DIXIE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000810 12/07/2022 Software: 8100.27	DIXIE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000810 12/07/2022 Software: 8100.27	DIXIE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000810 12/07/2022 Software: 8100.27	DIXIE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000810 12/07/2022 Software: 8100.27
Test g/210L Time	Test g/210L Time	Test g/210L Time	Test g/210L Time
Air Blank 0.000 11:40	Air Blank 0.000 11:44	Air Blank 0.000 11:49	Air Blank 0.000 11:57
Control Test 0.047 11:40	Control Test 0.077 11:45	Control Test 0.199 11:50	Control Test 0.074 11:57
Air Blank 0.000 11:41	Air Blank 0.000 11:46	Air Blank 0.000 11:50	Air Blank 0.000 11:57
Control Test 0.048 11:42	Control Test 0.077 11:46	Control Test 0.197 11:51	Control Test 0.074 11:58
Air Blank 0.000 11:42	Air Blank 0.000 11:47	Air Blank 0.000 11:51	Air Blank 0.000 11:58
Control Test 0.047 11:43	Control Test 0.077 11:48	Control Test 0.198 11:52	Control Test 0.074 11:59
Air Blank 0.000 11:43	Air Blank 0.000 11:48	Air Blank 0.000 11:53	Air Blank 0.000 11:59
Control Test Stats	Control Test Stats	Control Test Stats	Control Test Stats
Average 0.0473	Average 0.0770	Average 0.1980	Average 0.0740
Std Dev 0.0006	Std Dev 0.0000	Std Dev 0.0010	Std Dev 0.0000
Rel Std Dev(%) 1.2198	Rel Std Dev(%) 0.0000	Rel Std Dev(%) 0.5051	Rel Std Dev(%) 0.0000
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

Comments: N/A



## Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road.  
Suite B1032  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000810, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000810</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>DIXIE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/08/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:55</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

12/08/2022

Date

Phil Nicodemo

Digitally signed by Phil  
Nicodemo  
Date: 2022.12.08 15:31:59 -05'00'

PHIL NICODEMO,  
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

DIXIE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000810  
12/08/2022 10:20:53

OPTICAL BENCH CALIBRATION

ADJUSTMENT

P.N 12/8/22

Auto Calibration

pg 1 of 2

<<<< 3um >>>> <<<< 9um >>>>

-----  
Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.0840	(-0.0180)	0.1500	(-0.0150)
Sample #2	0.0670	(0.0450)	0.1300	(0.0220)
Sample #3	0.0600	(0.0860)	0.1610	(0.0240)
Sample #4	0.0750	(0.0770)	0.1540	(0.0240)
Avg % Abs	0.0673	(0.0693)	0.1483	(0.0233)
STD DEV	0.0075	(0.0215)	0.0163	(0.0012)
REL STD DEV	11.147	(31.079)	10.961	(4.949)

-----  
Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.8080	(-0.0190)	1.6160	(-0.0060)
Sample #2	0.7750	(0.0060)	1.6070	(0.0100)
Sample #3	0.7790	(0.0180)	1.6060	(0.0080)
Sample #4	0.7730	(0.0650)	1.5840	(0.0330)
Avg % Abs	0.7757	(0.0297)	1.5990	(0.0170)
STD DEV	0.0031	(0.0312)	0.0130	(0.0139)
REL STD DEV	0.394	(105.109)	0.813	(81.720)

-----  
Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	1.8510	(-0.0120)	3.7260	(0.0000)
Sample #2	1.8320	(0.0250)	3.6870	(0.0460)
Sample #3	1.8170	(0.0590)	3.6800	(0.0550)
Sample #4	1.8200	(0.0670)	3.6630	(0.0850)
Avg % Abs	1.8230	(0.0503)	3.6767	(0.0620)
STD DEV	0.0079	(0.0223)	0.0123	(0.0204)
REL STD DEV	0.435	(44.307)	0.336	(32.936)

-----  
Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	3.5560	(-0.0260)	7.1110	(-0.0320)
Sample #2	3.5110	(0.0010)	6.9940	(0.0500)
Sample #3	3.4980	(0.0540)	6.9910	(0.0670)
Sample #4	3.4790	(0.0510)	6.9800	(0.0730)
Avg % Abs	3.4960	(0.0353)	6.9883	(0.0633)
STD DEV	0.0161	(0.0298)	0.0074	(0.0119)
REL STD DEV	0.460	(84.259)	0.105	(18.837)

-----  
Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.1770	(-0.0360)	10.2440	(-0.0290)
Sample #2	5.1030	(0.0710)	10.1070	(0.1080)
Sample #3	5.1140	(0.0630)	10.0980	(0.1390)
Sample #4	5.1000	(0.0770)	10.1080	(0.1220)
Avg % Abs	5.1057	(0.0703)	10.1043	(0.1230)
STD DEV	0.0074	(0.0070)	0.0055	(0.0155)
REL STD DEV	0.144	(9.986)	0.055	(12.621)

-----

DIXIE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000810  
12/08/2022 10:20:53

OPTICAL BENCH CALIBRATION  
ADJUSTMENT

P.N 12/8/2022

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

-----  
Zero Order Coef -175.86  
First Order Coef 2645.37  
Second Order Coef 36.67  
-----

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0001
0.040	0.040	0.0001
0.100	0.100	-0.0001
0.200	0.200	0.0001
0.300	0.300	-0.0000

-----

<<<< 9um >>>>

-----  
-195.92  
1297.15  
13.46  
-----

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0001
0.040	0.040	-0.0002
0.100	0.100	0.0001
0.200	0.200	-0.0001
0.300	0.300	0.0000

-----

<<<< 3um >>>>

-----  
Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1  
Sample

Sample #1	3340.00	3248.00
Sample #2	3251.00	3275.00
Sample #3	3229.00	3249.00
Sample #4	3416.00	3281.00
Avg	3298.6667	3268.3333
STD DEV	102.2073	17.0098
REL STD DEV	3.098	0.520
H2O adjust (mg/l*10k)	511	541

Barometric Pressure = 1020

\*\*\*\*\*CALIBRATION SUCCESSFUL\*\*\*\*\*

Type of Test	Serial Number	Agency	Date	Performed By
Post Calibration Adjustment Stability Checks	80-000810	Dixie County Sheriff's Office	12/08/2022	PN

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤0.003 of Wet
DIXIE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000810 12/08/2022 Software: 8100.27	DIXIE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000810 12/08/2022 Software: 8100.27	DIXIE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000810 12/08/2022 Software: 8100.27	DIXIE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000810 12/08/2022 Software: 8100.27
Test g/210L Time	Test g/210L Time	Test g/210L Time	Test g/210L Time
Air Blank 0.000 11:57	Air Blank 0.000 12:02	Air Blank 0.000 12:06	Air Blank 0.000 12:11
Control Test 0.049 11:58	Control Test 0.078 12:03	Control Test 0.201 12:07	Control Test 0.080 12:11
Air Blank 0.000 11:58	Air Blank 0.000 12:03	Air Blank 0.000 12:08	Air Blank 0.000 12:12
Control Test 0.049 11:59	Control Test 0.077 12:04	Control Test 0.200 12:08	Control Test 0.079 12:12
Air Blank 0.000 12:00	Air Blank 0.000 12:04	Air Blank 0.000 12:09	Air Blank 0.000 12:13
Control Test 0.049 12:00	Control Test 0.077 12:05	Control Test 0.198 12:10	Control Test 0.080 12:13
Air Blank 0.000 12:01	Air Blank 0.000 12:06	Air Blank 0.000 12:10	Air Blank 0.000 12:14
Control Test Stats	Control Test Stats	Control Test Stats	Control Test Stats
Average 0.0490	Average 0.0773	Average 0.1997	Average 0.0797
Std Dev 0.0000	Std Dev 0.0006	Std Dev 0.0015	Std Dev 0.0006
Rel Std Dev(%) 0.0000	Rel Std Dev(%) 0.7466	Rel Std Dev(%) 0.7650	Rel Std Dev(%) 0.7247
Phil Modeno Operator's Signature	Phil Modeno Operator's Signature	Phil Modeno Operator's Signature	Phil Modeno Operator's Signature

Comments: N/A