



INSTRUMENT PROCESSING SHEET

Agency Hillsborough CSO

S/N 80-000830

Florida Department of
Law Enforcement

Date In 4/20/2022 DI Completion Date 4/22/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 4/22/2022	Flow Calibration	By	Date																												
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value 111 ☒ Flow Verification (L/s) Flow Column # ATP104 32 mm 0.160 (.139 - .169) 36 mm 0.171 (.156 - .190) 53 mm 0.242 (.228 - .278) 103 mm 0.500 (.447 - .547) ☒ Barometric Pressure Check Gauge ID # 68639 ☒ Stability Checks			Flow Column # ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value ☐ Post Calibration Verification (L/s) Flow Column # 32 mm (.139 - .169) 36 mm (.156 - .190) 53 mm (.228 - .278) 103 mm (.447 - .547)																														
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5092</td><td>202010A 10/05/2022</td></tr><tr><td>0.080</td><td>MP5093</td><td>202010B 10/05/2022</td></tr><tr><td>0.200</td><td>MP5094</td><td>202010D 10/06/2022</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202010A 10/05/2022	0.080	MP5093	202010B 10/05/2022	0.200	MP5094	202010D 10/06/2022	0.080 DGS	N/A	AG115904 06/08/2023																		
Simulator	Serial #	Lot #/Exp																																	
0.050	MP5092	202010A 10/05/2022																																	
0.080	MP5093	202010B 10/05/2022																																	
0.200	MP5094	202010D 10/06/2022																																	
0.080 DGS	N/A	AG115904 06/08/2023																																	
Calibration Adjustment		By		Department Inspection			By TDG																												
Barometric Pressure Gauge ID #				Barometric Pressure ID# 28199																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					Gauge 1023 Instrument 1022 Mouth Alcohol Solution Lot # 2021-B Acetone Stock Solution Lot # 2021-C			
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
☐ Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>SD3963</td></tr><tr><td>Interferent</td><td>SD3966</td></tr><tr><td>0.050</td><td>MP5092</td></tr><tr><td>0.080</td><td>MP5093</td></tr><tr><td>0.200</td><td>MP5094</td></tr></tbody></table>			Simulator	Serial Number	0.000	SD3963	Interferent	SD3966	0.050	MP5092	0.080	MP5093	0.200	MP5094			Attachments														
Simulator	Serial Number																																		
0.000	SD3963																																		
Interferent	SD3966																																		
0.050	MP5092																																		
0.080	MP5093																																		
0.200	MP5094																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment			☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other								
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
				Israel Soto Digitally signed by Israel Soto Date: 2022.04.25 10:45:07 -04'00'			26 12:59:22 04/001																												
				Tech Review / Date			Admin. Review / Date																												

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000830	Hillsborough CSD	04/22/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 04/22/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:46 Control Test 0.050 10:47 Air Blank 0.000 10:47 Control Test 0.050 10:48 Air Blank 0.000 10:49 Control Test 0.050 10:49 Air Blank 0.000 10:50 Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 04/22/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:53 Control Test 0.080 10:54 Air Blank 0.000 10:54 Control Test 0.080 10:55 Air Blank 0.000 10:55 Control Test 0.080 10:56 Air Blank 0.000 10:57 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 04/22/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:59 Control Test 0.201 11:00 Air Blank 0.000 11:01 Control Test 0.201 11:01 Air Blank 0.000 11:02 Control Test 0.201 11:03 Air Blank 0.000 11:03 Control Test Stats Average 0.2010 Std Dev 0.0000 Rel Std Dev(%) 0.0000	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 04/22/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:05 Control Test 0.082 11:06 Air Blank 0.000 11:06 Control Test 0.081 11:07 Air Blank 0.000 11:07 Control Test 0.082 11:07 Air Blank 0.000 11:08 Control Test Stats Average 0.0817 Std Dev 0.0006 Rel Std Dev(%) 0.7078
<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 13:26

Date of Inspection: 04/22/2022

Serial Number: 80-000830
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.080	0.200	0.083
0.000	0.049	0.079	0.201	0.082
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.079	0.200	0.082
0.000	0.051	0.080	0.200	0.082


Standard Deviations	0.0005	0.0004	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

04/22/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000830, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-000830	UNCERTAINTY* $\pm$
Owning Agency:	HILLSBOROUGH CO SO	0.050 g/ 210 L 0.004
Calibration Date:	04/22/2022	0.080 g/ 210 L 0.004
Calibration Time:	13:26	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/22/2022

Date



TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality