



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000861, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000861</u>	UNCERTAINTY* ±	
Owning Agency:	<u>SARASOTA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/26/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:01</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/26/2022 Date

DAVID E REYES-RIVERA,
Department Inspector

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 12:01

Date of Inspection: 07/26/2022

Serial Number: 80-000861
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.078	0.202	0.000 / 0.080
0.000	0.049	0.077	0.201	0.000 / 0.080
0.000	0.049	0.078	0.202	0.000 / 0.080
0.000	0.049	0.078	0.202	0.000 / 0.080
0.000	0.049	0.077	0.202	0.000 / 0.079
0.000	0.049	0.078	0.202	0.000 / 0.079
0.000	0.049	0.078	0.202	0.000 / 0.080
0.000	0.049	0.078	0.202	0.000 / 0.079
0.000	0.049	0.077	0.202	0.000 / 0.080
0.000	0.049	0.078	0.202	0.000 / 0.079
Standard Deviations	0.0000	0.0004	0.0003	0.0000 / 0.0005


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance DGS NOT CONNECTED..

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

DAVID E REYES-RIVERA

07/26/2022
 Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-00861	Sarasota County Sheriff's Office	7/26/2022	DERR <i>MLL</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>SARASOTA COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-000861 07/26/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:32 Control Test 0.048 09:33 Air Blank 0.000 09:33 Control Test 0.048 09:34 Air Blank 0.000 09:34 Control Test 0.048 09:35 Air Blank 0.000 09:36</p> <p>Control Test Stats Average 0.0480 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature <i>MLL</i></p>	<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>SARASOTA COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-000861 07/26/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:36 Control Test 0.078 09:36 Air Blank 0.000 09:39 Control Test 0.077 09:40 Air Blank 0.000 09:40 Control Test 0.078 09:41 Air Blank 0.000 09:41</p> <p>Control Test Stats Average 0.0777 Std Dev 0.0006 Rel Std Dev(%) 0.7434</p> <p>Operator's Signature <i>MLL</i></p>	<p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>SARASOTA COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-000861 07/26/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:43 Control Test 0.202 09:44 Air Blank 0.000 09:44 Control Test 0.200 09:45 Air Blank 0.000 09:45 Control Test 0.200 09:46 Air Blank 0.000 09:47</p> <p>Control Test Stats Average 0.2007 Std Dev 0.0012 Rel Std Dev(%) 0.5754</p> <p>Operator's Signature <i>MLL</i></p>	<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>SARASOTA COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-000861 07/26/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:48 Control Test 0.080 09:48 Air Blank 0.000 09:49 Control Test 0.080 09:49 Air Blank 0.000 09:50 Control Test 0.079 09:50 Air Blank 0.000 09:51</p> <p>Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247</p> <p>Operator's Signature <i>MLL</i></p>

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000861
 07/26/2022 08:31:26

Auto Calibration
 Max Power Res Value = 53
 Auto Range Res Value = 44

Sol Value = 0.000 g/210L ***
 Fit Value = 0.0000 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12827, Sum Io = 13364

Channel 1
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.0350 (-0.0120)
 Sample #2 = 0.0610 (-0.0150)
 Sample #3 = 0.0830 (-0.0550)
 Sample #4 = 0.0740 (-0.0280)
 Avg % Abs = 0.0727 (-0.0327)
 STD DEV = 0.0111 (0.0204)
 REL STD DEV = 15.22% (62.462)

Channel 2
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.5340 (-0.0010)
 Sample #2 = 1.5110 (0.0150)
 Sample #3 = 1.5330 (-0.0070)
 Sample #4 = 1.5330 (-0.0070)
 Avg % Abs = 1.5257 (0.0003)
 STD DEV = 0.0127 (0.0127)
 REL STD DEV = 0.833 (3810.515)

Sol Value = 0.100 g/210L ***
 Fit Value = 0.4762 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12825, Sum Io = 13362

Channel 1
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.8030 (-0.0080)
 Sample #2 = 1.8100 (-0.0200)
 Sample #3 = 1.8350 (-0.0150)
 Sample #4 = 1.8230 (-0.0220)
 Avg % Abs = 1.8240 (-0.0130)
 STD DEV = 0.0145 (0.0101)
 REL STD DEV = 0.796 (78.069)

Channel 2
 Sample % Abs (% Abs Ref)
 Sample #1 = 6.9920 (-0.0130)
 Sample #2 = 6.9210 (0.0560)
 Sample #3 = 6.9050 (0.0760)
 Sample #4 = 6.8980 (0.0790)
 Avg % Abs = 6.9080 (0.0703)
 STD DEV = 0.0118 (0.0125)
 REL STD DEV = 0.171 (17.777)

Sol Value = 0.300 g/210L ***
 Fit Value = 1.4286 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12817, Sum Io = 13356

Channel 1
 Sample % Abs (% Abs Ref)
 Sample #1 = 5.1090 (-0.0150)
 Sample #2 = 5.0240 (0.0550)
 Sample #3 = 4.9720 (0.0750)
 Sample #4 = 4.9630 (0.0450)
 Avg % Abs = 4.9853 (0.0583)
 STD DEV = 0.0329 (0.0153)
 REL STD DEV = 0.660 (26.186)

Channel 2
 Sample % Abs (% Abs Ref)
 Sample #1 = 10.0740 (-0.0070)
 Sample #2 = 9.6090 (0.0280)
 Sample #3 = 9.8270 (0.1230)
 Sample #4 = 9.7570 (0.1230)
 Avg % Abs = 9.8210 (0.1207)
 STD DEV = 0.0612 (0.0040)
 REL STD DEV = 0.623 (3.349)

Sol Value = 0.200 g/210L ***
 Fit Value = 0.9524 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12821, Sum Io = 13358

Channel 1
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.5270 (-0.0100)
 Sample #2 = 3.4900 (0.0120)
 Sample #3 = 3.4820 (0.0430)
 Sample #4 = 3.4610 (0.0430)
 Avg % Abs = 3.4777 (0.0327)
 STD DEV = 0.0150 (0.0179)
 REL STD DEV = 0.431 (54.789)

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0008
0.040	0.039	0.0007
0.100	0.099	0.0010
0.200	0.201	-0.0014
0.300	0.299	0.0006

Sol Value = 0.080 g/210L ***
 Fit Value = 0.3910 mg/l %
 Samples Taken = 4, Discarded = 1

Channel 1

Sample #	% Abs	(% Abs Ref)
1	3368.00	3368.00
2	3322.00	3322.00
3	3324.00	3324.00
4	3361.00	3361.00

Average Result = 3239.0000
 STD DEV = 19.9750
 REL STD DEV = 0.617

Channel 2

Sample #	% Abs	(% Abs Ref)
1	3368.00	3368.00
2	3352.00	3352.00
3	3358.00	3358.00
4	3381.00	3381.00

Average Result = 3363.6657
 STD DEV = 15.3080
 REL STD DEV = 0.455

Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1018
 3 um H2O Adjust (mg/l x 0.000) = 570
 9 um H2O Adjust (mg/l x 0.000) = 446
 **** AUTO CAL PASS

Optical Calibration

SN:	80-000861
Agency:	Sarasota CSO
Date:	7/26/2022
Quadratic Fit:	+/- 0.002g/210L
By:	DERR <i>[Signature]</i>

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0006
0.040	0.039	0.0007
0.100	0.100	0.0004
0.200	0.201	-0.0008
0.300	0.300	0.0003

Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12827, Sum Io = 13364

Channel 1

Sample #	% Abs	(% Abs Ref)
1	0.7710	(-0.0080)
2	0.7450	(0.0110)
3	0.7950	(-0.0310)
4	0.7830	(-0.0220)

Avg % Abs = 0.7743 (-0.0140)
 STD DEV = 0.0261 (0.0221)
 REL STD DEV = 3.371 (157.952)

Sol Value = 0.200 g/210L ***
 Fit Value = 0.9524 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12821, Sum Io = 13358

Channel 1

Sample #	% Abs	(% Abs Ref)
1	3.5270	(-0.0100)
2	3.4900	(0.0120)
3	3.4820	(0.0430)
4	3.4610	(0.0430)

Avg % Abs = 3.4777 (0.0327)
 STD DEV = 0.0150 (0.0179)
 REL STD DEV = 0.431 (54.789)

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000861	Sarasota County Sheriff's Office	7/26/2022	DERR 

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<p>SR99S079 COUNTY 50 Intoxilizer - Alcohol Analyzer Model 8000 SN 80-000861 07/26/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 1.000 07:27 Control Test 0.047 07:28 Air Blank 0.000 07:29 Control Test 0.047 07:29 Air Blank 0.000 07:31 Control Test 0.047 07:31 Air Blank 0.000 07:31</p> <p>Control Test Stats Average 0.0470 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p> Operator's Signature</p>	<p>SR99S079 COUNTY 50 Intoxilizer - Alcohol Analyzer Model 8000 SN 80-000861 07/26/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:34 Control Test 0.076 07:35 Air Blank 0.000 07:36 Control Test 0.076 07:36 Air Blank 0.000 07:37 Control Test 0.076 07:37 Air Blank 0.000 07:38</p> <p>Control Test Stats Average 0.0750 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p> Operator's Signature</p>	<p>SR99S079 COUNTY 50 Intoxilizer - Alcohol Analyzer Model 8000 SN 80-000861 07/26/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:41 Control Test 0.198 07:41 Air Blank 0.000 07:42 Control Test 0.197 07:43 Air Blank 0.000 07:43 Control Test 0.197 07:44 Air Blank 0.000 07:45</p> <p>Control Test Stats Average 0.1973 Std Dev 0.0006 Rel Std Dev(%) 0.2926</p> <p> Operator's Signature</p>	<p>SR99S079 COUNTY 50 Intoxilizer - Alcohol Analyzer Model 8000 SN 80-000861 07/26/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:46 Control Test 0.079 07:46 Air Blank 0.000 07:47 Control Test 0.079 07:47 Air Blank 0.000 07:48 Control Test 0.079 07:48 Air Blank 0.000 07:49</p> <p>Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p> Operator's Signature</p>
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>