



Agency North Miami Beach Police Department

S/N 80-000886

Date In 8/19/2022 DI Completion Date 8/30/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

FDLE/ATP Form 48 January 2022
Issuing Authority: Alcohol Testing Program

Page 1 of 1



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000886, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000886</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>NORTH MIAMI BEACH PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>08/30/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>10:16</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

08/30/2022

Date

DAVID E REYES-RIVERA,

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NORTH MIAMI BEACH PD
Time of Inspection: 10:16

Date of Inspection: 08/30/2022

Serial Number: 80-000886
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.079	0.201	0.080
0.000	0.050	0.079	0.201	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.202	0.080

Standard Deviations	0.0005	0.0006	0.0004	0.0004
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

A F / M A: DID NOT PROVIDE MA SAMPLE.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA
Signature and Printed Name

08/30/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000886	North Miami Beach Police Department	8/30/2022	DERR 

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
NORTH MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000886 08/30/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:45 Control Test 0.050 06:45 Air Blank 0.000 06:46 Control Test 0.049 06:47 Air Blank 0.000 06:47 Control Test 0.049 06:48 Air Blank 0.000 06:48 Control Test Stats Average 0.0493 Std Dev 0.0006 Rel Std Dev(%) 1.1703 Operator's Signature 	NORTH MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000886 08/30/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:50 Control Test 0.078 06:50 Air Blank 0.000 06:51 Control Test 0.078 06:52 Air Blank 0.000 06:52 Control Test 0.078 06:53 Air Blank 0.000 06:54 Control Test Stats Average 0.0780 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 	NORTH MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000886 08/30/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:54 Control Test 0.197 06:55 Air Blank 0.000 06:56 Control Test 0.196 06:56 Air Blank 0.000 06:57 Control Test 0.197 06:58 Air Blank 0.000 06:58 Control Test Stats Average 0.1967 Std Dev 0.0006 Rel Std Dev(%) 0.2956 Operator's Signature 	NORTH MIAMI BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000886 08/30/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:59 Control Test 0.079 07:00 Air Blank 0.000 07:00 Control Test 0.080 07:01 Air Blank 0.000 07:01 Control Test 0.079 07:01 Air Blank 0.000 07:02 Control Test Stats Average 0.0793 Std Dev 0.0006 Rel Std Dev(%) 0.7277 Operator's Signature 

INSTRUMENT PROCESSING SHEET

Agency North Miami Beach Police Department

S/N 80-000886

Florida Department of
Law Enforcement

Date In 6/23/2022 DI Completion Date 6/27/2022

☒ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake By <u>DERR</u> ☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ 	Quality Checks By <u>DERR</u> Date <u>6/27/2022</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>138</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP4863</td> <td>202201C</td> </tr> <tr> <td>01/11/2024</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP4864</td> <td>202201D</td> </tr> <tr> <td>01/18/2024</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5097</td> <td>202201E</td> </tr> <tr> <td>01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904</td> </tr> <tr> <td></td> <td></td> <td>06/08/2023</td> </tr> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	202201C	01/11/2024	0.080	MP4864	202201D	01/18/2024	0.200	MP5097	202201E	01/18/2024	0.080 DGS	N/A	AG115904			06/08/2023	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ 																																						
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP4863	202201C																																																											
		01/11/2024																																																											
0.080	MP4864	202201D																																																											
		01/18/2024																																																											
0.200	MP5097	202201E																																																											
		01/18/2024																																																											
0.080 DGS	N/A	AG115904																																																											
		06/08/2023																																																											
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> Notes/Suggested Service: <u>No tone from speaker I will send it back to CMI as a warranty repair. DERR</u> _____ _____ _____ _____ _____ 	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>DERR</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1015</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1017</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use Israel Soto Digitally signed by Israel Soto Date: 2022.06.27 14:12:30 +0400' <u>30</u> 11:03:02 04'00" Tech Review / Date Admin Review / Date	Simulator	Serial Number	0.000	SD1014	Interferent	SD1017	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD1014																																																												
Interferent	SD1017																																																												
0.050	MP4863																																																												
0.080	MP4864																																																												
0.200	MP5097																																																												

Return Material Authorization

Ship to:



CMI, Inc.



Enforcement Electronics

Shipment to repair facility authorized by: David Reyes-Rivera on 6/27/2022

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: 8000 Serial Number: 80-000886

Bill To Address:

Ship to Address:

FDLE

ATTN: David Reyes-Rivera

4700 Terminal Drive Suite 1

Fort Myers, Florida 33907

Reason for Return:

Instrument was received from CMI, while conducting quality checks the instrument did not have an audible tone. Tried Control F5 in the DVM menu but couldn't get the tone to work.

This is a warranty repair.


Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$_____.
- ☐ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: David Reyes

Phone #: 8504043700 Email: davidreyes@fdle.state.fl.us

ATP Contact Name: David Reyes-Rivera ATP Email: davidreyes@fdle.state.fl.us



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000886, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000886</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>NORTH MIAMI BEACH PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/27/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:25</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

06/27/2022

Date

DAVID E REYES-RIVERA,

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NORTH MIAMI BEACH PD
Time of Inspection: 08:25

Date of Inspection: 06/27/2022

Serial Number: 80-000886
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081


Standard Deviations	0.0004	0.0004	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

DAVID E REYES-RIVERA

06/27/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000886	North Miami Beach Police Department	6/27/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
NORTH MIAMI BEACH PD Intoxilyzer - Alcolon Analyzer Model 8000 06/27/2022 Software: 8100.27 SN 80-000886 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>06:32</td></tr><tr><td>Control Test</td><td>0.050</td><td>06:32</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:33</td></tr><tr><td>Control Test</td><td>0.049</td><td>06:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:34</td></tr><tr><td>Control Test</td><td>0.049</td><td>06:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:36</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	06:32	Control Test	0.050	06:32	Air Blank	0.000	06:33	Control Test	0.049	06:34	Air Blank	0.000	06:34	Control Test	0.049	06:35	Air Blank	0.000	06:36	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		NORTH MIAMI BEACH PD Intoxilyzer - Alcolon Analyzer Model 8000 06/27/2022 Software: 8100.27 SN 80-000886 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>06:37</td></tr><tr><td>Control Test</td><td>0.079</td><td>06:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:38</td></tr><tr><td>Control Test</td><td>0.079</td><td>06:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:39</td></tr><tr><td>Control Test</td><td>0.079</td><td>06:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:41</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	06:37	Control Test	0.079	06:38	Air Blank	0.000	06:38	Control Test	0.079	06:39	Air Blank	0.000	06:39	Control Test	0.079	06:40	Air Blank	0.000	06:41	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		NORTH MIAMI BEACH PD Intoxilyzer - Alcolon Analyzer Model 8000 06/27/2022 Software: 8100.27 SN 80-000886 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>06:42</td></tr><tr><td>Control Test</td><td>0.200</td><td>06:42</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:43</td></tr><tr><td>Control Test</td><td>0.199</td><td>06:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:44</td></tr><tr><td>Control Test</td><td>0.198</td><td>06:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:46</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1990</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.5025</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	06:42	Control Test	0.200	06:42	Air Blank	0.000	06:43	Control Test	0.199	06:44	Air Blank	0.000	06:44	Control Test	0.198	06:45	Air Blank	0.000	06:46	Control Test Stats			Average	0.1990		Std Dev	0.0010		Rel Std Dev(%)	0.5025		NORTH MIAMI BEACH PD Intoxilyzer - Alcolon Analyzer Model 8000 06/27/2022 Software: 8100.27 SN 80-000886 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>06:47</td></tr><tr><td>Control Test</td><td>0.080</td><td>06:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:47</td></tr><tr><td>Control Test</td><td>0.080</td><td>06:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:48</td></tr><tr><td>Control Test</td><td>0.080</td><td>06:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:49</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	06:47	Control Test	0.080	06:47	Air Blank	0.000	06:47	Control Test	0.080	06:48	Air Blank	0.000	06:48	Control Test	0.080	06:49	Air Blank	0.000	06:49	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:32																																																																																																																																																	
Control Test	0.050	06:32																																																																																																																																																	
Air Blank	0.000	06:33																																																																																																																																																	
Control Test	0.049	06:34																																																																																																																																																	
Air Blank	0.000	06:34																																																																																																																																																	
Control Test	0.049	06:35																																																																																																																																																	
Air Blank	0.000	06:36																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:37																																																																																																																																																	
Control Test	0.079	06:38																																																																																																																																																	
Air Blank	0.000	06:38																																																																																																																																																	
Control Test	0.079	06:39																																																																																																																																																	
Air Blank	0.000	06:39																																																																																																																																																	
Control Test	0.079	06:40																																																																																																																																																	
Air Blank	0.000	06:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:42																																																																																																																																																	
Control Test	0.200	06:42																																																																																																																																																	
Air Blank	0.000	06:43																																																																																																																																																	
Control Test	0.199	06:44																																																																																																																																																	
Air Blank	0.000	06:44																																																																																																																																																	
Control Test	0.198	06:45																																																																																																																																																	
Air Blank	0.000	06:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1990																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.5025																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:47																																																																																																																																																	
Control Test	0.080	06:47																																																																																																																																																	
Air Blank	0.000	06:47																																																																																																																																																	
Control Test	0.080	06:48																																																																																																																																																	
Air Blank	0.000	06:48																																																																																																																																																	
Control Test	0.080	06:49																																																																																																																																																	
Air Blank	0.000	06:49																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		