

INSTRUMENT PROCESSING SHEET

Agency Pinellas CSO

S/N 80-000890

Florida Department of
Law Enforcement

Date In 2/17/2022 DI Completion Date 3/17/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

[illegible]

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO

Time of Inspection: 10:20

Date of Inspection: 03/17/2022

Serial Number: 80-000890

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Not determined

mg
03/17/2022

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

03/17/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 0890	Pinellas CSO	03/17/2022	TDG MK

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L																																																																																																																																																	
0.047 to 0.053	✓	0.077 to 0.083	✓	0.194 to 0.206	✓	0.077 to 0.083	✓																																																																																																																																																
PINELLAS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000890 03/17/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:49</td></tr><tr><td>Control Test</td><td>0.051</td><td>10:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:50</td></tr><tr><td>Control Test</td><td>0.050</td><td>10:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:52</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0500</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>2.0000</td><td></td></tr></tbody></table> MK Operator's Signature		Test	g/210L	Time	Air Blank	0.000	10:49	Control Test	0.051	10:49	Air Blank	0.000	10:50	Control Test	0.050	10:51	Air Blank	0.000	10:51	Control Test	0.049	10:52	Air Blank	0.000	10:52	Control Test Stats			Average	0.0500		Std Dev	0.0010		Rel Std Dev(%)	2.0000		PINELLAS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000890 03/17/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:58</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:01</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2658</td><td></td></tr></tbody></table> MK Operator's Signature		Test	g/210L	Time	Air Blank	0.000	10:58	Control Test	0.078	10:59	Air Blank	0.000	11:00	Control Test	0.080	11:01	Air Blank	0.000	11:01	Control Test	0.079	11:02	Air Blank	0.000	11:02	Control Test Stats			Average	0.0790		Std Dev	0.0010		Rel Std Dev(%)	1.2658		PINELLAS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000890 03/17/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:09</td></tr><tr><td>Control Test</td><td>0.197</td><td>11:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:11</td></tr><tr><td>Control Test</td><td>0.198</td><td>11:11</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:12</td></tr><tr><td>Control Test</td><td>0.198</td><td>11:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:13</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1977</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr></tbody></table> MK Operator's Signature		Test	g/210L	Time	Air Blank	0.000	11:09	Control Test	0.197	11:10	Air Blank	0.000	11:11	Control Test	0.198	11:11	Air Blank	0.000	11:12	Control Test	0.198	11:13	Air Blank	0.000	11:13	Control Test Stats			Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		PINELLAS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000890 03/17/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr><tr><td>Control Test</td><td>0.078</td><td>11:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:16</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></tbody></table> MK Operator's Signature		Test	g/210L	Time	Air Blank	0.000	11:15	Control Test	0.078	11:15	Air Blank	0.000	11:15	Control Test	0.079	11:16	Air Blank	0.000	11:16	Control Test	0.079	11:17	Air Blank	0.000	11:17	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	10:49																																																																																																																																																					
Control Test	0.051	10:49																																																																																																																																																					
Air Blank	0.000	10:50																																																																																																																																																					
Control Test	0.050	10:51																																																																																																																																																					
Air Blank	0.000	10:51																																																																																																																																																					
Control Test	0.049	10:52																																																																																																																																																					
Air Blank	0.000	10:52																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0500																																																																																																																																																						
Std Dev	0.0010																																																																																																																																																						
Rel Std Dev(%)	2.0000																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	10:58																																																																																																																																																					
Control Test	0.078	10:59																																																																																																																																																					
Air Blank	0.000	11:00																																																																																																																																																					
Control Test	0.080	11:01																																																																																																																																																					
Air Blank	0.000	11:01																																																																																																																																																					
Control Test	0.079	11:02																																																																																																																																																					
Air Blank	0.000	11:02																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0790																																																																																																																																																						
Std Dev	0.0010																																																																																																																																																						
Rel Std Dev(%)	1.2658																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	11:09																																																																																																																																																					
Control Test	0.197	11:10																																																																																																																																																					
Air Blank	0.000	11:11																																																																																																																																																					
Control Test	0.198	11:11																																																																																																																																																					
Air Blank	0.000	11:12																																																																																																																																																					
Control Test	0.198	11:13																																																																																																																																																					
Air Blank	0.000	11:13																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.1977																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	0.2921																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	11:15																																																																																																																																																					
Control Test	0.078	11:15																																																																																																																																																					
Air Blank	0.000	11:15																																																																																																																																																					
Control Test	0.079	11:16																																																																																																																																																					
Air Blank	0.000	11:16																																																																																																																																																					
Control Test	0.079	11:17																																																																																																																																																					
Air Blank	0.000	11:17																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0787																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	0.7339																																																																																																																																																						

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 15:01

Date of Inspection: 03/17/2022

Serial Number: 80-000890
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.050	0.080	0.198	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.050	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.050	0.079	0.198	0.078

Standard Deviations	0.0004	0.0003	0.0000	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

03/17/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000890, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000890</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>PINELLAS COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>03/17/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:01</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/17/2022

Date


TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality