



INSTRUMENT PROCESSING SHEET

Agency Florida Fish and Wildlife Conservation Commission S/N 80-000904

Florida Department of
Law Enforcement

Date In 7/26/2022 DI Completion Date 7/27/2022 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>DERR</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: Pelican Size 50 hard case DGS. DERR	Quality Checks By <u>DERR</u> Date <u>7/27/2022</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>118</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.183</u> (.156 - .190) 53 mm <u>0.273</u> (.228 - .278) 103 mm <u>0.519</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28199</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td>MP5092</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904 06/08/2023</td> </tr> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202201C 01/11/2024	0.080	MP5093	202201D 01/18/2024	0.200	MP5094	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023	Flow Calibration By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																														
Simulator	Serial #	Lot #/Exp																																																													
0.050	MP5092	202201C 01/11/2024																																																													
0.080	MP5093	202201D 01/18/2024																																																													
0.200	MP5094	202201E 01/18/2024																																																													
0.080 DGS	N/A	AG115904 06/08/2023																																																													
		Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																																																													
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>DERR</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1017</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP5095</td> </tr> <tr> <td>Interferent</td> <td>MP5097</td> </tr> <tr> <td>0.050</td> <td>MP5092</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> </tr> </table>		Simulator	Serial Number	0.000	MP5095	Interferent	MP5097	0.050	MP5092	0.080	MP5093	0.200	MP5094
Simulator	Serial #	Lot #	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial #	Lot #	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	MP5095																																																														
Interferent	MP5097																																																														
0.050	MP5092																																																														
0.080	MP5093																																																														
0.200	MP5094																																																														
Notes/Suggested Service: <u>Tech review identified a miss type on the form 41. Corrected and resubmitted. DERR 7/27/2022</u>		Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																																													
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use		Israel Soto <small>Digitally signed by Israel Soto Date: 2022.07.27 10:33:58 -04'00'</small> Tech Review / Date _____ Admin Review / Date _____																																																													



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000904, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000904</u>	UNCERTAINTY * \pm
Owning Agency:	<u>FEWCC SOUTH REGION</u>	0.050 g/210 L 0.004
Calibration Date:	<u>07/27/2022</u>	0.080 g/210 L 0.004
Calibration Time:	<u>09:24</u>	0.200 g/210 L 0.007
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/27/2022

Date

DAVID E RIVERA-RIVERA,
Department Inspector

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC SOUTH REGION

Time of Inspection: 09:24

Date of Inspection: 07/27/2022

Serial Number: 80-000904

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) 2 Lot#:202201E Exp: 01/18/2024 DERR 7/27/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.048	0.078	0.198	0.080
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.079	0.200	0.080


Standard Deviations	0.0004	0.0004	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

DAVID E REYES-RIVERA

07/27/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000904	Florida Fish and Wildlife Conservation Commission	7/27/2022	DERR <i>[Signature]</i>

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>
<p>FFWC SOUTH REGION Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-000904 07/27/2022 Software: 8.00.27</p> <p>Test g/210L Time</p> <p>Alc Blank 0.000 07:06 Control Test 0.049 07:07 Alc Blank 0.000 07:08 Control Test 1.069 07:08 Alc Blank 1.000 07:09 Control Test 1.049 07:09 Alc Blank 0.000 07:10 Control Test Status</p> <p>Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature <i>[Signature]</i></p>	<p>FFWC SOUTH REGION Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-000904 07/27/2022 Software: 8.00.27</p> <p>Test g/210L Time</p> <p>Alc Blank 0.000 07:11 Control Test 0.078 07:12 Alc Blank 0.000 07:13 Control Test 0.078 07:13 Alc Blank 0.000 07:14 Control Test 0.079 07:14 Alc Blank 0.000 07:15 Control Test Status</p> <p>Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370</p> <p>Operator's Signature <i>[Signature]</i></p>	<p>FFWC SOUTH REGION Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-000904 07/27/2022 Software: 8.00.27</p> <p>Test g/210L Time</p> <p>Alc Blank 0.000 07:17 Control Test 0.199 07:17 Alc Blank 0.000 07:18 Control Test 0.199 07:18 Alc Blank 0.000 07:19 Control Test 0.199 07:19 Alc Blank 0.000 07:20 Control Test Status</p> <p>Average 0.1990 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature <i>[Signature]</i></p>	<p>FFWC SOUTH REGION Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-000904 07/27/2022 Software: 8.00.27</p> <p>Test g/210L Time</p> <p>Alc Blank 0.000 07:21 Control Test 0.080 07:22 Alc Blank 0.000 07:22 Control Test 0.078 07:23 Alc Blank 0.000 07:23 Control Test 0.079 07:23 Alc Blank 0.000 07:24 Control Test Status</p> <p>Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.2658</p> <p>Operator's Signature <i>[Signature]</i></p>