





INSTRUMENT PROCESSING SHEET

Agency Florida Fish and Wildlife Conservation CommissionS/N 80-000906Florida Department of
Law EnforcementDate In 7/26/2022 DI Completion Date 7/27/2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>DERR</u>	Quality Checks	By <u>DERR</u>	Date <u>7/27/2022</u>	Flow Calibration	By	Date																																							
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: Pelican Size 50 hard case, DGS. DERR		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>225</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5092</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5093</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5094</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202201C 01/11/2024	0.080	MP5093	202201D 01/18/2024	0.200	MP5094	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023																													
Simulator	Serial #	Lot #/Exp																																												
0.050	MP5092	202201C 01/11/2024																																												
0.080	MP5093	202201D 01/18/2024																																												
0.200	MP5094	202201E 01/18/2024																																												
0.080 DGS	N/A	AG115904 06/08/2023																																												
Calibration Adjustment		By _____		Department Inspection			By <u>DERR</u>																																							
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>26932</u> Gauge <u>1017</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5095</td></tr><tr><td>Interferent</td><td>MP5097</td></tr><tr><td>0.050</td><td>MP5092</td></tr><tr><td>0.080</td><td>MP5093</td></tr><tr><td>0.200</td><td>MP5094</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP5095	Interferent	MP5097	0.050	MP5092	0.080	MP5093	0.200	MP5094	
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP5095																																													
Interferent	MP5097																																													
0.050	MP5092																																													
0.080	MP5093																																													
0.200	MP5094																																													
☐ Post Calibration Adjustment Stability Checks				Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																						
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																										
		Israel Soto		Digitally signed by Israel Soto Date: 2022.07.27 10:12:16 +0400'			Date: 2022.07.27 11:10:17 -04'00'																																							
		Tech Review / Date		Admin Review / Date																																										



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000906, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000906</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FISH & WILDLIFE CC</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/27/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>09:33</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

07/27/2022

Date

DAVID E REYES-RIVERA,

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FISH & WILDLIFE CC
Time of Inspection: 09:33

Date of Inspection: 07/27/2022

Serial Number: 80-000906
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.048	0.077	0.200	0.078
0.000	0.048	0.077	0.200	0.079
0.000	0.049	0.077	0.200	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.049	0.077	0.199	0.079
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.078	0.200	0.078
0.000	0.049	0.078	0.200	0.078

Standard Deviations	0.0005	0.0006	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera Signature and Printed Name

07/27/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000906	Florida Fish and Wildlife Conservation Commission	7/27/2022	DERR <i>MLL</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
FISH & WILDLIFE CC Intoxilyzer - Alcohol Analyzer Model 8000 07/27/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 07:11 Control Test 0.049 07:12 Air Blank 0.000 07:12 Control Test 0.049 07:13 Air Blank 0.000 07:14 Control Test 0.049 07:14 Air Blank 0.000 07:15 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>MLL</i>	FISH & WILDLIFE CC Intoxilyzer - Alcohol Analyzer Model 8000 07/27/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 07:16 Control Test 0.079 07:17 Air Blank 0.000 07:17 Control Test 0.078 07:18 Air Blank 0.000 07:19 Control Test 0.078 07:19 Air Blank 0.000 07:20 Control Test Stats Average 0.0783 Std Dev 0.0005 Rel Std Dev(%) 0.7370 Operator's Signature <i>MLL</i>	FISH & WILDLIFE CC Intoxilyzer - Alcohol Analyzer Model 8000 07/27/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 07:21 Control Test 0.200 07:22 Air Blank 0.000 07:22 Control Test 0.198 07:23 Air Blank 0.000 07:24 Control Test 0.198 07:24 Air Blank 0.000 07:25 Control Test Stats Average 0.1987 Std Dev 0.0012 Rel Std Dev(%) 0.5812 Operator's Signature <i>MLL</i>	FISH & WILDLIFE CC Intoxilyzer - Alcohol Analyzer Model 8000 07/27/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 07:26 Control Test 0.079 07:27 Air Blank 0.000 07:27 Control Test 0.079 07:27 Air Blank 0.000 07:28 Control Test 0.079 07:28 Air Blank 0.000 07:29 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>MLL</i>