

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Jen Gaytan on 10/25/2022

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000938

Bill To Address:

Lee County Sheriff's Office

Attn: Jen Gaytan

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Alcohol Testing Program

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

The analog connector piece broke during upload and retracted into the instrument. Records
have all been uploaded by FDLE.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$ _____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Jen Gaytan

Phone #: 239-707-8476

Email: jgaytan@sheriffleefl.org

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000938	Lee CSO	10/25/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/25/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:02 Control Test 0.050 10:03 Air Blank 0.000 10:03 Control Test 0.049 10:04 Air Blank 0.000 10:05 Control Test 0.049 10:05 Air Blank 0.000 10:06 Control Test Stats Average 0.0493 Std Dev 0.0006 Rel Std Dev(%) 1.1703	LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/25/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:09 Control Test 0.079 10:10 Air Blank 0.000 10:10 Control Test 0.078 10:11 Air Blank 0.000 10:12 Control Test 0.079 10:12 Air Blank 0.000 10:13 Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7339	LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/25/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:16 Control Test 0.200 10:17 Air Blank 0.000 10:17 Control Test 0.201 10:18 Air Blank 0.000 10:18 Control Test 0.200 10:19 Air Blank 0.000 10:20 Control Test Stats Average 0.2003 Std Dev 0.0006 Rel Std Dev(%) 0.2882	LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/25/2022 Software: 8101.27 Test g/210L Time Air Blank 0.000 09:58 Control Test 0.079 09:58 Air Blank 0.000 09:59 Control Test 0.077 09:59 Air Blank 0.000 10:00 Control Test 0.078 10:00 Air Blank 0.000 10:01 Control Test Stats Average 0.0780 Std Dev 0.0010 Rel Std Dev(%) 1.2821
<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEE COUNTY SO
Time of Inspection: 12:28

Date of Inspection: 10/25/2022

Serial Number: 80-000938
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.050	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.049	0.078	0.200	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.079	0.199	0.077
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.078	0.199	0.078
Standard Deviations	0.0004	0.0004	0.0005	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

TAYLOR D GUTSCHOW

10/25/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000938, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000938</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>LEE COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/25/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:28</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

10/25/2022

Date



TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality