



INSTRUMENT PROCESSING SHEET

Agency Glades County Sheriff's OfficeS/N 80-000947Florida Department of
Law EnforcementDate In 5/13/2022DI Completion Date 5/19/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>DERR</u>	Quality Checks	By <u>DERR</u>	Date <u>5/19/2022</u>	Flow Calibration	By _____	Date _____																																						
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>230</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																								
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP6287</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP6288</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP6287	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023			Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____																									
Simulator	Serial #	Lot #/Exp																																											
0.050	MP6286	202201C 01/11/2024																																											
0.080	MP6287	202201D 01/18/2024																																											
0.200	MP6288	202201E 01/18/2024																																											
0.080 DGS	N/A	AG115904 06/08/2023																																											
Calibration Adjustment By _____		Department Inspection By <u>DERR</u>																																											
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>28199</u> Gauge <u>1015</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	MP6284																																												
Interferent	MP6285																																												
0.050	MP6286																																												
0.080	MP6287																																												
0.200	MP6288																																												
☐ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☒ Other <u>Prior work docs</u>																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																							
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		Israel Soto Digitally signed by Israel Soto Date: 2022.05.20 08:35:43 +0400 Tech Review / Date _____ Admin Review / Date <u>2022.05.20 10:25:00</u>																																											



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000947, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000947</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>GLADES COUNTY S.O.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/19/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:28</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/19/2022

Date

DAVID EREYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GLADES COUNTY S.O.
Time of Inspection: 08:28

Date of Inspection: 05/19/2022

Serial Number: 80-000947
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.048	0.080	0.203	0.080
0.000	0.049	0.080	0.202	0.081
0.000	0.049	0.080	0.202	0.080
0.000	0.050	0.080	0.202	0.080
0.000	0.049	0.080	0.202	0.080
0.000	0.050	0.080	0.203	0.080
0.000	0.049	0.080	0.202	0.081
0.000	0.049	0.080	0.203	0.080
0.000	0.050	0.081	0.203	0.080
0.000	0.050	0.080	0.202	0.081


Standard Deviations	0.0006	0.0003	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

DAVID E REYES-RIVERA

05/19/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000947	Glades County Sheriff's Office	5/19/2022	DERR <i>hll</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒																																																																																																																																																
GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 05/19/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>06:37</td></tr><tr><td>Control Test</td><td>0.050</td><td>06:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:38</td></tr><tr><td>Control Test</td><td>0.049</td><td>06:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:39</td></tr><tr><td>Control Test</td><td>0.049</td><td>06:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:41</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></table> Operator's Signature <i>hll</i>	Test	g/210L	Time	Air Blank	0.000	06:37	Control Test	0.050	06:38	Air Blank	0.000	06:38	Control Test	0.049	06:39	Air Blank	0.000	06:39	Control Test	0.049	06:40	Air Blank	0.000	06:41	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 05/19/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>06:42</td></tr><tr><td>Control Test</td><td>0.079</td><td>06:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:43</td></tr><tr><td>Control Test</td><td>0.078</td><td>06:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:44</td></tr><tr><td>Control Test</td><td>0.078</td><td>06:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:46</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></table> Operator's Signature <i>hll</i>	Test	g/210L	Time	Air Blank	0.000	06:42	Control Test	0.079	06:43	Air Blank	0.000	06:43	Control Test	0.078	06:44	Air Blank	0.000	06:44	Control Test	0.078	06:45	Air Blank	0.000	06:46	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 05/19/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>06:46</td></tr><tr><td>Control Test</td><td>0.202</td><td>06:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:48</td></tr><tr><td>Control Test</td><td>0.202</td><td>06:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:49</td></tr><tr><td>Control Test</td><td>0.201</td><td>06:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:50</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2017</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2863</td><td></td></tr></table> Operator's Signature <i>hll</i>	Test	g/210L	Time	Air Blank	0.000	06:46	Control Test	0.202	06:47	Air Blank	0.000	06:48	Control Test	0.202	06:48	Air Blank	0.000	06:49	Control Test	0.201	06:50	Air Blank	0.000	06:50	Control Test Stats			Average	0.2017		Std Dev	0.0006		Rel Std Dev(%)	0.2863		GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 05/19/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>06:51</td></tr><tr><td>Control Test</td><td>0.080</td><td>06:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:52</td></tr><tr><td>Control Test</td><td>0.080</td><td>06:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:53</td></tr><tr><td>Control Test</td><td>0.079</td><td>06:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:54</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></table> Operator's Signature <i>hll</i>	Test	g/210L	Time	Air Blank	0.000	06:51	Control Test	0.080	06:52	Air Blank	0.000	06:52	Control Test	0.080	06:53	Air Blank	0.000	06:53	Control Test	0.079	06:53	Air Blank	0.000	06:54	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:37																																																																																																																																																	
Control Test	0.050	06:38																																																																																																																																																	
Air Blank	0.000	06:38																																																																																																																																																	
Control Test	0.049	06:39																																																																																																																																																	
Air Blank	0.000	06:39																																																																																																																																																	
Control Test	0.049	06:40																																																																																																																																																	
Air Blank	0.000	06:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:42																																																																																																																																																	
Control Test	0.079	06:43																																																																																																																																																	
Air Blank	0.000	06:43																																																																																																																																																	
Control Test	0.078	06:44																																																																																																																																																	
Air Blank	0.000	06:44																																																																																																																																																	
Control Test	0.078	06:45																																																																																																																																																	
Air Blank	0.000	06:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:46																																																																																																																																																	
Control Test	0.202	06:47																																																																																																																																																	
Air Blank	0.000	06:48																																																																																																																																																	
Control Test	0.202	06:48																																																																																																																																																	
Air Blank	0.000	06:49																																																																																																																																																	
Control Test	0.201	06:50																																																																																																																																																	
Air Blank	0.000	06:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2017																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:51																																																																																																																																																	
Control Test	0.080	06:52																																																																																																																																																	
Air Blank	0.000	06:52																																																																																																																																																	
Control Test	0.080	06:53																																																																																																																																																	
Air Blank	0.000	06:53																																																																																																																																																	
Control Test	0.079	06:53																																																																																																																																																	
Air Blank	0.000	06:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Sergeant J. Bonilla on 08/18/2021

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000947

Bill To Address:

Glades County SO

1297 E. State Road 78

Moore Haven, FL., 33471

Ship to Address:

FDLE ATP

ATTN: Michael Haughey

4700 Terminal Drive Suite 1

Fort Myers FL 33907

Reason for Return:

Unable to calibrate instrument. Checked DVMS 3 is higher then 9, and when flicked signal jumps up. Dry Gas leaks.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Sergeant J. Bonilla

Phone #: 863-946-1600

Email: jbonilla@gladessheriff.org

ATP Contact Name: Michael Haughey

ATP Email: michaelhaughey@fdle.state.fl.us



INSTRUMENT PROCESSING SHEET

Agency Glades County SOS/N 80-000947Florida Department of
Law EnforcementDate In 7/6/2021 DI Completion Date _____☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By MH	Quality Checks	By MH	Date 08/05/2021	Flow Calibration	By	Date																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>DG Leak</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>235</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 106</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5092</td><td>202010A 10/05/2022</td></tr><tr><td>0.080</td><td>MP5093</td><td>202010B 10/05/2022</td></tr><tr><td>0.200</td><td>MP5097</td><td>202010D 10/06/2022</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG026705 09/23/2022</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202010A 10/05/2022	0.080	MP5093	202010B 10/05/2022	0.200	MP5097	202010D 10/06/2022	0.080 DGS	N/A	AG026705 09/23/2022			Maintenance By MH ☐ Battery Replacement ☒ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>DG Calibration</u>															
Simulator	Serial #	Lot #/Exp																																	
0.050	MP5092	202010A 10/05/2022																																	
0.080	MP5093	202010B 10/05/2022																																	
0.200	MP5097	202010D 10/06/2022																																	
0.080 DGS	N/A	AG026705 09/23/2022																																	
					DI Temp. Checks By MH ☒ Lab Temp °C External Digital Therm. ID#: <u>300504</u> ☒ 34°C +/- .2 Serial #: <u>MP5092</u> ☒ 34°C +/- .2 Serial #: <u>MP5093</u> ☒ 34°C +/- .2 Serial #: <u>MP5097</u>																														
Calibration Adjustment By MH		Department Inspection By MH																																	
Barometric Pressure Gauge <u>28663</u> ID # <u>1019</u>		Barometric Pressure ID# <u>68639</u>		Gauge _____ Instrument _____																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Mouth Alcohol Solution Lot # <u>2021-A</u>		Acetone Stock Solution Lot # <u>2020-A</u>			
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
☐ Post Calibration Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>SD1014</td></tr><tr><td>Interferent</td><td>SD1015</td></tr><tr><td>0.050</td><td>MP4863</td></tr><tr><td>0.080</td><td>MP4864</td></tr><tr><td>0.200</td><td>MP5097</td></tr></tbody></table>		Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5097	Attachments																			
Simulator	Serial Number																																		
0.000	SD1014																																		
Interferent	SD1015																																		
0.050	MP4863																																		
0.080	MP4864																																		
0.200	MP5097																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment		☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____											
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Notes/Suggested Service: <u>Unstable signal unable to calibration. Recommend repair.</u> <u>Sent to repair with leaking Regulator.</u>		☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																	
		Israel Soto Digitally signed by Israel Soto Date: 2022.05.20 09:36:59 +0400		Tech Review / Date _____ Admin Review / Date _____																															

GLADES COUNTY S.C.

Intoxilyzer - Alcotest Analyzer

Model 8500

08/11/2021 09:03:35

Auto Calibration

Max Power Res Value = 85

Auto Range Res Value = 4:

Sol Value = 0.000 g/210L ***

Fit Value = 0.0000 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12127, Sum Io = 12100

**** CHANNEL 1 ****

Sample % Abs (-0.0200)

Sample #1 = 0.2440 (0.1590)

Sample #2 = 0.2120 (0.1590)

Sample #3 = 0.1630 (0.2200)

Sample #4 = 0.1530 (0.4280)

Avg % Abs = 0.1760 (0.2930)

STD DEV = 0.0316 (0.1345)

REL STD DEV = 17.941 (45.905)

**** CHANNEL 2 ****

Sample % Abs (-0.0230)

Sample #1 = 0.2210 (0.1500)

Sample #2 = 0.1420 (0.2310)

Sample #3 = 0.1730 (0.2830)

Sample #4 = 0.1637 (0.2247)

Avg % Abs = 0.1688 (0.0617)

STD DEV = 11.501 (27.483)

REL STD DEV = 11.501 (27.483)

Sol Value = 0.040 g/210L ***

Fit Value = 0.1905 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12652, Sum Io = 12049

**** CHANNEL 1 ****

Sample % Abs (-0.0250)

Sample #1 = 0.0300 (0.1080)

Sample #2 = 0.0640 (0.1580)

Sample #3 = 0.0940 (0.2250)

Sample #4 = 0.0650 (0.1550)

Avg % Abs = 0.0657 (0.1550)

STD DEV = 0.0333 (0.0608)

REL STD DEV = 3.759 (35.850)

**** CHANNEL 2 ****

Sample % Abs (-0.0180)

Sample #1 = 1.5770 (0.0390)

Sample #2 = 1.5840 (0.1080)

Sample #3 = 1.5450 (0.1140)

Sample #4 = 1.4920 (0.0870)

Avg % Abs = 1.5403 (0.0870)

STD DEV = 0.3452 (0.0417)

REL STD DEV = 2.998 (47.905)

Sol Value = 0.100 g/210L ***

Fit Value = 0.4762 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12596, Sum Io = 12018

**** CHANNEL 1 ****

Sample % Abs (-0.0190)

Sample #1 = 1.0970 (0.0190)

Sample #2 = 1.9930 (0.0620)

Sample #3 = 1.9790 (0.0210)

Sample #4 = 1.9240 (0.1130)

Avg % Abs = 1.9653 (0.0240)

STD DEV = 0.0365 (0.0875)

REL STD DEV = 1.856 (364.744)

**** CHANNEL 2 ****

Sample % Abs (-0.0070)

Sample #1 = 3.4460 (0.0990)

Sample #2 = 3.5350 (0.0720)

Sample #3 = 3.5030 (0.0040)

Sample #4 = 3.5430 (0.0583)

Avg % Abs = 3.5537 (0.0490)

STD DEV = 0.0257 (0.0490)

REL STD DEV = 0.724 (83.918)

Sol Value = 0.200 g/210L ***

Fit Value = 0.9524 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12551, Sum Io = 11993

**** CHANNEL 1 ****

Sample % Abs (-0.0230)

Sample #1 = 3.7570 (0.0230)

Sample #2 = 3.7710 (0.0170)

Sample #3 = 3.7710 (0.0590)

Sample #4 = 3.7400 (0.1110)

Avg % Abs = 3.7607 (0.0623)

STD DEV = 0.0179 (0.0471)

REL STD DEV = 0.476 (75.543)

**** AUTO CAL DATA ****

**** CHANNEL 1 ****

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.176

Std Dev = 0.03 Rel Std Dev = 17.94

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.386

Std Dev = 0.03 Rel Std Dev = 3.76

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 1.965

Std Dev = 0.04 Rel Std Dev = 1.86

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.761

Std Dev = 0.02 Rel Std Dev = 0.48

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 5.429

Std Dev = 0.01 Rel Std Dev = 0.21

Zero Order Coef = -414.96

First Order Coef = 2559.93

Second Order Coef = 26.44

Standard Deviation = 46.525543

**** CHANNEL 2 ****

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.164

Std Dev = 0.02 Rel Std Dev = 11.50

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.540

Std Dev = 0.05 Rel Std Dev = 3.00

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.554

Std Dev = 0.03 Rel Std Dev = 0.72

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 6.800

Std Dev = 0.01 Rel Std Dev = 0.16

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 9.807

Std Dev = 0.03 Rel Std Dev = 0.30

Zero Order Coef = -205.58

First Order Coef = 1342.47

Second Order Coef = 13.66

Standard Deviation = 22.471313

Optical Calibration



SN: 80-000947

Agency: Glades County

Date: 8/11/21

By: JMA

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities - Post	80-000947	Glades County SO	08/11/2021	MMO

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 ☐	0.077 to 0.083 ☐	0.194 to 0.206 ☐	0.077 to 0.083 ☐																																																																																																																																																
GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000947 08/11/2021 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>09:49</td></tr><tr><td>Control Test</td><td>0.051</td><td>09:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr><tr><td>Control Test</td><td>0.049</td><td>09:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:51</td></tr><tr><td>Control Test</td><td>0.050</td><td>09:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:53</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0500</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>2.0000</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	09:49	Control Test	0.051	09:50	Air Blank	0.000	09:50	Control Test	0.049	09:51	Air Blank	0.000	09:51	Control Test	0.050	09:52	Air Blank	0.000	09:53	Control Test Stats			Average	0.0500		Std Dev	0.0010		Rel Std Dev(%)	2.0000		GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000947 08/11/2021 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>09:53</td></tr><tr><td>Control Test</td><td>0.084</td><td>09:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:55</td></tr><tr><td>Control Test</td><td>0.079</td><td>09:55</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:56</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:57</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0810</td><td></td></tr><tr><td>Std Dev</td><td>0.0026</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>3.2664</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	09:53	Control Test	0.084	09:54	Air Blank	0.000	09:55	Control Test	0.079	09:55	Air Blank	0.000	09:56	Control Test	0.080	09:57	Air Blank	0.000	09:57	Control Test Stats			Average	0.0810		Std Dev	0.0026		Rel Std Dev(%)	3.2664		GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000947 08/11/2021 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>09:58</td></tr><tr><td>Control Test</td><td>0.199</td><td>09:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:59</td></tr><tr><td>Control Test</td><td>0.198</td><td>10:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:00</td></tr><tr><td>Control Test</td><td>0.198</td><td>10:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:02</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1983</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	09:58	Control Test	0.199	09:59	Air Blank	0.000	09:59	Control Test	0.198	10:00	Air Blank	0.000	10:00	Control Test	0.198	10:01	Air Blank	0.000	10:02	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000947 08/11/2021 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:04</td></tr><tr><td>Control Test</td><td>0.077</td><td>10:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:05</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:05</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr><tr><td>Control Test</td><td>0.077</td><td>10:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0777</td><td></td></tr><tr><td>Std Dev</td><td>0.0012</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.4667</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	10:04	Control Test	0.077	10:04	Air Blank	0.000	10:05	Control Test	0.079	10:05	Air Blank	0.000	10:06	Control Test	0.077	10:06	Air Blank	0.000	10:06	Control Test Stats			Average	0.0777		Std Dev	0.0012		Rel Std Dev(%)	1.4667	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:49																																																																																																																																																	
Control Test	0.051	09:50																																																																																																																																																	
Air Blank	0.000	09:50																																																																																																																																																	
Control Test	0.049	09:51																																																																																																																																																	
Air Blank	0.000	09:51																																																																																																																																																	
Control Test	0.050	09:52																																																																																																																																																	
Air Blank	0.000	09:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	2.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:53																																																																																																																																																	
Control Test	0.084	09:54																																																																																																																																																	
Air Blank	0.000	09:55																																																																																																																																																	
Control Test	0.079	09:55																																																																																																																																																	
Air Blank	0.000	09:56																																																																																																																																																	
Control Test	0.080	09:57																																																																																																																																																	
Air Blank	0.000	09:57																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0026																																																																																																																																																		
Rel Std Dev(%)	3.2664																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:58																																																																																																																																																	
Control Test	0.199	09:59																																																																																																																																																	
Air Blank	0.000	09:59																																																																																																																																																	
Control Test	0.198	10:00																																																																																																																																																	
Air Blank	0.000	10:00																																																																																																																																																	
Control Test	0.198	10:01																																																																																																																																																	
Air Blank	0.000	10:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:04																																																																																																																																																	
Control Test	0.077	10:04																																																																																																																																																	
Air Blank	0.000	10:05																																																																																																																																																	
Control Test	0.079	10:05																																																																																																																																																	
Air Blank	0.000	10:06																																																																																																																																																	
Control Test	0.077	10:06																																																																																																																																																	
Air Blank	0.000	10:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	1.4667																																																																																																																																																		
																																																																																																																																																			
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature																																																																																																																																																

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities - POST	80-000947	Glades County SO	8/11/21	MMH

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																								
0.047 to 0.053 ☒	0.077 to 0.083 ☐	0.194 to 0.206 ☐	0.077 to 0.083 ☐																								
GLADES COUNTY S.O. Intoxilyzer - Etico. Analyzer Model 8000 08/11/2021 Software: 8.0000 SN 80-000947 <table><thead><tr><th>Test</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000 09:45</td></tr><tr><td>Control Test</td><td>0.038 09:45</td></tr><tr><td>Air Blank</td><td>0.000 09:46</td></tr><tr><td>Control Test</td><td>0.051 09:47</td></tr><tr><td>Air Blank</td><td>0.000 09:47</td></tr><tr><td>Control Test</td><td>0.048 09:48</td></tr><tr><td>Air Blank</td><td>0.000 09:48</td></tr><tr><td colspan="2">Control Test Stats</td></tr><tr><td>Average</td><td>0.0457</td></tr><tr><td>Std Dev</td><td>0.0068</td></tr><tr><td>Rel Std Dev(%)</td><td>14.9055</td></tr></tbody></table> MMH Tighten simulator calibration port after 0.038g/210L May have been loose.				Test	Time	Air Blank	0.000 09:45	Control Test	0.038 09:45	Air Blank	0.000 09:46	Control Test	0.051 09:47	Air Blank	0.000 09:47	Control Test	0.048 09:48	Air Blank	0.000 09:48	Control Test Stats		Average	0.0457	Std Dev	0.0068	Rel Std Dev(%)	14.9055
Test	Time																										
Air Blank	0.000 09:45																										
Control Test	0.038 09:45																										
Air Blank	0.000 09:46																										
Control Test	0.051 09:47																										
Air Blank	0.000 09:47																										
Control Test	0.048 09:48																										
Air Blank	0.000 09:48																										
Control Test Stats																											
Average	0.0457																										
Std Dev	0.0068																										
Rel Std Dev(%)	14.9055																										

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000947	Glades County SO	08/05/21	MX

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒
GLADES COUNTY S.O. Intoxilyzer - Alcotest Analyzer Model 8000 08/05/2021 Software: 8100.27 Test 9/210L Time Air Blank 0.000 09:53 Control Test 0.048 09:54 Air Blank 0.000 09:55 Control Test 0.050 09:55 Air Blank 0.000 09:56 Control Test 0.049 09:57 Air Blank 0.000 09:57 Control Test Stats Average 0.0490 Std Dev 0.0010 Rel Std Dev(%) 2.0408	GLADES COUNTY S.O. Intoxilyzer - Alcotest Analyzer Model 8000 08/05/2021 Software: 8100.27 Test 9/210L Time Air Blank 0.000 09:59 Control Test 0.078 09:59 Air Blank 0.000 10:00 Control Test 0.078 10:01 Air Blank 0.000 10:01 Control Test 0.077 10:02 Air Blank 0.000 10:02 Control Test Stats Average 0.0777 Std Dev 0.0006 Rel Std Dev(%) 0.7434	GLADES COUNTY S.O. Intoxilyzer - Alcotest Analyzer Model 8000 08/05/2021 Software: 8100.27 Test 9/210L Time Air Blank 0.000 10:05 Control Test 0.201 10:05 Air Blank 0.000 10:06 Control Test 0.200 10:07 Air Blank 0.000 10:07 Control Test 0.201 10:08 Air Blank 0.000 10:08 Control Test Stats Average 0.2007 Std Dev 0.0006 Rel Std Dev(%) 0.2877	GLADES COUNTY S.O. Intoxilyzer - Alcotest Analyzer Model 8000 08/05/2021 Software: 8100.27 Test 9/210L Time Air Blank 0.000 10:11 Control Test 0.083 10:12 Air Blank 0.000 10:12 Control Test 0.080 10:13 Air Blank 0.000 10:13 Control Test 0.082 10:14 Air Blank 0.000 10:14 Control Test Stats Average 0.0817 Std Dev 0.0015 Rel Std Dev(%) 1.8704

MX

MX

MX

MX

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GLADES COUNTY S.O.

Time of Inspection: 09:02

Date of Inspection: 08/05/2021

Serial Number: 80-000947

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____


Number of Simulators Used: _____

Remarks:

COMPLAINT NOT DETERMINED. BYPASS AI.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

08/05/2021
Date