



Agency Franklin County Sheriff's Office

S/N 80-000952

Date In 12/28/2021 DI Completion Date 1/3/2022

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

FDLE/ATP Form 48 January 2021
Issuing Authority: Alcohol Testing Program

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FRANKLIN COUNTY SO
Time of Inspection: 13:59

Date of Inspection: 01/03/2022

Serial Number: 80-000952
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.081	0.203	0.079
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.080	0.203	0.079
0.000	0.050	0.081	0.203	0.078
0.000	0.050	0.080	0.202	0.078
0.000	0.050	0.081	0.203	0.078
0.000	0.051	0.080	0.202	0.079
0.000	0.050	0.081	0.202	0.078
0.000	0.050	0.081	0.202	0.078
Standard Deviations	0.0003	0.0004	0.0005	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

01/03/2022
Date

Stability Checks

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:31
Control Test	0.051	11:32
Air Blank	0.000	11:32
Control Test	0.050	11:33
Air Blank	0.000	11:33
Control Test	0.050	11:34
Air Blank	0.000	11:35
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
01/03/2022
Software: 8100.27

Test*	g/210L	Time
Air Blank	0.000	11:36
Control Test	0.080	11:37
Air Blank	0.000	11:37
Control Test	0.080	11:38
Air Blank	0.000	11:38
Control Test	0.080	11:39
Air Blank	0.000	11:40
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet

Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:41
Control Test	0.202	11:42
Air Blank	0.000	11:42
Control Test	0.202	11:43
Air Blank	0.000	11:43
Control Test	0.201	11:44
Air Blank	0.000	11:45
Control Test Stats		
Average	0.2017	
Std Dev	0.0006	
Rel Std Dev(%)	0.2863	

Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:47
Control Test	0.081	11:47
Air Blank	0.000	11:48
Control Test	0.080	11:48
Air Blank	0.000	11:49
Control Test	0.081	11:49
Air Blank	0.000	11:49
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Dry

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000952, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000952</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FRANKLIN COUNTY SO</u>	0.050 g/ 210 L 0.005
Calibration Date:	<u>01/03/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:59</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Date: 2022.01.03 14:58:30
-05'00'

01/03/2022

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Franklin County Sheriff's Office

S/N 80-000952

Florida Department of
Law Enforcement

Date In 7/7/2022

DI Completion Date 07-22-2022

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>PN</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ Agency Inspector indicated Dry Gas Regulator was leaking. Confirmed regulator was leaking in lab. - IS _____ _____	Quality Checks By <u>PN</u> Date <u>7/22/2022</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>220</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 103</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG121002 07/29/2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01/11/2024	0.080	MP6292	202201D 01/18/2024	0.200	MP6293	202201E 01/18/2024	0.080 DGS	N/A	AG121002 07/29/2023	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By <u>IS</u> ☐ Battery Replacement ☒ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Dry Gas Regulator Tare</u> _____ _____ _____																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	MP6291	202201C 01/11/2024																																																
0.080	MP6292	202201D 01/18/2024																																																
0.200	MP6293	202201E 01/18/2024																																																
0.080 DGS	N/A	AG121002 07/29/2023																																																
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: <u>Regulator replaced prior to</u> <u>Quality Checks and Department Inspection. - IS</u> _____ _____ _____ _____ _____			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial #	Lot #	Expiration																																															
0.000		N/A	N/A																																															
0.040																																																		
0.100																																																		
0.200																																																		
0.300																																																		
0.080 DGS	N/A																																																	
Simulator	Serial #	Lot #	Expiration																																															
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
Department Inspection By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1015</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP6291</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2022.07.22 15:03:14 -0400 <u>2022.07.26</u> <u>11:18:10</u> Tech Review / Date Admin Review / Date			Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293																																				
Simulator	Serial Number																																																	
0.000	MP6289																																																	
Interferent	MP6290																																																	
0.050	MP6291																																																	
0.080	MP6292																																																	
0.200	MP6293																																																	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FRANKLIN COUNTY SO
Time of Inspection: 13:18

Date of Inspection: 07/22/2022

Serial Number: 80-000952
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.080	0.205	0.080
0.000	0.049	0.081	0.205	0.079
0.000	0.050	0.080	0.205	0.079
0.000	0.050	0.081	0.205	0.079
0.000	0.050	0.081	0.205	0.079
0.000	0.050	0.081	0.204	0.079
0.000	0.049	0.081	0.203	0.079
0.000	0.048	0.081	0.202	0.079
0.000	0.047	0.081	0.202	0.080
0.000	0.047	0.081	0.202	0.079

Standard Deviations	0.0012	0.0004	0.0013	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0008 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

07/22/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	80-000952	Franklin County Sheriff's Office	7/22/2022	PN

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-000952
07/22/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:29
Control Test	0.050	10:30
Air Blank	0.000	10:30
Control Test	0.049	10:31
Air Blank	0.000	10:31
Control Test	0.049	10:32
Air Blank	0.000	10:33
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	


Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-000952
07/22/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:34
Control Test	0.080	10:34
Air Blank	0.000	10:35
Control Test	0.079	10:36
Air Blank	0.000	10:36
Control Test	0.080	10:37
Air Blank	0.000	10:37
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	


Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-000952
07/22/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:38
Control Test	0.204	10:39
Air Blank	0.000	10:39
Control Test	0.204	10:40
Air Blank	0.000	10:41
Control Test	0.204	10:41
Air Blank	0.000	10:42
Control Test Stats		
Average	0.2040	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-000952
07/22/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:43
Control Test	0.079	10:43
Air Blank	0.000	10:44
Control Test	0.080	10:44
Air Blank	0.000	10:44
Control Test	0.080	10:45
Air Blank	0.000	10:45
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	


Operator's Signature

Comments: N/A



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000952, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000952</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FRANKLIN COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/22/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:18</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel
Date: 2022.07.22
13:32:29 -04'00'

07/22/2022

Date

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality