

INSTRUMENT PROCESSING SHEET

Agency Leon County Sheriff's Office

S/N 80-000957

Date In 01-27-2022 DI Completion Date 01-31-2022

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	Quality Checks	Flow Calibration																																																																								
By <u>IS</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>208</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202010A</td> </tr> <tr> <td>10-05-2022</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202010B</td> </tr> <tr> <td>10-05-2022</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202010D</td> </tr> <tr> <td>10-06-2022</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG113403</td> </tr> <tr> <td>05-14-2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202010A	10-05-2022	0.080	MP5089	202010B	10-05-2022	0.200	MP5090	202010D	10-06-2022	0.080 DGS	N/A	AG113403	05-14-2023	Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Maintenance</th> <th>By</th> </tr> </thead> <tbody> <tr> <td>☐ Battery Replacement</td> <td></td> </tr> <tr> <td>☐ Dry Gas Regulator Replacement</td> <td></td> </tr> <tr> <td>☐ Breath Tube Replacement</td> <td></td> </tr> <tr> <td>☐ Other _____</td> <td></td> </tr> <tr><td> </td><td></td></tr> <tr><td> </td><td></td></tr> <tr><td> </td><td></td></tr> <tr><td> </td><td></td></tr> </tbody> </table>	Maintenance	By	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____																																												
Simulator	Serial #	Lot #/Exp																																																																								
0.050	MP5088	202010A																																																																								
		10-05-2022																																																																								
0.080	MP5089	202010B																																																																								
		10-05-2022																																																																								
0.200	MP5090	202010D																																																																								
		10-06-2022																																																																								
0.080 DGS	N/A	AG113403																																																																								
		05-14-2023																																																																								
Maintenance	By																																																																									
☐ Battery Replacement																																																																										
☐ Dry Gas Regulator Replacement																																																																										
☐ Breath Tube Replacement																																																																										
☐ Other _____																																																																										
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1021</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>MP5086</td></tr> <tr><td>Interferent</td><td>MP5087</td></tr> <tr><td>0.050</td><td>MP5088</td></tr> <tr><td>0.080</td><td>MP5089</td></tr> <tr><td>0.200</td><td>MP5090</td></tr> </tbody> </table> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Attachments</th> </tr> </thead> <tbody> <tr> <td>☒ Form 41</td> <td>☐ Post-Stability Checks</td> </tr> <tr> <td>☒ Stability Checks</td> <td>☐ Flow Calibration</td> </tr> <tr> <td>☒ Calibration Certificate</td> <td>☐ Form 40</td> </tr> <tr> <td>☐ Calibration Adjustment</td> <td>☐ Other _____</td> </tr> </tbody> </table> <table border="1" style="width:100%; border-collapse: collapse;"> <tbody> <tr> <td> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use </td> </tr> </tbody> </table> Richard A Williams Digitally signed by Richard A Williams Date: 2022.02.03 16:30:12 -0500 2022.02.03 13:51:11 Tech Review / Date Admin Review / Date		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	Attachments		☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other _____	☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use
Simulator	Serial #	Lot #	Expiration																																																																							
0.000		N/A	N/A																																																																							
0.040																																																																										
0.100																																																																										
0.200																																																																										
0.300																																																																										
0.080 DGS	N/A																																																																									
Simulator	Serial #	Lot #	Expiration																																																																							
0.050																																																																										
0.080																																																																										
0.200																																																																										
0.080 DGS	N/A																																																																									
Simulator	Serial Number																																																																									
0.000	MP5086																																																																									
Interferent	MP5087																																																																									
0.050	MP5088																																																																									
0.080	MP5089																																																																									
0.200	MP5090																																																																									
Attachments																																																																										
☒ Form 41	☐ Post-Stability Checks																																																																									
☒ Stability Checks	☐ Flow Calibration																																																																									
☒ Calibration Certificate	☐ Form 40																																																																									
☐ Calibration Adjustment	☐ Other _____																																																																									
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																																										

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEON COUNTY SO
Time of Inspection: 11:32

Date of Inspection: 01/31/2022

Serial Number: 80-000957
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.051	0.081	0.201	0.081
0.000	0.050	0.081	0.201	0.080
0.000	0.051	0.081	0.202	0.080
0.000	0.051	0.081	0.201	0.079
0.000	0.051	0.081	0.202	0.079
0.000	0.050	0.081	0.201	0.079
0.000	0.051	0.082	0.201	0.079
0.000	0.051	0.081	0.201	0.079
0.000	0.051	0.081	0.202	0.079
0.000	0.051	0.081	0.202	0.079


Standard Deviations	0.0004	0.0003	0.0005	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISRAEL SOTO

Signature and Printed Name

01/31/2022
Date

stability checks

LEON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000957
01/28/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:12
Control Test	0.051	15:12
Air Blank	0.000	15:13
Control Test	0.049	15:14
Air Blank	0.000	15:14
Control Test	0.049	15:15
Air Blank	0.000	15:15
Control Test Stats		
Average	0.0497	
Std Dev	0.0012	
Rel Std Dev(%)	2.3249	

Operator's Signature

LEON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000957
01/28/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:17
Control Test	0.079	15:18
Air Blank	0.000	15:18
Control Test	0.080	15:19
Air Blank	0.000	15:19
Control Test	0.079	15:20
Air Blank	0.000	15:21
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

wet

Operator's Signature

LEON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000957
01/28/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:22
Control Test	0.200	15:23
Air Blank	0.000	15:23
Control Test	0.199	15:24
Air Blank	0.000	15:25
Control Test	0.198	15:25
Air Blank	0.000	15:26
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

Operator's Signature

LEON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000957
01/28/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:27
Control Test	0.081	15:28
Air Blank	0.000	15:28
Control Test	0.081	15:28
Air Blank	0.000	15:29
Control Test	0.081	15:29
Air Blank	0.000	15:30
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000957, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000957</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>LEON COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/31/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:32</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel Soto
Date: 2022.01.31 13:34:45 -05'00'

01/31/2022

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 January 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality