

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: APOPKA P.D.
Time of Inspection: 09:51

Date of Inspection: 03/29/2022

Serial Number: 80-000967
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Apopka PD was the original owning agency. Ownership was transferred to Port Richey PD shortly after this Form 40 was generated. MB 03/29/2022

Not determined MB 03/29/2022

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

03/29/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000967	Port Richey PD	03/29/2022	TDG ML

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L	
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083	
✓		✓		✓		✓	
PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 03/29/2022 Software: 8100.27 SN 80-000967		PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 03/29/2022 Software: 8100.27 SN 80-000967		PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 03/29/2022 Software: 8100.27 SN 80-000967		PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 03/29/2022 Software: 8100.27 SN 80-000967	
Test g/210L Time		Test g/210L Time		Test g/210L Time		Test g/210L Time	
Air Blank 0.000 11:07		Air Blank 0.000 11:16		Air Blank 0.000 11:23		Air Blank 0.000 10:56	
Control Test 0.050 11:08		Control Test 0.078 11:16		Control Test 0.195 11:24		Control Test 0.078 10:56	
Air Blank 0.000 11:08		Air Blank 0.000 11:17		Air Blank 0.000 11:24		Air Blank 0.000 10:57	
Control Test 0.048 11:09		Control Test 0.078 11:18		Control Test 0.196 11:25		Control Test 0.079 10:57	
Air Blank 0.000 11:10		Air Blank 0.000 11:18		Air Blank 0.000 11:26		Air Blank 0.000 10:57	
Control Test 0.048 11:10		Control Test 0.078 11:19		Control Test 0.197 11:26		Control Test 0.078 10:58	
Air Blank 0.000 11:11		Air Blank 0.000 11:19		Air Blank 0.000 11:27		Air Blank 0.000 10:58	
Control Test Stats		Control Test Stats		Control Test Stats		Control Test Stats	
Average 0.0487		Average 0.0780		Average 0.1960		Average 0.0783	
Std Dev 0.0012		Std Dev 0.0000		Std Dev 0.0010		Std Dev 0.0006	
Rel Std Dev(%) 2.3727		Rel Std Dev(%) 0.0000		Rel Std Dev(%) 0.5102		Rel Std Dev(%) 0.7370	
ML Operator's Signature		ML Operator's Signature		ML Operator's Signature		ML Operator's Signature	

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT RICHEY PD
Time of Inspection: 14:21

Date of Inspection: 03/29/2022

Serial Number: 80-000967
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.047	0.078	0.197	0.079
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.078	0.196	0.078
0.000	0.048	0.078	0.196	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.077	0.196	0.077
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.078	0.196	0.078


Standard Deviations	0.0003	0.0004	0.0005	0.0007
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

03/29/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000967, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000967</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>PORT RICHEY PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>03/29/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:21</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/29/2022

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-000967

OWNING AGENCY: Port Richey PD

DATE OF DEPARTMENT INSPECTION: 3/29/2022

AGENCY INSPECTOR: Damon Pulaski (FWC)

ADDRESS: 6333 Ridge Road

CITY, STATE, ZIP: Port Richey, FL 34668

TELEPHONE NUMBER: 352-398-9945 (Damon Pulaski, FWC)

FAX NUMBER: _____

EMAIL ADDRESS (if available): Damon.Pulaski@myfwc.com

For Program Office Use Only:



Registration Issued



Instrument Added to Evidentiary Instrument Database



Instrument Added to Monthly Statistics Database



Contact Information Added to Instrument Database

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000967	Port Richey PD	11/30/2022	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 ≤0.003 of Wet
PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000967 11/30/2022 Software: 8100.27	PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000967 11/30/2022 Software: 8100.27	PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000967 11/30/2022 Software: 8100.27	DGS PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000967 11/30/2022 Software: 8100.27
Test g/210L Time	Test g/210L Time	Test g/210L Time	Test g/210L Time
Air Blank 0.000 11:22	Air Blank 0.000 11:33	Air Blank 0.000 11:49	Air Blank 0.000 11:26
Control Test 0.048 11:22	Control Test 0.078 11:34	Control Test 0.196 11:49	Control Test 0.078 11:27
Air Blank 0.000 11:23	Air Blank 0.000 11:34	Air Blank 0.000 11:50	Air Blank 0.000 11:27
Control Test 0.048 11:24	Control Test 0.077 11:35	Control Test 0.197 11:51	Control Test 0.079 11:27
Air Blank 0.000 11:24	Air Blank 0.000 11:36	Air Blank 0.000 11:51	Air Blank 0.000 11:28
Control Test 0.047 11:25	Control Test 0.077 11:36	Control Test 0.197 11:52	Control Test 0.078 11:28
Air Blank 0.000 11:25	Air Blank 0.000 11:37	Air Blank 0.000 11:52	Air Blank 0.000 11:29
Control Test Stats	Control Test Stats	Control Test Stats	Control Test Stats
Average 0.0477	Average 0.0773	Average 0.1967	Average 0.0783
Std Dev 0.0006	Std Dev 0.0006	Std Dev 0.0006	Std Dev 0.0006
Rel Std Dev(%) 1.2112	Rel Std Dev(%) 0.7466	Rel Std Dev(%) 0.2936	Rel Std Dev(%) 0.7370
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT RICHEY PD
Time of Inspection: 14:09

Date of Inspection: 11/30/2022

Serial Number: 80-000967
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
/ 0.000	0.048	0.078	0.196	0.078
/ 0.000	0.048	0.078	0.196	0.078
/ 0.000	0.048	0.077	0.196	0.078
/ 0.000	0.049	0.078	0.196	0.078
/ 0.000	0.048	0.078	0.197	0.078
/ 0.000	0.048	0.077	0.197	0.078
/ 0.000	0.048	0.078	0.196	0.078
/ 0.000	0.048	0.077	0.196	0.078
/ 0.000	0.048	0.078	0.196	0.078
/ 0.000	0.048	0.078	0.196	0.077

Standard Deviations	0.0003	0.0004	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

00: Ambient Fail. UNKNOWN. AIRED OUT ROOM FOR ABOUT 10 MIN.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 TAYLOR D GUTSCHOW

Signature and Printed Name

11/30/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000967, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000967</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PORT RICHEY PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/30/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:09</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/30/2022

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1