

INSTRUMENT PROCESSING SHEET

Agency Indiatlantic Police Department

S/N 80-001006

Florida Department of
Law Enforcement

Date In 08/01/2022 DI Completion Date 08/02/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By _____ DERR	Quality Checks	By _____ DERR	Date <u>8/2/2022</u>	Flow Calibration	By _____	Date _____																					
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Keyboard ☒ Feet ☒ Ports ☒ Handle ☒ Dry Gas Shelf ☒ Breath Tube ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>194</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.183</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.539</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks	Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____																									
	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td rowspan="2">MP5092</td> <td>202201C</td> </tr> <tr> <td></td> <td>01/11/2024</td> </tr> <tr> <td>0.080</td> <td rowspan="2">MP5093</td> <td>202201D</td> </tr> <tr> <td></td> <td>01/18/2024</td> </tr> <tr> <td>0.200</td> <td rowspan="2">MP5094</td> <td>202201E</td> </tr> <tr> <td></td> <td>01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>00521080A2</td> </tr> <tr> <td></td> <td></td> <td>02/05/2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202201C		01/11/2024	0.080	MP5093	202201D		01/18/2024	0.200	MP5094	202201E		01/18/2024	0.080 DGS	N/A	00521080A2			02/05/2023			
Simulator	Serial #	Lot #/Exp																										
0.050	MP5092	202201C																										
		01/11/2024																										
0.080	MP5093	202201D																										
		01/18/2024																										
0.200	MP5094	202201E																										
		01/18/2024																										
0.080 DGS	N/A	00521080A2																										
		02/05/2023																										

Calibration Adjustment				By <u>DERR</u>	
Barometric Pressure Gauge <u>1019</u> ID # <u>68639</u>					
Simulator	Serial #	Lot #	Expiration		
0.000	MP5099	N/A	N/A		
0.040	MP5096	21070	03/01/2023		
0.100	MP5098	21080	03/08/2023		
0.200	MP5100	20510	12/03/2022		
0.300	MP5101	21030	02/02/2023		
0.080 DGS	N/A	AG115904	06/08/2023		
☒ Post Calibration Adjustment Stability Checks					
Simulator	Serial #	Lot #	Expiration		
0.050	MP5092	202201C	01/11/2024		
0.080	MP5093	202201D	01/18/2024		
0.200	MP5094	202201E	01/18/2024		
0.080 DGS	N/A	00521080A2	02/05/2023		

Department Inspection		By <u>DERR</u>	
Barometric Pressure ID# <u>26932</u>			
Gauge <u>1019</u>		Instrument <u>1019</u>	
Mouth Alcohol Solution Lot # <u>2021-D</u>			
Acetone Stock Solution Lot # <u>2021-C</u>			
Simulator	Serial Number		
0.000	MP5095		
Interferent	MP5097		
0.050	MP5092		
0.080	MP5093		
0.200	MP5094		
Attachments			
☒ Form 41		☒ Post-Stability Checks	
☒ Stability Checks		☐ Flow Calibration	
☒ Calibration Certificate		☐ Form 40	
☒ Calibration Adjustment		☐ Other _____	

Notes/Suggested Service: Instrument was calibrated to to bring values closer to nominal. During the first optical calibration data entry the 0.040 value was omitted during data entry, so the calibration was aborted and started over from the beginning. DERR

☒ Instrument Complies with Chapter 11D-8, FAC
☐ Instrument Does Not Comply with Chapter 11D-8, FAC

☒ Return to/Place into Evidentiary Use
☐ Remain Out of Evidentiary Use

☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor Gutschow

Digitally signed by Taylor
Gutschow
Date: 2022.08.02 14:21:25 -04'

Tech Review / Date

Admin Review ^{-04'00'} Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001006, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001006</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>INDIALANTIC P.D.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/02/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:28</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/02/2022

Date

DAVID E REYES-RIVERA,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIALANTIC P.D.
Time of Inspection: 13:28

Date of Inspection: 08/02/2022

Serial Number: 80-001006
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.050	0.078	0.200	0.079
0.000	0.050	0.078	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.078

Standard Deviations	0.0005	0.0004	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

08/02/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-001006	Indiatlantic Police Department	8/2/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒
INDIANLANTIC P.D. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001006 08/02/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:37 Control Test 0.049 11:38 Air Blank 0.000 11:38 Control Test 0.049 11:39 Air Blank 0.000 11:39 Control Test 0.049 11:40 Air Blank 0.000 11:41 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>[Signature]</i> Operator's Signature	INDIANLANTIC P.D. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001006 08/02/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:41 Control Test 0.078 11:42 Air Blank 0.000 11:43 Control Test 0.077 11:43 Air Blank 0.000 11:44 Control Test 0.078 11:45 Air Blank 0.000 11:45 Control Test Stats Average 0.0777 Std Dev 0.0006 Rel Std Dev(%) 0.7434 <i>[Signature]</i> Operator's Signature	INDIANLANTIC P.D. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001006 08/02/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:46 Control Test 0.200 11:47 Air Blank 0.000 11:47 Control Test 0.199 11:48 Air Blank 0.000 11:49 Control Test 0.197 11:49 Air Blank 0.000 11:50 Control Test Stats Average 0.1987 Std Dev 0.0015 Rel Std Dev(%) 0.7689 <i>[Signature]</i> Operator's Signature	INDIANLANTIC P.D. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001006 08/02/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:51 Control Test 0.079 11:51 Air Blank 0.000 11:52 Control Test 0.080 11:52 Air Blank 0.000 11:53 Control Test 0.079 11:53 Air Blank 0.000 11:53 Control Test Stats Average 0.0793 Std Dev 0.0006 Rel Std Dev(%) 0.7277 <i>[Signature]</i> Operator's Signature

Auto Calibration

Max Power Res Value = 116
Auto Range Res Value = 85

Sol Value = 0.000 g/210L ***
Fit Value = 0.000 mg/l ****
Samples Taken = 4, Discarded = 1
Sum Io = 12657, Sum Io = 13346

Sample % Abs (% Abs Ref)
Sample #1 = 1.5470 (-0.0170)
Sample #2 = 1.5050 (-0.0150)
Sample #3 = 1.4860 (-0.0220)
Sample #4 = 1.4970 (-0.0260)
Avg % Abs = 1.4930 (-0.0243)
STD DEV = 0.0113 (-0.0086)
REL STD DEV = 0.755 (35.432)

INSTRUMENT P.C.
Intoxilyzer - Micro Analyzer
Model 8000
SN 81-001015
08-02/2022

Auto Calibration

Max Power Res Value = 116
Auto Range Res Value = 85

Sol Value = 0.000 g/210L ***

Fit Value = 0.000 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12657, Sum Io = 13346

Sample % Abs (% Abs Ref)

Sample #1 = 1.1020 (-0.0180)

Sample #2 = 1.1290 (-0.0260)

Sample #3 = 1.1290 (-0.0260)

Sample #4 = 1.1030 (-0.0390)

Avg % Abs = 1.1203 (-0.0203)

STD DEV = 0.0150 (-0.0235)

REL STD DEV = 12.45 (-115.373)

Channel 2

Sample % Abs (% Abs Ref)

Sample #1 = 0.1060 (-0.0090)

Sample #2 = 1.1020 (-0.0260)

Sample #3 = 1.1170 (-0.0070)

Sample #4 = 0.0990 (-0.0090)

Avg % Abs = 1.1060 (-0.0063)

STD DEV = 0.0096 (-0.0174)

REL STD DEV = 9.096 (-274.543)

Sol Value = 0.000 g/210L ***

Fit Value = 0.000 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12657, Sum Io = 13346

Sample % Abs (% Abs Ref)

Sample #1 = 0.0930 (-0.0120)

Sample #2 = 0.0870 (-0.0260)

Sample #3 = 0.0790 (-0.0130)

Sample #4 = 0.0730 (-0.0190)

Avg % Abs = 0.0877 (-0.0193)

STD DEV = 0.0081 (-0.0055)

REL STD DEV = 0.921 (-33.654)

Channel 2

Sample % Abs (% Abs Ref)
Sample #1 = 6.8510 (0.0020)
Sample #2 = 6.7530 (0.0020)
Sample #3 = 6.7640 (0.0040)
Sample #4 = 6.7430 (0.0080)
Avg % Abs = 6.7553 (0.0080)
STD DEV = 0.0105 (0.0067)
REL STD DEV = 0.156 (8.896)

Sol Value = 0.300 g/210L ***

Fit Value = 1.4286 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12653, Sum Io = 13338

Sample % Abs (% Abs Ref)

Sample #1 = 5.5360 (0.0040)

Sample #2 = 5.4470 (0.0070)

Sample #3 = 5.4330 (0.0090)

Sample #4 = 5.4570 (0.0090)

Avg % Abs = 5.457 (0.0092)

STD DEV = 0.0121 (0.0160)

REL STD DEV = 0.221 (17.374)

Channel 2

Sample % Abs (% Abs Ref)
Sample #1 = 9.9010 (0.0100)
Sample #2 = 9.8120 (0.0240)
Sample #3 = 9.7740 (0.0200)
Sample #4 = 9.7510 (0.0600)
Avg % Abs = 9.7990 (0.0453)
STD DEV = 0.0308 (0.0189)
REL STD DEV = 0.315 (13.007)

Sol Value = 0.220 g/210L ***

Fit Value = 0.9524 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12653, Sum Io = 13341

Sample % Abs (% Abs Ref)

Sample #1 = 3.8140 (-0.0150)

Sample #2 = 3.7480 (0.0470)

Sample #3 = 3.7450 (0.0560)

Sample #4 = 3.7900 (0.0350)

Avg % Abs = 3.7443 (0.0466)

STD DEV = 0.0660 (0.0103)

REL STD DEV = 0.161 (22.964)

Optical Calibration 2

SN: 80-001006
Agency: Indiantantic PD
Date: 8/2/2022
Quadratic Fit: +/- 0.002g/210L
By: DERR

Auto Cal Data

Channel 1
Sol Val = 0.000 mg/l or 0.000 g/210L
% Abs = 0.120
Std Dev = 0.02 Rel Std Dev = 12.47
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.076
Std Dev = 0.01 Rel Std Dev = 0.92
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.945
Std Dev = 0.01 Rel Std Dev = 0.39
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.744
Std Dev = 0.01 Rel Std Dev = 0.16
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.446
Std Dev = 0.01 Rel Std Dev = 0.22
Zero Order Coef = -311.43
First Order Coef = 2535.24
Second Order Coef = 26.39
Standard Deviation = 29.676660

Sol Value = 0.000 g/210L ***

Fit Value = 0.3810 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12653, Sum Io = 13338

Sample % Abs (% Abs Ref)

Sample #1 = 2.9010 (0.0100)

Sample #2 = 2.9400 (0.0100)

Sample #3 = 2.9030 (0.0100)

Sample #4 = 2.9000 (0.0100)

Average Result = 2.915.6667

STD DEV = 24.9832

REL STD DEV = 0.843

Channel 2

Sol Val = 0.000 mg/l or 0.000 g/210L
% Abs = 0.106
Std Dev = 0.01 Rel Std Dev = 9.10
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.493
Std Dev = 0.01 Rel Std Dev = 0.75
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.510
Std Dev = 0.02 Rel Std Dev = 0.49
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.753
Std Dev = 0.01 Rel Std Dev = 0.16
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.779
Std Dev = 0.03 Rel Std Dev = 0.32
Zero Order Coef = -134.03
First Order Coef = 1342.86
Second Order Coef = 13.38
Standard Deviation = 15.291855

Sol Value = 0.000 g/210L ***

Fit Value = 0.3810 mg/l ****

Samples Taken = 4, Discarded = 1

Sum Io = 12653, Sum Io = 13338

Sample % Abs (% Abs Ref)

Sample #1 = 2.9010 (0.0100)

Sample #2 = 2.9400 (0.0100)

Sample #3 = 2.9030 (0.0100)

Sample #4 = 2.9000 (0.0100)

Average Result = 2.915.6667

STD DEV = 24.9832

REL STD DEV = 0.843

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.000
0.040 0.041 -0.0006
0.100 0.099 0.0009
0.200 0.201 -0.0006
0.300 0.300 0.0002

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1019
3 um H2O Adjust (mg/l x 10,000) = 894
9 um H2O Adjust (mg/l x 10,000) = 396
**** AUTO CAL PASS

INDIATLANTIC P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-01106
08/02/2022 10:54:30

Auto Calibration
Max Power Res Value = 120
Auto Range Res Value = 84

xxxx AUTO CAL FAIL

Optical Calibration 1	
SN:	80-001006
Agency: Indiatlantic PD	
Date:	8/2/2022
Quadratic Fit: +/- 0.002g/210L	
By:	DERR <i>[Signature]</i>

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001006	Indiatlantic Police Department	8/2/2022	DERR

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒
INDIANTLANTIC P.D. Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001006 08/02/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 07:56 Control Test 0.048 07:57 Air Blank 0.000 07:58 Control Test 0.048 07:59 Air Blank 0.000 07:59 Control Test 0.047 07:59 Air Blank 0.000 08:00 Control Test Stats Average 0.0477 Std Dev 0.0006 Rel Std Dev(%) 1.2112 Operator's Signature 	INDIANTLANTIC P.D. Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001006 08/02/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:02 Control Test 0.076 08:02 Air Blank 0.000 08:03 Control Test 0.077 08:04 Air Blank 0.000 08:04 Control Test 0.077 08:05 Air Blank 0.000 08:05 Control Test Stats Average 0.0767 Std Dev 0.0006 Rel Std Dev(%) 0.7531 Operator's Signature 	INDIANTLANTIC P.D. Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001006 08/02/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:07 Control Test 0.200 08:08 Air Blank 0.000 08:08 Control Test 0.198 08:09 Air Blank 0.000 08:09 Control Test 0.198 08:10 Air Blank 0.000 08:11 Control Test Stats Average 0.1997 Std Dev 0.0012 Rel Std Dev(%) 0.5812 Operator's Signature 	INDIANTLANTIC P.D. Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001006 08/02/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:12 Control Test 0.080 08:13 Air Blank 0.000 08:13 Control Test 0.080 08:13 Air Blank 0.000 08:14 Control Test 0.080 08:14 Air Blank 0.000 08:15 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 

Instruments 0.08g/210L values are just out of the acceptable range for the Department Inspection. Will calibrate the instrument to bring values closer to nominal. DERR