



INSTRUMENT PROCESSING SHEET

Agency Brevard County Sheriff's OfficeS/N 80-001010Florida Department of
Law EnforcementDate In 11-3-2022 DI Completion Date 11-4-2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By IS	Quality Checks	By IS	Date <u>11-3-2022</u>	Flow Calibration	By	Date																																							
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>230</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C 01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202201D 01-18-2024</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E 01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>08121080A1 05-05-2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202201D 01-18-2024	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	08121080A1 05-05-2023																													
Simulator	Serial #	Lot #/Exp																																												
0.050	MP6291	202201C 01-11-2024																																												
0.080	MP6292	202201D 01-18-2024																																												
0.200	MP6293	202201E 01-18-2024																																												
0.080 DGS	N/A	08121080A1 05-05-2023																																												
Calibration Adjustment		By _____		Department Inspection			By IS _____																																							
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28427</u> Gauge <u>1016</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP6289																																													
Interferent	MP6290																																													
0.050	MP6291																																													
0.080	MP6292																																													
0.200	MP6293																																													
☐ Post Calibration Adjustment Stability Checks				Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					<table border="1"><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☒ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></tbody></table>			☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☒ Form 40	☐ Calibration Adjustment	☐ Other _____												
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
☒ Form 41	☐ Post-Stability Checks																																													
☒ Stability Checks	☐ Flow Calibration																																													
☒ Calibration Certificate	☒ Form 40																																													
☐ Calibration Adjustment	☐ Other _____																																													
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2022.11.08 Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2022.11.04 09:36:31 -0400 Tech Review / Date _____ Admin Review / Date <u>09:27:37</u> -05'00"																																										

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.
Time of Inspection: 09:05

Date of Inspection: 11/04/2022

Serial Number: 80-001010
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08121080A1 Exp: 05/05/2023
0.000	0.051	0.081	0.207	0.078
0.000	0.052	0.081	0.207	0.078
0.000	0.051	0.081	0.207	0.078
0.000	0.051	0.081	0.207	0.078
0.000	0.051	0.081	0.206	0.078
0.000	0.052	0.081	0.206	0.078
0.000	0.051	0.081	0.206	0.078
0.000	0.052	0.081	0.206	0.078
0.000	0.051	0.081	0.206	0.079
0.000	0.051	0.081	0.205	0.078

Standard Deviations	0.0004	0.0000	0.0006	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

11/04/2022
Date

Stability checks

BREVARD COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001010
11/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:19
Control Test	0.052	14:20
Air Blank	0.000	14:20
Control Test	0.051	14:21
Air Blank	0.000	14:22
Control Test	0.050	14:22
Air Blank	0.000	14:23
Control Test Stats		
Average	0.0510	
Std Dev	0.0010	
Rel Std Dev(%)	1.9608	

Operator's Signature

BREVARD COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001010
11/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:24
Control Test	0.081	14:24
Air Blank	0.000	14:25
Control Test	0.080	14:26
Air Blank	0.000	14:26
Control Test	0.080	14:27
Air Blank	0.000	14:28
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Operator's Signature

BREVARD COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001010
11/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:28
Control Test	0.206	14:29
Air Blank	0.000	14:30
Control Test	0.203	14:30
Air Blank	0.000	14:31
Control Test	0.202	14:32
Air Blank	0.000	14:32
Control Test Stats		
Average	0.2037	
Std Dev	0.0021	
Rel Std Dev(%)	1.0221	

Operator's Signature

wet

BREVARD COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001010
11/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:34
Control Test	0.078	14:34
Air Blank	0.000	14:35
Control Test	0.078	14:35
Air Blank	0.000	14:36
Control Test	0.079	14:36
Air Blank	0.000	14:36
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Operator's Signature

Dry



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001010, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001010</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>BREVARD COUNTY S.O.</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>11/04/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>09:05</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Date: 2022.11.04 09:22:32
-04'00'

Digitally signed by Israel Soto
Date: 2022.11.04 09:22:32
-04'00'

11/04/2022

Date

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.

Time of Inspection: 14:14

Date of Inspection: 11/03/2022

Serial Number: 80-001010

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

N/A compliance not determined

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

11/03/2022
Date