



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001044, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001044</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>COCONUT CREEK PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/13/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>09:00</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/13/2022

Date

David E Reyes-Rivera
DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD
Time of Inspection: 09:00

Date of Inspection: 01/13/2022

Serial Number: 80-001044
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.081	0.200	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.051	0.080	0.200	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.051	0.081	0.201	0.080
0.000	0.051	0.081	0.201	0.079
0.000	0.050	0.081	0.200	0.079


Standard Deviations	0.0004	0.0005	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

DAVID E REYES-RIVERA

01/13/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001044	Coconut Creek Police Department	1/13/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒																																																																																																																																																
COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 01/13/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>06:51</td></tr><tr><td>Control Test</td><td>0.051</td><td>06:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:53</td></tr><tr><td>Control Test</td><td>0.050</td><td>06:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:54</td></tr><tr><td>Control Test</td><td>0.050</td><td>06:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:55</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0503</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1471</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	06:51	Control Test	0.051	06:52	Air Blank	0.000	06:53	Control Test	0.050	06:53	Air Blank	0.000	06:54	Control Test	0.050	06:54	Air Blank	0.000	06:55	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel Std Dev(%)	1.1471		COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 01/13/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>06:56</td></tr><tr><td>Control Test</td><td>0.081</td><td>06:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:57</td></tr><tr><td>Control Test</td><td>0.080</td><td>06:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:58</td></tr><tr><td>Control Test</td><td>0.080</td><td>06:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:00</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	06:56	Control Test	0.081	06:57	Air Blank	0.000	06:57	Control Test	0.080	06:58	Air Blank	0.000	06:58	Control Test	0.080	06:59	Air Blank	0.000	07:00	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187		COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 01/13/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:02</td></tr><tr><td>Control Test</td><td>0.200</td><td>07:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:03</td></tr><tr><td>Control Test</td><td>0.201</td><td>07:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:04</td></tr><tr><td>Control Test</td><td>0.200</td><td>07:05</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:05</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2003</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2882</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	07:02	Control Test	0.200	07:02	Air Blank	0.000	07:03	Control Test	0.201	07:04	Air Blank	0.000	07:04	Control Test	0.200	07:05	Air Blank	0.000	07:05	Control Test Stats			Average	0.2003		Std Dev	0.0006		Rel Std Dev(%)	0.2882		COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 01/13/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:07</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:07</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:08</td></tr><tr><td>Control Test</td><td>0.081</td><td>07:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:09</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0012</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.4494</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	07:07	Control Test	0.079	07:07	Air Blank	0.000	07:07	Control Test	0.079	07:08	Air Blank	0.000	07:08	Control Test	0.081	07:08	Air Blank	0.000	07:09	Control Test Stats			Average	0.0797		Std Dev	0.0012		Rel Std Dev(%)	1.4494	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:51																																																																																																																																																	
Control Test	0.051	06:52																																																																																																																																																	
Air Blank	0.000	06:53																																																																																																																																																	
Control Test	0.050	06:53																																																																																																																																																	
Air Blank	0.000	06:54																																																																																																																																																	
Control Test	0.050	06:54																																																																																																																																																	
Air Blank	0.000	06:55																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0503																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1471																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:56																																																																																																																																																	
Control Test	0.081	06:57																																																																																																																																																	
Air Blank	0.000	06:57																																																																																																																																																	
Control Test	0.080	06:58																																																																																																																																																	
Air Blank	0.000	06:58																																																																																																																																																	
Control Test	0.080	06:59																																																																																																																																																	
Air Blank	0.000	07:00																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:02																																																																																																																																																	
Control Test	0.200	07:02																																																																																																																																																	
Air Blank	0.000	07:03																																																																																																																																																	
Control Test	0.201	07:04																																																																																																																																																	
Air Blank	0.000	07:04																																																																																																																																																	
Control Test	0.200	07:05																																																																																																																																																	
Air Blank	0.000	07:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2003																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2882																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:07																																																																																																																																																	
Control Test	0.079	07:07																																																																																																																																																	
Air Blank	0.000	07:07																																																																																																																																																	
Control Test	0.079	07:08																																																																																																																																																	
Air Blank	0.000	07:08																																																																																																																																																	
Control Test	0.081	07:08																																																																																																																																																	
Air Blank	0.000	07:09																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	1.4494																																																																																																																																																		
 Operator's Signature	 Operator's Signature	 Operator's Signature	 Operator's Signature																																																																																																																																																

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD
Time of Inspection: 06:50

Date of Inspection: 01/13/2022

Serial Number: 80-001044
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____


Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA

Signature and Printed Name

01/13/2022
Date