



INSTRUMENT PROCESSING SHEET

Agency Lakeland PDS/N 80-001057Florida Department of
Law EnforcementDate In 5/23/2022 DI Completion Date 6/9/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>6/9/2022</u>	Flow Calibration	By	Date																												
☒ Annual ☒ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>AI would like to re-register this deregistered instrument and put into evidential use.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>210</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
					Maintenance By <u>TDG</u> ☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																														
Calibration Adjustment By _____				Department Inspection By <u>TDG</u>																															
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28199</u>																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1016</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2021-C</u>			
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
☐ Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>SD1014</td></tr><tr><td>Interferent</td><td>SD1017</td></tr><tr><td>0.050</td><td>MP4863</td></tr><tr><td>0.080</td><td>MP4864</td></tr><tr><td>0.200</td><td>MP5097</td></tr></tbody></table>				Simulator	Serial Number	0.000	SD1014	Interferent	SD1017	0.050	MP4863	0.080	MP4864	0.200	MP5097																
Simulator	Serial Number																																		
0.000	SD1014																																		
Interferent	SD1017																																		
0.050	MP4863																																		
0.080	MP4864																																		
0.200	MP5097																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☒ Form 40 ☐ Calibration Adjustment ☒ Other <u>Form 47</u>											
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Notes/Suggested Service: <u>Checked breath tube screen and replaced o-rings on 5/27. Replaced battery and verified setup on 6/8. Added static bag for shipment back to the agency (TDG)</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
				Israel Soto Digitally signed by Israel Soto Date: 2022.06.10 12:37:49 +04'00' <u>2022.06.13</u> 09:06:33 Tech Review / Date Admin Review / Date																															



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-001057

OWNING AGENCY: Lakeland Police Department

DATE OF DEPARTMENT INSPECTION: 6/9/2022

AGENCY INSPECTOR: Camilo Almeida

ADDRESS: 219 North Massachusetts Avenue

CITY, STATE, ZIP: Lakeland, FL 33801

TELEPHONE NUMBER: 813-600-8985

FAX NUMBER: n/a

EMAIL ADDRESS (if available): Camilo.Almeida@lakelandgov.net

For Program Office Use Only:



Registration Issued



Instrument Added to Evidentiary Instrument Database



Instrument Added to Monthly Statistics Database



Contact Information Added to Instrument Database

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 11:57

Date of Inspection: 06/09/2022

Serial Number: 80-001057
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:


AI NOT CONDUCTED. BYPASSED TO OPERATE INSTRUMENT

Not determined

MG
06/09/2022

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/09/2022
Date

LAKELAND PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001057
06/09/2022
Software: 8100.27

Test	g/210L	Time
Air Blānk	0.000	12:08
Control Test	0.051	12:09
Air Blank	0.000	12:10
Control Test	0.050	12:10
Air Blank	0.000	12:11
Control Test	0.050	12:12
Air Blank	0.000	12:12
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	



Operator's Signature

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001057	Lakeland PD	06/09/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																			
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																			
✓	✓	✓	✓																																																																																																			
Inadvertently printed to external printer. Results within the acceptable range. Will attach. <i>ML</i> 06/09/2022	LAKELAND PD Intoxilyzer - Alcohol Analyzer Model 8000 06/09/2022 Software: 8100.27 SN 80-001057 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:17</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:18</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr> <tr><td>Control Test Status</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature <i>ML</i>	Air Blank	0.000	12:16	Control Test	0.078	12:17	Air Blank	0.000	12:17	Control Test	0.078	12:18	Air Blank	0.000	12:18	Control Test	0.078	12:19	Air Blank	0.000	12:20	Control Test Status			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		LAKELAND PD Intoxilyzer - Alcohol Analyzer Model 8000 06/09/2022 Software: 8100.27 SN 80-001057 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr> <tr><td>Control Test</td><td>0.197</td><td>12:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:27</td></tr> <tr><td>Control Test</td><td>0.197</td><td>12:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr> <tr><td>Control Test Status</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1973</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2926</td><td></td></tr> </table> Operator's Signature <i>ML</i>	Air Blank	0.000	12:25	Control Test	0.198	12:25	Air Blank	0.000	12:26	Control Test	0.197	12:27	Air Blank	0.000	12:27	Control Test	0.197	12:28	Air Blank	0.000	12:28	Control Test Status			Average	0.1973		Std Dev	0.0006		Rel Std Dev(%)	0.2926		<i>ML</i> LAKELAND PD Intoxilyzer - Alcohol Analyzer Model 8000 06/09/2022 Software: 8100.27 SN 80-001057 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:31</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:32</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr> <tr><td>Control Test Status</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2821</td><td></td></tr> </table> Operator's Signature <i>ML</i>	Air Blank	0.000	12:31	Control Test	0.077	12:31	Air Blank	0.000	12:32	Control Test	0.079	12:32	Air Blank	0.000	12:33	Control Test	0.078	12:33	Air Blank	0.000	12:33	Control Test Status			Average	0.0780		Std Dev	0.0010		Rel Std Dev(%)	1.2821	
Air Blank	0.000	12:16																																																																																																				
Control Test	0.078	12:17																																																																																																				
Air Blank	0.000	12:17																																																																																																				
Control Test	0.078	12:18																																																																																																				
Air Blank	0.000	12:18																																																																																																				
Control Test	0.078	12:19																																																																																																				
Air Blank	0.000	12:20																																																																																																				
Control Test Status																																																																																																						
Average	0.0780																																																																																																					
Std Dev	0.0000																																																																																																					
Rel Std Dev(%)	0.0000																																																																																																					
Air Blank	0.000	12:25																																																																																																				
Control Test	0.198	12:25																																																																																																				
Air Blank	0.000	12:26																																																																																																				
Control Test	0.197	12:27																																																																																																				
Air Blank	0.000	12:27																																																																																																				
Control Test	0.197	12:28																																																																																																				
Air Blank	0.000	12:28																																																																																																				
Control Test Status																																																																																																						
Average	0.1973																																																																																																					
Std Dev	0.0006																																																																																																					
Rel Std Dev(%)	0.2926																																																																																																					
Air Blank	0.000	12:31																																																																																																				
Control Test	0.077	12:31																																																																																																				
Air Blank	0.000	12:32																																																																																																				
Control Test	0.079	12:32																																																																																																				
Air Blank	0.000	12:33																																																																																																				
Control Test	0.078	12:33																																																																																																				
Air Blank	0.000	12:33																																																																																																				
Control Test Status																																																																																																						
Average	0.0780																																																																																																					
Std Dev	0.0010																																																																																																					
Rel Std Dev(%)	1.2821																																																																																																					

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE LAND PD
Time of Inspection: 14:35

Date of Inspection: 06/09/2022

Serial Number: 80-001057
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.050	0.079	0.198	0.077
0.000	0.050	0.079	0.198	0.077
0.000	0.050	0.079	0.198	0.077
0.000	0.050	0.079	0.198	0.077
0.000	0.050	0.079	0.199	0.077
0.000	0.050	0.079	0.199	0.077
0.000	0.050	0.079	0.199	0.078
0.000	0.050	0.080	0.198	0.077
0.000	0.050	0.079	0.198	0.077
0.000	0.050	0.079	0.198	0.077


Standard Deviations	0.0000	0.0003	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/09/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001057, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001057</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>LAKE LAND PD</u>	0.050 g/210 L 0.004
Calibration Date:	<u>06/09/2022</u>	0.080 g/210 L 0.004
Calibration Time:	<u>14:35</u>	0.200 g/210 L 0.007
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/09/2022

Date

Taylor D Gutschow

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality