



INSTRUMENT PROCESSING SHEET

Agency Winter Haven PDS/N 80-001060Florida Department of
Law EnforcementDate In 6/30/2022 DI Completion Date 7/5/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>7/5/2022</u>	Flow Calibration	By	Date																																							
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>226</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP4863</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP4864</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5097</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	202201C 01/11/2024	0.080	MP4864	202201D 01/18/2024	0.200	MP5097	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023																													
Simulator	Serial #	Lot #/Exp																																												
0.050	MP4863	202201C 01/11/2024																																												
0.080	MP4864	202201D 01/18/2024																																												
0.200	MP5097	202201E 01/18/2024																																												
0.080 DGS	N/A	AG115904 06/08/2023																																												
Calibration Adjustment		By _____		Department Inspection			By <u>TDG</u>																																							
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28199</u> Gauge <u>1017</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>SD1014</td></tr><tr><td>Interferent</td><td>SD1017</td></tr><tr><td>0.050</td><td>MP4863</td></tr><tr><td>0.080</td><td>MP4864</td></tr><tr><td>0.200</td><td>MP5097</td></tr></tbody></table>			Simulator	Serial Number	0.000	SD1014	Interferent	SD1017	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	SD1014																																													
Interferent	SD1017																																													
0.050	MP4863																																													
0.080	MP4864																																													
0.200	MP5097																																													
☐ Post Calibration Adjustment Stability Checks				Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																						
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____				Israel Soto Digitally signed by Israel Soto Date: 2022.07.05 17:18:06 +0400 Tech Review / Date _____ 08:58:43 Admin Review / Date _____ 04:00																																										

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001060	Winter Haven	07/05/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
WINTER HAVEN PD Intoxilyzer - Alcohol Analyzer Model 8000 07/05/2022 Software: 8100.27 SN 80-001060 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr><tr><td>Control Test</td><td>0.049</td><td>11:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr><tr><td>Control Test</td><td>0.048</td><td>11:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:25</td></tr><tr><td>Control Test</td><td>0.049</td><td>11:26</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:27</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0487</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr></tbody></table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:23	Control Test	0.049	11:24	Air Blank	0.000	11:24	Control Test	0.048	11:25	Air Blank	0.000	11:25	Control Test	0.049	11:26	Air Blank	0.000	11:27	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		WINTER HAVEN PD Intoxilyzer - Alcohol Analyzer Model 8000 07/05/2022 Software: 8100.27 SN 80-001060 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr><tr><td>Control Test</td><td>0.078</td><td>11:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:37</td></tr><tr><td>Control Test</td><td>0.078</td><td>11:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:38</td></tr><tr><td>Control Test</td><td>0.078</td><td>11:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:39</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0780</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:35	Control Test	0.078	11:36	Air Blank	0.000	11:37	Control Test	0.078	11:37	Air Blank	0.000	11:38	Control Test	0.078	11:38	Air Blank	0.000	11:39	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		WINTER HAVEN PD Intoxilyzer - Alcohol Analyzer Model 8000 07/05/2022 Software: 8100.27 SN 80-001060 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:50</td></tr><tr><td>Control Test</td><td>0.199</td><td>11:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:51</td></tr><tr><td>Control Test</td><td>0.198</td><td>11:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:53</td></tr><tr><td>Control Test</td><td>0.198</td><td>11:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:54</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1983</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr></tbody></table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:50	Control Test	0.199	11:51	Air Blank	0.000	11:51	Control Test	0.198	11:52	Air Blank	0.000	11:53	Control Test	0.198	11:53	Air Blank	0.000	11:54	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		WINTER HAVEN PD Intoxilyzer - Alcohol Analyzer Model 8000 07/05/2022 Software: 8100.27 SN 80-001060 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:36</td></tr><tr><td>Control Test</td><td>0.081</td><td>10:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:37</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr><tr><td>Control Test</td><td>0.081</td><td>10:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:39</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0807</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr></tbody></table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	10:36	Control Test	0.081	10:36	Air Blank	0.000	10:37	Control Test	0.080	10:37	Air Blank	0.000	10:38	Control Test	0.081	10:38	Air Blank	0.000	10:39	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:23																																																																																																																																																	
Control Test	0.049	11:24																																																																																																																																																	
Air Blank	0.000	11:24																																																																																																																																																	
Control Test	0.048	11:25																																																																																																																																																	
Air Blank	0.000	11:25																																																																																																																																																	
Control Test	0.049	11:26																																																																																																																																																	
Air Blank	0.000	11:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:35																																																																																																																																																	
Control Test	0.078	11:36																																																																																																																																																	
Air Blank	0.000	11:37																																																																																																																																																	
Control Test	0.078	11:37																																																																																																																																																	
Air Blank	0.000	11:38																																																																																																																																																	
Control Test	0.078	11:38																																																																																																																																																	
Air Blank	0.000	11:39																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:50																																																																																																																																																	
Control Test	0.199	11:51																																																																																																																																																	
Air Blank	0.000	11:51																																																																																																																																																	
Control Test	0.198	11:52																																																																																																																																																	
Air Blank	0.000	11:53																																																																																																																																																	
Control Test	0.198	11:53																																																																																																																																																	
Air Blank	0.000	11:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:36																																																																																																																																																	
Control Test	0.081	10:36																																																																																																																																																	
Air Blank	0.000	10:37																																																																																																																																																	
Control Test	0.080	10:37																																																																																																																																																	
Air Blank	0.000	10:38																																																																																																																																																	
Control Test	0.081	10:38																																																																																																																																																	
Air Blank	0.000	10:39																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0807																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7157																																																																																																																																																		

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WINTER HAVEN PD
Time of Inspection: 15:43

Date of Inspection: 07/05/2022

Serial Number: 80-001060
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.048	0.078	0.198	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.048	0.078	0.198	0.080
0.000	0.048	0.078	0.198	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.080
0.000	0.050	0.079	0.199	0.081


Standard Deviations	0.0006	0.0005	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/05/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number **80-001060**, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001060	UNCERTAINTY* $\pm$	
Owning Agency:	WINTER HAVEN PD	0.050 g/ 210 L	0.004
Calibration Date:	07/05/2022	0.080 g/ 210 L	0.004
Calibration Time:	15:43	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/05/2022

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Winter Haven PDS/N 80-001060Florida Department of
Law EnforcementDate In 7/25/2022DI Completion Date 8/1/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 8/1/2022	Flow Calibration	By TDG	Date 8/1/2022																														
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: AI reports a DSP Fail and Temp Regulation Fail during the Diagnostic sequence.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>229</u> ☒ Flow Verification (L/s) Flow Column # ATP104 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.156</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # 68639 ☒ Stability Checks			Flow Column # ATP106 ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>230</u> ☒ Post Calibration Verification (L/s) Flow Column # ATP104 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547)																																
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5092</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5093</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5094</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>00521080A2 02/05/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202201C 01/11/2024	0.080	MP5093	202201D 01/18/2024	0.200	MP5094	202201E 01/18/2024	0.080 DGS	N/A	00521080A2 02/05/2023	<table border="1"><thead><tr><th>Maintenance</th><th>By</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other</td><td></td></tr></tbody></table>					Maintenance	By	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other						
Simulator	Serial #	Lot #/Exp																																			
0.050	MP5092	202201C 01/11/2024																																			
0.080	MP5093	202201D 01/18/2024																																			
0.200	MP5094	202201E 01/18/2024																																			
0.080 DGS	N/A	00521080A2 02/05/2023																																			
Maintenance	By																																				
☐ Battery Replacement																																					
☐ Dry Gas Regulator Replacement																																					
☐ Breath Tube Replacement																																					
☐ Other																																					
Calibration Adjustment		Department Inspection																																			
Barometric Pressure Gauge ID #		Barometric Pressure ID# <u>28199</u>																																			
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1020</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>							
Simulator	Serial #	Lot #	Expiration																																		
0.000		N/A	N/A																																		
0.040																																					
0.100																																					
0.200																																					
0.300																																					
0.080 DGS	N/A																																				
☐ Post Calibration Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5095</td></tr><tr><td>Interferent</td><td>SD3966</td></tr><tr><td>0.050</td><td>MP5092</td></tr><tr><td>0.080</td><td>MP5093</td></tr><tr><td>0.200</td><td>MP5094</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP5095	Interferent	SD3966	0.050	MP5092	0.080	MP5093	0.200	MP5094																		
Simulator	Serial Number																																				
0.000	MP5095																																				
Interferent	SD3966																																				
0.050	MP5092																																				
0.080	MP5093																																				
0.200	MP5094																																				
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1"><thead><tr><th>Attachments</th><th></th></tr></thead><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☒ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☒ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other</td></tr></tbody></table>						Attachments		☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☒ Flow Calibration	☒ Calibration Certificate	☒ Form 40	☐ Calibration Adjustment	☐ Other
Simulator	Serial #	Lot #	Expiration																																		
0.050																																					
0.080																																					
0.200																																					
0.080 DGS	N/A																																				
Attachments																																					
☒ Form 41	☐ Post-Stability Checks																																				
☒ Stability Checks	☒ Flow Calibration																																				
☒ Calibration Certificate	☒ Form 40																																				
☐ Calibration Adjustment	☐ Other																																				
Notes/Suggested Service: <u>Since several of the flow values were at the low end of nominal, a discretionary flow cal was performed to bring the values closer to nominal. The DSP Fail and Temp Regulation Fail could not be duplicated in the lab. Manipulation of the breath hose showed stable temperature. If these messages persist at the agency, the AI shall contact me to discuss. (TDG)</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																			
		Israel Soto Digitally signed by Israel Soto Date: 2022.08.02 09:35:18 +0400 13:39:13																																			
		Tech Review / Date <u>04/00</u> Adm'r, Review / Date																																			

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WINTER HAVEN PD
Time of Inspection: 09:58

Date of Inspection: 08/01/2022

Serial Number: 80-001060
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. AI NOT CONDUCTED.

Not determined ^{mg} 08/01/2022

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW
Signature and Printed Name

08/01/2022
Date

80-001060

Flow Cal

08/01/2022

MG

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
08/01/2022
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 6.633

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.746

3: Rate (Liters/min) = 30

SQRT(Diff) = 21.352

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 657

Rounded Intercept = -559408

Correlation = 0.99823

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001060	Winter Haven PD	08/01/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
WINTER HAVEN PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001060 08/01/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:35 Control Test 0.050 10:36 Air Blank 0.000 10:37 Control Test 0.049 10:37 Air Blank 0.000 10:38 Control Test 0.049 10:39 Air Blank 0.000 10:39 Control Test Stats Average 0.0493 Std Dev 0.0006 Rel Std Dev(%) 1.1703	WINTER HAVEN PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001060 08/01/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:40 Control Test 0.079 10:41 Air Blank 0.000 10:41 Control Test 0.079 10:42 Air Blank 0.000 10:42 Control Test 0.078 10:43 Air Blank 0.000 10:44 Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7339	WINTER HAVEN PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001060 08/01/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:45 Control Test 0.199 10:45 Air Blank 0.000 10:46 Control Test 0.198 10:47 Air Blank 0.000 10:47 Control Test 0.198 10:48 Air Blank 0.000 10:48 Control Test Stats Average 0.1983 Std Dev 0.0006 Rel Std Dev(%) 0.2911	WINTER HAVEN PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001060 08/01/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:51 Control Test 0.081 10:51 Air Blank 0.000 10:52 Control Test 0.080 10:52 Air Blank 0.000 10:52 Control Test 0.081 10:53 Air Blank 0.000 10:53 Control Test Stats Average 0.0807 Std Dev 0.0006 Rel Std Dev(%) 0.7157
<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WINTER HAVEN PD
Time of Inspection: 12:58

Date of Inspection: 08/01/2022

Serial Number: 80-001060
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.049	0.078	0.198	0.081
0.000	0.049	0.082	0.199	0.080
0.000	0.049	0.082	0.198	0.081
0.000	0.049	0.082	0.198	0.081
0.000	0.049	0.082	0.198	0.081
0.000	0.049	0.082	0.198	0.081
0.000	0.049	0.083	0.199	0.081
0.000	0.049	0.082	0.199	0.081
0.000	0.049	0.082	0.199	0.081
0.000	0.049	0.082	0.199	0.081


Standard Deviations	0.0000	0.0013	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

08/01/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001060, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001060</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>WINTER HAVEN PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>08/01/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:58</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/01/2022

Date

Taylor D Gutschow,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality