

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: Clay County Sheriff's Office
Time of Inspection: N/A

Date of Inspection: 2022

Serial Number: 80-001071
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume Check: OK			Barometric Pressure Sensor Check: OK		
Alcohol Free Subject Test: 0.000			Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect			Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

Department Inspection was not performed for 2022. Instrument was sent to an authorized repair facility and did not return during the calendar year.

Compliance not determined.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Phil Nicodemo

Signature and Printed Name

12/28/2022

Date