

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: St. Petersburg PD
Time of Inspection: n/a

Date of Inspection: 2022

Serial Number: 80-001078
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume Check: OK			Barometric Pressure Sensor Check: OK		
Alcohol Free Subject Test: 0.000			Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect			Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:
Instrument did not receive a Department Inspection for 2022. It was sent to repair and did not return within the calendar year.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.
I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor Gutschow Digitally signed by Taylor Gutschow
Date: 2022.12.29 14:42:47 -0500'
Taylor D. Gutschow

Signature and Printed Name

12/29/2022
Date