

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 1113	FHP	09/01/2022	TDG M6

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 09/01/2022 Software: 8100.27 SN 80-001113 Test g/210L Time Air Blank 0.000 14:42 Control Test 0.049 14:43 Air Blank 0.000 14:43 Control Test 0.049 14:44 Air Blank 0.000 14:44 Control Test 0.049 14:45 Air Blank 0.000 14:46 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature M6	FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 09/01/2022 Software: 8100.27 SN 80-001113 Test g/210L Time Air Blank 0.000 14:48 Control Test 0.079 14:50 Air Blank 0.000 14:50 Control Test 0.079 14:51 Air Blank 0.000 14:52 Control Test 0.079 14:52 Air Blank 0.000 14:53 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature M6	FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 09/01/2022 Software: 8100.27 SN 80-001113 Test g/210L Time Air Blank 0.000 14:56 Control Test 0.202 14:56 Air Blank 0.000 14:57 Control Test 0.200 14:57 Air Blank 0.000 14:58 Control Test 0.200 14:59 Air Blank 0.000 14:59 Control Test Stats Average 0.2007 Std Dev 0.0012 Rel Std Dev(%) 0.5754 Operator's Signature M6	FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 09/01/2022 Software: 8100.27 SN 80-001113 Test g/210L Time Air Blank 0.000 14:37 Control Test 0.083 14:37 Air Blank 0.000 14:37 Control Test 0.083 14:38 Air Blank 0.000 14:38 Control Test 0.083 14:39 Air Blank 0.000 14:39 Control Test Stats Average 0.0830 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature M6

Comments: Needs an Optical Cal Adjust. M6 9/1/2022

Solution Stats Quadratic Fit Chan 1			
Act	Fit	Residual	
g/210L	g/210L	g/210L	
0.000	0.000	-0.0002	
0.040	0.040	0.0002	
0.100	0.100	0.0002	
0.200	0.200	-0.0003	
0.300	0.300	0.0001	

Solution Stats Quadratic Fit Chan 2			
Act	Fit	Residual	
g/210L	g/210L	g/210L	
0.000	0.000	-0.0001	
0.040	0.040	0.0001	
0.100	0.100	0.0000	
0.200	0.200	-0.0001	
0.300	0.300	0.0000	

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1 *****
Sample #1 = 2894.00
Sample #2 = 2856.00
Sample #3 = 2821.00
Sample #4 = 2851.00
Average Result = 2842.6667
STD DEV = 18.9297
REL STD DEV = 0.666

***** CHANNEL 2 *****
Sample #1 = 3420.00
Sample #2 = 3368.00
Sample #3 = 3392.00
Sample #4 = 3423.00
Average Result = 3401.0000
STD DEV = 19.1572
REL STD DEV = 0.563

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1014
3 um H2O Adjust (mg/l*10,000) = 967
9 um H2O Adjust (mg/l*10,000) = 406
***** AUTO CAL PASS *****

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.139
Std Dev = 0.00 Rel Std Dev = 3.30
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.882
Std Dev = 0.01 Rel Std Dev = 1.42
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.987
Std Dev = 0.01 Rel Std Dev = 0.28
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.795
Std Dev = 0.01 Rel Std Dev = 0.29
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.533
Std Dev = 0.02 Rel Std Dev = 0.35
Zero Order Coef = -341.72
First Order Coef = 2519.11
Second Order Coef = 22.30
Standard Deviation = 11.135437

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.096
Std Dev = 0.01 Rel Std Dev = 14.01
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.426
Std Dev = 0.01 Rel Std Dev = 0.61
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.369
Std Dev = 0.01 Rel Std Dev = 0.25
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.461
Std Dev = 0.01 Rel Std Dev = 0.12
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.391
Std Dev = 0.03 Rel Std Dev = 0.28
Zero Order Coef = -132.30
First Order Coef = 1406.33
Second Order Coef = 13.72
Standard Deviation = 2.933315

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.5170 (-0.0100)
Sample #2 = 6.4660 (0.0470)
Sample #3 = 6.4650 (0.0550)
Sample #4 = 6.4520 (0.0650)
Avg % Abs = 6.4610 (0.0557)
STD DEV = 0.0078 (0.0090)
REL STD DEV = 0.121 (16.201)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12559, Sum Io = 13218
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.5900 (-0.0010)
Sample #2 = 5.5540 (0.0430)
Sample #3 = 5.5310 (0.0660)
Sample #4 = 5.5150 (0.0820)
Avg % Abs = 5.5333 (0.0637)
STD DEV = 0.0196 (0.0196)
REL STD DEV = 0.354 (30.792)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.4920 (0.0020)
Sample #2 = 9.4190 (0.0380)
Sample #3 = 9.3670 (0.1120)
Sample #4 = 9.3660 (0.1080)
Avg % Abs = 9.3907 (0.1010)
STD DEV = 0.0263 (0.0157)
REL STD DEV = 0.280 (15.561)

<<<<< CHANNEL 1 >>>>>
Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12574, Sum Io = 13225
<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 2.0090 (0.0000)
Sample #2 = 1.9930 (0.0310)
Sample #3 = 1.9820 (0.0400)
Sample #4 = 1.9860 (0.0570)
Avg % Abs = 1.9870 (0.0427)
STD DEV = 0.0056 (0.0132)
REL STD DEV = 0.280 (30.946)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.3990 (-0.0040)
Sample #2 = 3.3690 (0.0380)
Sample #3 = 3.3600 (0.0370)
Sample #4 = 3.3770 (0.0300)
Avg % Abs = 3.3687 (0.0350)
STD DEV = 0.0095 (0.0044)
REL STD DEV = 0.252 (12.454)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12566, Sum Io = 13219
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6320 (0.0000)
Sample #2 = 3.6050 (0.0380)
Sample #3 = 3.7960 (0.0550)
Sample #4 = 3.7830 (0.0680)
Avg % Abs = 3.7947 (0.0537)
STD DEV = 0.0111 (0.0150)
REL STD DEV = 0.291 (28.033)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4260 (0.0040)
Sample #2 = 1.4360 (0.0140)
Sample #3 = 1.4220 (0.0130)
Sample #4 = 1.4200 (0.0220)
Avg % Abs = 1.4260 (0.0163)
STD DEV = 0.0087 (0.0049)
REL STD DEV = 0.611 (30.201)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12574, Sum Io = 13225
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 2.0090 (0.0000)
Sample #2 = 1.9930 (0.0310)
Sample #3 = 1.9820 (0.0400)
Sample #4 = 1.9860 (0.0570)
Avg % Abs = 1.9870 (0.0427)
STD DEV = 0.0056 (0.0132)
REL STD DEV = 0.280 (30.946)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.3990 (-0.0040)
Sample #2 = 3.3690 (0.0380)
Sample #3 = 3.3600 (0.0370)
Sample #4 = 3.3770 (0.0300)
Avg % Abs = 3.3687 (0.0350)
STD DEV = 0.0095 (0.0044)
REL STD DEV = 0.252 (12.454)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12566, Sum Io = 13219
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6320 (0.0000)
Sample #2 = 3.6050 (0.0380)
Sample #3 = 3.7960 (0.0550)
Sample #4 = 3.7830 (0.0680)
Avg % Abs = 3.7947 (0.0537)
STD DEV = 0.0111 (0.0150)
REL STD DEV = 0.291 (28.033)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4260 (0.0040)
Sample #2 = 1.4360 (0.0140)
Sample #3 = 1.4220 (0.0130)
Sample #4 = 1.4200 (0.0220)
Avg % Abs = 1.4260 (0.0163)
STD DEV = 0.0087 (0.0049)
REL STD DEV = 0.611 (30.201)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12574, Sum Io = 13225
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 2.0090 (0.0000)
Sample #2 = 1.9930 (0.0310)
Sample #3 = 1.9820 (0.0400)
Sample #4 = 1.9860 (0.0570)
Avg % Abs = 1.9870 (0.0427)
STD DEV = 0.0056 (0.0132)
REL STD DEV = 0.280 (30.946)

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
SN 80-001113
09/14/2022 08:58:06
Auto Calibration
Max Power Res Value = 78
Auto Range Res Value = 54

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12625, Sum Io = 13253
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1440 (-0.0020)
Sample #2 = 0.1440 (0.0380)
Sample #3 = 0.1380 (0.0570)
Sample #4 = 0.1350 (0.0860)
Avg % Abs = 0.1390 (0.0603)
STD DEV = 0.0046 (0.0242)
REL STD DEV = 3.297 (40.066)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1030 (-0.0130)
Sample #2 = 0.1070 (-0.0090)
Sample #3 = 0.1000 (0.0020)
Sample #4 = 0.0810 (0.0240)
Avg % Abs = 0.0960 (0.0057)
STD DEV = 0.0135 (0.0168)
REL STD DEV = 14.014 (296.520)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12590, Sum Io = 13234
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8790 (0.0100)
Sample #2 = 0.8820 (0.0230)
Sample #3 = 0.8940 (0.0270)
Sample #4 = 0.8690 (0.0570)
Avg % Abs = 0.8817 (0.0357)
STD DEV = 0.0125 (0.0186)
REL STD DEV = 1.418 (52.102)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1030 (-0.0130)
Sample #2 = 0.1070 (-0.0090)
Sample #3 = 0.1000 (0.0020)
Sample #4 = 0.0810 (0.0240)
Avg % Abs = 0.0960 (0.0057)
STD DEV = 0.0135 (0.0168)
REL STD DEV = 14.014 (296.520)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12590, Sum Io = 13234
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8790 (0.0100)
Sample #2 = 0.8820 (0.0230)
Sample #3 = 0.8940 (0.0270)
Sample #4 = 0.8690 (0.0570)
Avg % Abs = 0.8817 (0.0357)
STD DEV = 0.0125 (0.0186)
REL STD DEV = 1.418 (52.102)

Optical Calibration	
SN:	80-00 1113
Agency:	FHP
Date:	09/14/2022
Quadratic Fit: +/- 0.002g/210L	✓
By:	TDG <i>MD</i>

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities (Post-Cal)	80-00 1113	FHP	09/14/2022	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
FL HIGHWAY PATROL Intoxilyzer - Alcolon Analyzer Model 8000 09/14/2022 Software: 8100.27 SN 80-001113 Test g/210L Time Air Blank 0.000 12:08 Control Test 0.049 12:08 Air Blank 0.000 12:09 Control Test 0.049 12:10 Air Blank 0.000 12:10 Control Test 0.049 12:11 Air Blank 0.000 12:11 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000	FL HIGHWAY PATROL Intoxilyzer - Alcolon Analyzer Model 8000 09/14/2022 Software: 8100.27 SN 80-001113 Test g/210L Time Air Blank 0.000 12:15 Control Test 0.079 12:16 Air Blank 0.000 12:16 Control Test 0.078 12:17 Air Blank 0.000 12:18 Control Test 0.078 12:18 Air Blank 0.000 12:19 Control Test Stats Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370	FL HIGHWAY PATROL Intoxilyzer - Alcolon Analyzer Model 8000 09/14/2022 Software: 8100.27 SN 80-001113 Test g/210L Time Air Blank 0.000 12:29 Control Test 0.200 12:30 Air Blank 0.000 12:30 Control Test 0.199 12:31 Air Blank 0.000 12:32 Control Test 0.199 12:32 Air Blank 0.000 12:33 Control Test Stats Average 0.1993 Std Dev 0.0005 Rel Std Dev(%) 0.2896	FL HIGHWAY PATROL Intoxilyzer - Alcolon Analyzer Model 8000 09/14/2022 Software: 8100.27 SN 80-001113 Test g/210L Time Air Blank 0.000 12:34 Control Test 0.080 12:34 Air Blank 0.000 12:35 Control Test 0.080 12:35 Air Blank 0.000 12:36 Control Test 0.080 12:36 Air Blank 0.000 12:36 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000
MG Operator's Signature	MG Operator's Signature	MG Operator's Signature	MG Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 15:31

Date of Inspection: 09/14/2022

Serial Number: 80-001113
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.050	0.078	0.200	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.078	0.200	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.080


Standard Deviations	0.0003	0.0004	0.0000	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

09/14/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001113, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001113</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>09/14/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:31</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/14/2022

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality