



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol

S/N 80-001118

Florida Department of
Law Enforcement

Date In 03-29-2022 DI Completion Date 04-01-2022

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>IS</u> ☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ 	Quality Checks By <u>IS</u> Date <u>03-30-2022</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>237</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.484</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202201C 01-11-2024</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202201D 01-18-2024</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202201E 01-18-2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG113403 05-14-2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202201D 01-18-2024	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	AG113403 05-14-2023	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By <u>IS</u> ☐ Battery Replacement ☒ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Digital readout did not match analog readout on dry gas regulator, performed dry gas tare and values still did not match. Performed second dry gas tare after reg. replacement.</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP6291	202201C 01-11-2024																																																											
0.080	MP6292	202201D 01-18-2024																																																											
0.200	MP6293	202201E 01-18-2024																																																											
0.080 DGS	N/A	AG113403 05-14-2023																																																											
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: <u>Replaced dry gas regulator after Quality Checks and prior to Department Inspection IS</u> _____ _____ _____ _____ _____ 	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1014</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP6291</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☒ Form 40 ☐ Calibration Adjustment ☒ Other <u>Form 51</u> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2022.04. Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2022.04.04 09:15:35 -0400 05 15:06:22 Tech Review / Date Admin Review <u>04/01</u> Date	Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	MP6289																																																												
Interferent	MP6290																																																												
0.050	MP6291																																																												
0.080	MP6292																																																												
0.200	MP6293																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Time of Inspection: 12:27

Date of Inspection: 04/01/2022

Serial Number: 80-001118

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.080	0.201	0.081
0.000	0.049	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.049	0.080	0.201	0.080
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.082
0.000	0.050	0.080	0.201	0.081

Standard Deviations	0.0004	0.0003	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Israel Soto ISRAEL SOTO
Signature and Printed Name

04/01/2022
Date

Stability Checks

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001118
03/30/2022
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	09:54
Control Test	0.049	09:55
Air Blank	0.000	09:55
Control Test	0.050	09:56
Air Blank	0.000	09:56
Control Test	0.048	09:57
Air Blank	0.000	09:57
Control Test Stats		
Average	0.0490	
Std Dev	0.0010	
Rel Std Dev(%)	2.0408	


Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001118
03/30/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:58
Control Test	0.079	09:59
Air Blank	0.000	10:00
Control Test	0.080	10:00
Air Blank	0.000	10:01
Control Test	0.080	10:02
Air Blank	0.000	10:02
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

wet


Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001118
03/30/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:06
Control Test	0.201	10:07
Air Blank	0.000	10:07
Control Test	0.202	10:08
Air Blank	0.000	10:08
Control Test	0.201	10:09
Air Blank	0.000	10:10
Control Test Stats		
Average	0.2013	
Std Dev	0.0006	
Rel Std Dev(%)	0.2868	


Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001118
03/30/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:13
Control Test	0.081	10:13
Air Blank	0.000	10:14
Control Test	0.080	10:14
Air Blank	0.000	10:15
Control Test	0.081	10:15
Air Blank	0.000	10:15
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Dry


Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001118, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001118</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>04/01/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:27</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel
Date: 2022.04.01 14:06:50
-04'00'

04/01/2022

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Time of Inspection: 07:24

Date of Inspection: 04/01/2022

Serial Number: 80-001118

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

The above instrument complies (N/A) does not comply (X) with Chapter 11D-8, FAC. *Compliance not determined*

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO
Signature and Printed Name

04/01/2022
Date

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Susan Barge on 09-21-2021

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001118

Bill To Address:

Florida Highway Patrol

Ship to Address:

Alcohol Testing Program - FDLE

Tallahassee, Florida

Reason for Return:

Instrument lost Form 41, when trying to upload forms an error was shown on Cobra V3 and
the instrument entered Disabled Mode showing "Error 016: Contact CMI".

Records on instrument have already been downloaded and instrument is cleared.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Susan Barge

Phone #: 352-620-4701

Email: susanbarga@flhsmv.gov

ATP Contact Name: Israel Soto

ATP Email: israelsoto@fdle.state.fl.us



INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-001118Florida Department of
Law EnforcementDate In 5/4/2022DI Completion Date 05-05-2022☐ Ship ☐ P/U ☒ H/D ☐ CMI ☐ EE

Intake	By PN	Quality Checks	By PN	Date <u>5/4/2022</u>	Flow Calibration	By IS	Date <u>05-04-2022</u>																														
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Instrument arrived with DGS</u> <u>and without power cord.</u> <u>Instrument lost agency codes.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>237</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 103</u> 32 mm <u>.144</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.218*</u> (.228 - .278) 103 mm <u>.476</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks			☒ Flow Column # <u>ATP-102</u> ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>236</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547)																																
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5089</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5090</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG121002 07/29/2023</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202201C 01/11/2024	0.080	MP5089	202201D 01/18/2024	0.200	MP5090	202201E 01/18/2024	0.080 DGS	N/A	AG121002 07/29/2023	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By IS</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td><td></td></tr><tr><td>☒ Other CMI re-uploaded Agency Codes through phone line</td><td></td><td></td></tr></tbody></table>			Maintenance		By IS	☐ Battery Replacement			☐ Dry Gas Regulator Replacement			☐ Breath Tube Replacement			☒ Other CMI re-uploaded Agency Codes through phone line		
Simulator	Serial #	Lot #/Exp																																			
0.050	MP5088	202201C 01/11/2024																																			
0.080	MP5089	202201D 01/18/2024																																			
0.200	MP5090	202201E 01/18/2024																																			
0.080 DGS	N/A	AG121002 07/29/2023																																			
Maintenance		By IS																																			
☐ Battery Replacement																																					
☐ Dry Gas Regulator Replacement																																					
☐ Breath Tube Replacement																																					
☒ Other CMI re-uploaded Agency Codes through phone line																																					
Calibration Adjustment		By			Department Inspection			By IS																													
Barometric Pressure Gauge <u> </u> ID # <u> </u>					Barometric Pressure ID# <u>26932</u>																																
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A						Gauge <u>1014</u> Instrument <u>1013</u>				
Simulator	Serial #	Lot #	Expiration																																		
0.000		N/A	N/A																																		
0.040																																					
0.100																																					
0.200																																					
0.300																																					
0.080 DGS	N/A																																				
☐ Post Calibration Adjustment Stability Checks					Mouth Alcohol Solution Lot # <u>2021-D</u>																																
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A						Acetone Stock Solution Lot # <u>2021-C</u>												
Simulator	Serial #	Lot #	Expiration																																		
0.050																																					
0.080																																					
0.200																																					
0.080 DGS	N/A																																				
Notes/Suggested Service: <u>Agency Codes were uploaded prior to Quality Checks and Department Inspection.</u> <u>Performed a System Check breath test to ensure Agency Codes were uploaded. IS</u>					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5084</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5084																			
Simulator	Serial Number																																				
0.000	MP5086																																				
Interferent	MP5087																																				
0.050	MP5088																																				
0.080	MP5089																																				
0.200	MP5084																																				
					Attachments																																
					☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment		☐ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☒ Other <u>System Check</u>																														
					☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																																
					☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																																
					☒ Conduct an Agency Inspection Before Evidentiary Use																																
					David Eliezer Reyes-Rivera		05 14:35:4																														
					Tech Review / Date		Admin Review / Date																														

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 08:43

Date of Inspection: 05/05/2022

Serial Number: 80-001118
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.079	0.203	0.080
0.000	0.049	0.079	0.202	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.079	0.201	0.080
0.000	0.051	0.079	0.199	0.080
0.000	0.051	0.080	0.200	0.080
0.000	0.051	0.080	0.200	0.080
0.000	0.051	0.081	0.200	0.080
0.000	0.051	0.081	0.201	0.081

Standard Deviations	0.0006	0.0008	0.0011	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISRAEL SOTO

Signature and Printed Name

05/05/2022
Date

Stability Checks 80-001118

FL HIGHWAY PATROL
Intoxilyzer - Alcotest Analyzer
Model: 8000 SN 80-001118
05/04/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:39
Control Test	0.049	12:40
Air Blank	0.000	12:40
Control Test	0.049	12:41
Air Blank	0.000	12:42
Control Test	0.049	12:42
Air Blank	0.000	12:43
Control Test Stats		
Average	0.049	
Std Dev	0.000	
Rel Std Dev(%)	0.000	


Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcotest Analyzer
Model: 8000 SN 80-001118
05/04/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:45
Control Test	0.079	12:45
Air Blank	0.000	12:46
Control Test	0.080	12:47
Air Blank	0.000	12:47
Control Test	0.078	12:48
Air Blank	0.000	12:48
Control Test Stats		
Average	0.079	
Std Dev	0.000	
Rel Std Dev(%)	1.2658	


Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcotest Analyzer
Model: 8000 SN 80-001118
05/04/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:51
Control Test	0.202	12:51
Air Blank	0.000	12:52
Control Test	0.201	12:53
Air Blank	0.000	12:53
Control Test	0.202	12:54
Air Blank	0.000	12:54
Control Test Stats		
Average	0.2017	
Std Dev	0.0006	
Rel Std Dev(%)	0.2863	


Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcotest Analyzer
Model: 8000 SN 80-001118
05/04/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:57
Control Test	0.080	12:58
Air Blank	0.000	12:58
Control Test	0.080	12:58
Air Blank	0.000	12:59
Control Test	0.081	12:59
Air Blank	0.000	13:00
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

D65


Operator's Signature

Flow Calibration 80-001118

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 6090 SN 80-001118
05/04/2022
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SRT(Diff)) = 7.277
2: Rate (Liters/min) = 15
SRT(Diff)) = 11.832
3: Rate (Liters/min) = 30
SRT(Diff)) = 20.949

Dependent Data Scale Factor = 1000.00 L/min

Independent Data Scale Factor = 256

Rounded Slope = 704

Rounded Intercept = -739981

Correlation = 0.99716

DA

FLORIDA DEPARTMENT OF LAW ENFORCEMENT
ALCOHOL TESTING PROGRAM
BREATH ALCOHOL TEST AFFIDAVIT

Instrument Type: Intoxilyzer 8000
Instrument Registered To: FL HIGHWAY PATROL
Instrument Serial Number: 80-001118 Software: 8100.27
Date of Test: 05/04/2022

Date of Last Agency Inspection: 04/08/2022

Observation Period Began: 12:00

Subject's Name: TEST T TEST

DOB: 01/01/1970 Sex: M

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	13:13
	Air Blank	0.000	13:14
	Control Test	0.081	13:14
	Air Blank	0.000	13:14
	Subject Sample #1	0.000	13:15
	Air Blank	0.000	13:15
	Air Blank	0.000	13:17
	Subject Sample #2	0.000	13:18
	Air Blank	0.000	13:18
	Control Test	0.081	13:18
	Air Blank	0.000	13:19
	Diagnostics Check	OK	13:19

System Check
checking Agency Code upload
codes uploaded successfully


Cylinder Lot: 402328118
Exp: 01/20/2025

State of Florida, County of _____,

Personally appeared before me the undersigned authority, who (____) is personally known to me or (____) produced _____ as identification, and who after being placed under oath, states:

I ISRAEL SOTO, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: _____ Date: _____
Signature

Sworn to (or affirmed) before me this _____ day of _____, _____

Signature of Notary Public-State of Florida Printed Name of Notary Public-State of Florida

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001118, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001118</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/05/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>08:43</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel Soto
Date: 2022.05.05 09:58:47 -04'00'

05/05/2022

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-001118Florida Department of
Law EnforcementDate In 10-10-2022 DI Completion Date 10-18-2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By IS	Quality Checks	By IS	Date <u>10-18-2022</u>	Flow Calibration	By	Date																															
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Agency Inspector</u> <u>indicated instrument lost forms.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>233</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.527</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks			☐ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																	
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C 01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202201D 01-18-2024</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E 01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>08121080A1 05-05-2023</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202201D 01-18-2024	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	08121080A1 05-05-2023	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By IS</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td>☒ Other <u>Forms uploaded to</u></td><td></td></tr><tr><td colspan="2"><u>instrument by CMI through phone</u></td><td></td></tr><tr><td colspan="2"><u>line.</u></td><td></td></tr></tbody></table>			Maintenance		By IS	☐ Battery Replacement	☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement	☒ Other <u>Forms uploaded to</u>		<u>instrument by CMI through phone</u>			<u>line.</u>			
Simulator	Serial #	Lot #/Exp																																				
0.050	MP6291	202201C 01-11-2024																																				
0.080	MP6292	202201D 01-18-2024																																				
0.200	MP6293	202201E 01-18-2024																																				
0.080 DGS	N/A	08121080A1 05-05-2023																																				
Maintenance		By IS																																				
☐ Battery Replacement	☐ Dry Gas Regulator Replacement																																					
☐ Breath Tube Replacement	☒ Other <u>Forms uploaded to</u>																																					
<u>instrument by CMI through phone</u>																																						
<u>line.</u>																																						
Calibration Adjustment		By _____			Department Inspection			By IS _____																														
Barometric Pressure Gauge _____ ID # _____					Barometric Pressure ID# <u>30793</u>																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A						Gauge <u>1009</u> Instrument <u>1008</u>					
Simulator	Serial #	Lot #	Expiration																																			
0.000		N/A	N/A																																			
0.040																																						
0.100																																						
0.200																																						
0.300																																						
0.080 DGS	N/A																																					
☐ Post Calibration Adjustment Stability Checks					Mouth Alcohol Solution Lot # <u>2022-A</u>																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A						Acetone Stock Solution Lot # <u>2022-B</u>													
Simulator	Serial #	Lot #	Expiration																																			
0.050																																						
0.080																																						
0.200																																						
0.080 DGS	N/A																																					
					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293																			
Simulator	Serial Number																																					
0.000	MP6289																																					
Interferent	MP6290																																					
0.050	MP6291																																					
0.080	MP6292																																					
0.200	MP6293																																					
Notes/Suggested Service: <u>Forms uploaded by CMI prior to</u>					Attachments																																	
<u>Quality Checks and Department Inspection.</u>					☒ Form 41			☐ Post-Stability Checks																														
<u>Instrument received Interferent Detect on Dry Gas</u>					☒ Stability Checks			☐ Flow Calibration																														
<u>test during Department Inspection. Test was repeated</u>					☒ Calibration Certificate			☐ Form 40																														
<u>with nominal results. Performed extra stability tests with</u>					☐ Calibration Adjustment			☒ Other <u>Extra Stabilities</u>																														
<u>0.080 dry gas standard, all results nominal. IS</u>																																						
					☒ Instrument Complies with Chapter 11D-8, FAC																																	
					☐ Instrument Does Not Comply with Chapter 11D-8, FAC																																	
					☒ Return to/Place into Evidentiary Use																																	
					☐ Remain Out of Evidentiary Use																																	
					☒ Conduct an Agency Inspection Before Evidentiary Use																																	
					David Eliezer Reyes-Rivera			Digitally signed by David Eliezer Reyes-Rivera Date: 2022.10.18 12:48:12 -0400																														
					Tech Review / Date			Admin Review / Date																														

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 09:12

Date of Inspection: 10/18/2022

Serial Number: 80-001118
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08121080A1 Exp: 05/05/2023
0.000	0.048	0.078	0.198	0.081 / 0.082
0.000	0.048	0.078	0.198	0.081 / 0.082
0.000	0.048	0.078	0.198	0.081 / 0.082
0.000	0.048	0.078	0.198	INT / 0.082
0.000	0.049	0.078	0.198	/ 0.082
0.000	0.048	0.078	0.198	/ 0.082
0.000	0.048	0.078	0.199	/ 0.082
0.000	0.048	0.078	0.198	/ 0.082
0.000	0.048	0.079	0.199	/ 0.082
0.000	0.048	0.078	0.198	/ 0.082

Standard Deviations	0.0003	0.0003	0.0004	/ 0.0000
---------------------	--------	--------	--------	----------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

08: Interferent Detect CLEARED AREA.

Cleared area near instrument, repeated test with
nominal results. will perform extra stability tests
JS

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ISRAEL SOTO

Signature and Printed Name

10/18/2022
Date

Stability Checks

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001118
10/18/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:08
Control Test	0.049	07:08
Air Blank	0.000	07:09
Control Test	0.049	07:09
Air Blank	0.000	07:10
Control Test	0.050	07:11
Air Blank	0.000	07:11
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001118
10/18/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:13
Control Test	0.079	07:13
Air Blank	0.000	07:14
Control Test	0.079	07:15
Air Blank	0.000	07:15
Control Test	0.079	07:16
Air Blank	0.000	07:16
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001118
10/18/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:18
Control Test	0.201	07:18
Air Blank	0.000	07:19
Control Test	0.201	07:19
Air Blank	0.000	07:20
Control Test	0.201	07:21
Air Blank	0.000	07:21
Control Test Stats		
Average	0.2010	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

wet

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001118
10/18/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:24
Control Test	0.081	07:24
Air Blank	0.000	07:25
Control Test	0.081	07:25
Air Blank	0.000	07:26
Control Test	0.081	07:26
Air Blank	0.000	07:27
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Dry



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001118, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001118</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/18/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>09:12</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Date: 2022.10.18 11:08:51
-04'00'

10/18/2022

Date

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001118
10/18/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:17
Control Test	0.082	09:17
Air Blank	0.000	09:18
Control Test	0.081	09:18
Air Blank	0.000	09:19
Control Test	0.082	09:19
Air Blank	0.000	09:19
Control Test	0.082	09:20
Air Blank	0.000	09:20
Control Test	0.081	09:20
Air Blank	0.000	09:21
Control Test	0.082	09:21
Air Blank	0.000	09:22
Control Test	0.081	09:22
Air Blank	0.000	09:23
Control Test	0.081	09:23
Air Blank	0.000	09:23
Control Test	0.082	09:24
Air Blank	0.000	09:24
Control Test	0.082	09:25
Air Blank	0.000	09:25
Control Test	0.081	09:25
Air Blank	0.000	09:26
Control Test	0.081	09:26
Air Blank	0.000	09:27
Control Test	0.081	09:27
Air Blank	0.000	09:28
Control Test	0.081	09:28
Air Blank	0.000	09:28
Control Test	0.081	09:29
Air Blank	0.000	09:29
Control Test	0.081	09:30
Air Blank	0.000	09:30
Control Test	0.082	09:30
Air Blank	0.000	09:31
Control Test	0.082	09:31
Air Blank	0.000	09:32
Control Test	0.081	09:32
Air Blank	0.000	09:32
Control Test	0.081	09:33
Air Blank	0.000	09:33
Control Test Stats		
Average	0.0814	
Std Dev	0.0005	
Rel Std Dev(%)	0.6175	

Extra 0.080 dry
gas stability tests

