



INSTRUMENT PROCESSING SHEET

Agency Daytona Beach Police Department

S/N 80-001135

Florida Department of
Law Enforcement



Date In 3/18/2022

DI Completion Date 3/21/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By DERR	Quality Checks	By DERR	Date 3/21/2022	Flow Calibration	By	Date	
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>234</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)			
				Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____				

Calibration Adjustment	By	Department Inspection	By DERR																																																													
Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A				Barometric Pressure ID# <u>28199</u> Gauge <u>1022</u> Instrument <u>1021</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☒ Other <u>Form 48,51</u> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Israel Soto Digitally signed by Israel Soto Date: 2022.03.21 14:06:47 +04'00' 3/21/2022 21:44:21 1 Tech Review / Date _____ Admin. Review / Date _____	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288		
Simulator	Serial #	Lot #	Expiration																																																													
0.000		N/A	N/A																																																													
0.040																																																																
0.100																																																																
0.200																																																																
0.300																																																																
0.080 DGS	N/A																																																															
Simulator	Serial #	Lot #	Expiration																																																													
0.050																																																																
0.080																																																																
0.200																																																																
0.080 DGS	N/A																																																															
Simulator	Serial Number																																																															
0.000	MP6284																																																															
Interferent	MP6285																																																															
0.050	MP6286																																																															
0.080	MP6287																																																															
0.200	MP6288																																																															



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001135, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001135</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>DAYTONA BEACH P.D.</u>	0.050 g/210 L	0.004
Calibration Date:	<u>03/21/2022</u>	0.080 g/210 L	0.004
Calibration Time:	<u>12:52</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/21/2022

Date


DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DAYTONA BEACH P.D.
Time of Inspection: 12:52

Date of Inspection: 03/21/2022

Serial Number: 80-001135
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.078	0.201	0.080
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079

Standard Deviations	0.0003	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

A F / M A:DID NOT PROVIDE M/A SAMPLE.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 Signature and Printed Name

DAVID E REYES-RIVERA

03/21/2022
 Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001135	Daytona Beach Police Department	3/21/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒																																																																																																																																																
DAYTONA BEACH P.D. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001135 03/21/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:41</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:42</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:43</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:45</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0490</td><td></td></tr><tr><td>Std Dev</td><td>0.0002</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	10:41	Control Test	0.049	10:41	Air Blank	0.000	10:42	Control Test	0.049	10:43	Air Blank	0.000	10:43	Control Test	0.049	10:44	Air Blank	0.000	10:45	Control Test Stats			Average	0.0490		Std Dev	0.0002		Rel Std Dev(%)	0.0000		DAYTONA BEACH P.D. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001135 03/21/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:46</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:47</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:48</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:50</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	10:46	Control Test	0.078	10:47	Air Blank	0.000	10:47	Control Test	0.078	10:48	Air Blank	0.000	10:48	Control Test	0.079	10:49	Air Blank	0.000	10:50	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		DAYTONA BEACH P.D. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001135 03/21/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr><tr><td>Control Test</td><td>0.201</td><td>10:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:52</td></tr><tr><td>Control Test</td><td>0.201</td><td>10:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:54</td></tr><tr><td>Control Test</td><td>0.200</td><td>10:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:55</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2007</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2877</td><td></td></tr></table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	10:51	Control Test	0.201	10:52	Air Blank	0.000	10:52	Control Test	0.201	10:53	Air Blank	0.000	10:54	Control Test	0.200	10:54	Air Blank	0.000	10:55	Control Test Stats			Average	0.2007		Std Dev	0.0006		Rel Std Dev(%)	0.2877		DAYTONA BEACH P.D. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001135 03/21/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:57</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:58</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:59</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0793</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr></table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	10:57	Control Test	0.079	10:57	Air Blank	0.000	10:58	Control Test	0.080	10:58	Air Blank	0.000	10:59	Control Test	0.079	10:59	Air Blank	0.000	11:00	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:41																																																																																																																																																	
Control Test	0.049	10:41																																																																																																																																																	
Air Blank	0.000	10:42																																																																																																																																																	
Control Test	0.049	10:43																																																																																																																																																	
Air Blank	0.000	10:43																																																																																																																																																	
Control Test	0.049	10:44																																																																																																																																																	
Air Blank	0.000	10:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0002																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:46																																																																																																																																																	
Control Test	0.078	10:47																																																																																																																																																	
Air Blank	0.000	10:47																																																																																																																																																	
Control Test	0.078	10:48																																																																																																																																																	
Air Blank	0.000	10:48																																																																																																																																																	
Control Test	0.079	10:49																																																																																																																																																	
Air Blank	0.000	10:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:51																																																																																																																																																	
Control Test	0.201	10:52																																																																																																																																																	
Air Blank	0.000	10:52																																																																																																																																																	
Control Test	0.201	10:53																																																																																																																																																	
Air Blank	0.000	10:54																																																																																																																																																	
Control Test	0.200	10:54																																																																																																																																																	
Air Blank	0.000	10:55																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2007																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2877																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:57																																																																																																																																																	
Control Test	0.079	10:57																																																																																																																																																	
Air Blank	0.000	10:58																																																																																																																																																	
Control Test	0.080	10:58																																																																																																																																																	
Air Blank	0.000	10:59																																																																																																																																																	
Control Test	0.079	10:59																																																																																																																																																	
Air Blank	0.000	11:00																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		

Return Material Authorization

Ship to:

☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Justin Dunne on 11/29/2021

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001135

Bill To Address:

Daytona Beach Police Department

Attn: Justin Dunne

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Alcohol Testing Program

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Displays "Err: 012, Contact CMI". Uploaded records prior to sending to repair. No records remain on the instrument.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Justin Dunne

Phone #: 386-316-5594

Email: DunneJustin@DBPD.US

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us



INSTRUMENT PROCESSING SHEET

Agency Daytona Beach PDS/N 80-001135Florida Department of
Law EnforcementDate In 11/29/2021

DI Completion Date _____

☒ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By <u>TDG</u>	Quality Checks	By _____	Date _____	Flow Calibration	By _____	Date _____																														
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>AI Justin Dunne reports an Error 12 and a second error that occurs when he enters the agency code ("Error- Check File Agency"). Agency was unable to upload. I uploaded via direct connect on 11/29/2021.</u>		☐ Breath Tube Screen ☐ Replace External O-Rings ☐ Instrument Set Up Verified ☐ R-Value _____ ☐ Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) ☐ Barometric Pressure Check Gauge ID # _____ ☐ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A		<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By _____</th></tr></thead><tbody><tr><td colspan="2">☐ Battery Replacement</td><td></td></tr><tr><td colspan="2">☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td colspan="2">☐ Breath Tube Replacement</td><td></td></tr><tr><td colspan="2">☐ Other _____</td><td></td></tr></tbody></table>			Maintenance		By _____	☐ Battery Replacement			☐ Dry Gas Regulator Replacement			☐ Breath Tube Replacement			☐ Other _____		
Simulator	Serial #	Lot #/Exp																																			
0.050																																					
0.080																																					
0.200																																					
0.080 DGS	N/A																																				
Maintenance		By _____																																			
☐ Battery Replacement																																					
☐ Dry Gas Regulator Replacement																																					
☐ Breath Tube Replacement																																					
☐ Other _____																																					
		<table border="1"><thead><tr><th colspan="2">DI Temp. Checks</th><th>By _____</th></tr></thead><tbody><tr><td colspan="2">☐ Lab Temp °C</td><td></td></tr><tr><td colspan="2">External Digital Therm. ID#:</td><td></td></tr><tr><td colspan="2">☐ 34°C +- .2 Serial #:</td><td></td></tr><tr><td colspan="2">☐ 34°C +- .2 Serial #:</td><td></td></tr><tr><td colspan="2">☐ 34°C +- .2 Serial #:</td><td></td></tr></tbody></table>			DI Temp. Checks		By _____	☐ Lab Temp °C			External Digital Therm. ID#:			☐ 34°C +- .2 Serial #:			☐ 34°C +- .2 Serial #:			☐ 34°C +- .2 Serial #:																	
DI Temp. Checks		By _____																																			
☐ Lab Temp °C																																					
External Digital Therm. ID#:																																					
☐ 34°C +- .2 Serial #:																																					
☐ 34°C +- .2 Serial #:																																					
☐ 34°C +- .2 Serial #:																																					

Calibration Adjustment	By _____	Department Inspection	By _____																												
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# _____																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A				Gauge _____ Instrument _____	
Simulator	Serial #	Lot #	Expiration																												
0.000		N/A	N/A																												
0.040																															
0.100																															
0.200																															
0.300																															
0.080 DGS	N/A																														
		Mouth Alcohol Solution Lot # _____																													
		Acetone Stock Solution Lot # _____																													
		<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table>	Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200																		
Simulator	Serial Number																														
0.000																															
Interferent																															
0.050																															
0.080																															
0.200																															
☐ Post Calibration Adjustment Stability Checks		<table border="1"><thead><tr><th colspan="2">Attachments</th></tr></thead><tbody><tr><td>☐ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☐ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☐ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☒ Other <u>Form 51</u></td></tr></tbody></table>		Attachments		☐ Form 41	☐ Post-Stability Checks	☐ Stability Checks	☐ Flow Calibration	☐ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☒ Other <u>Form 51</u>																		
Attachments																															
☐ Form 41	☐ Post-Stability Checks																														
☐ Stability Checks	☐ Flow Calibration																														
☐ Calibration Certificate	☐ Form 40																														
☐ Calibration Adjustment	☒ Other <u>Form 51</u>																														

Notes/Suggested Service: Instrument still displays Error 12 after direct-connect upload. Will send to CMI. (TDG)

- | |
|--|
| ☐ Instrument Complies with Chapter 11D-8, FAC |
| ☐ Instrument Does Not Comply with Chapter 11D-8, FAC |
| ☐ Return to/Place into Evidentiary Use |
| ☒ Remain Out of Evidentiary Use |
| ☐ Conduct an Agency Inspection Before Evidentiary Use |

Israel Soto

Digitally signed by Israel Soto
Date: 2022.03.21 14:06:20
+0400

Tech Review / Date _____

Admin Review / Date _____