



INSTRUMENT PROCESSING SHEET

Agency Panama City Police Department

S/N 80-001204

Florida Department of
Law Enforcement

Date In 04-04-2022 DI Completion Date 04-05-2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>IS</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ <u>Agency Inspector indicated</u> <u>instrument is displaying Error</u> <u>Code 12.</u> <u>Keyboard was not working, two</u> <u>pins in the cable appear broken.</u> 	Quality Checks By <u>IS</u> Date <u>04-05-2022</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>236</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.519</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202201C 01-11-2024</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202201D 01-18-2024</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202201E 01-18-2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG113403 05-14-2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202201D 01-18-2024	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	AG113403 05-14-2023	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By <u>IS</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>CMI cleared error 12</u> <u>remotely through phone line prior to</u> <u>Quality Checks and Department</u> <u>Inspection</u> 																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP6291	202201C 01-11-2024																																																											
0.080	MP6292	202201D 01-18-2024																																																											
0.200	MP6293	202201E 01-18-2024																																																											
0.080 DGS	N/A	AG113403 05-14-2023																																																											
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: _____ 	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1007</u> Instrument <u>1004</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP6291</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use David Eliezer Reyes-Rivera Digitally signed by David Eliezer Reyes-Rivera Date: 2022.04.06 07:50:12 -0400 2022.04.06 10:16:1 3-04'00" Tech Review / Date Admin Review / Date	Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	MP6289																																																												
Interferent	MP6290																																																												
0.050	MP6291																																																												
0.080	MP6292																																																												
0.200	MP6293																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PANAMA CITY PD

Time of Inspection: 14:25

Date of Inspection: 04/05/2022

Serial Number: 80-001204

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.079	0.200	0.078

Standard Deviations	0.0003	0.0003	0.0000	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISRAEL SOTO

Signature and Printed Name

04/05/2022
Date

Stability checks

PANAMA CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001204
04/05/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:27
Control Test	0.049	12:28
Air Blank	0.000	12:28
Control Test	0.050	12:29
Air Blank	0.000	12:29
Control Test	0.050	12:30
Air Blank	0.000	12:31
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

Operator's Signature

PANAMA CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001204
04/05/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:32
Control Test	0.079	12:33
Air Blank	0.000	12:33
Control Test	0.080	12:34
Air Blank	0.000	12:35
Control Test	0.080	12:35
Air Blank	0.000	12:36
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Operator's Signature

PANAMA CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001204
04/05/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:37
Control Test	0.200	12:38
Air Blank	0.000	12:38
Control Test	0.200	12:39
Air Blank	0.000	12:40
Control Test	0.200	12:40
Air Blank	0.000	12:41
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

PANAMA CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001204
04/05/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:46
Control Test	0.079	12:46
Air Blank	0.000	12:46
Control Test	0.079	12:47
Air Blank	0.000	12:47
Control Test	0.079	12:48
Air Blank	0.000	12:48
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

wet

Dry



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001204, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001204</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PANAMA CITY PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/05/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:25</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Date: 2022.04.05 15:49:13
-04'00'

04/05/2022
Date
ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PANAMA CITY PD

Time of Inspection: 12:08

Date of Inspection: 04/05/2022

Serial Number: 80-001204

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL ~~X~~ SOTO

Signature and Printed Name

04/05/2022
Date

no middle initial
IS

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PANAMA CITY PD
Time of Inspection: 12:31

Date of Inspection: 05/31/2022

Serial Number: 80-001204
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.080	0.201	0.078
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.079	0.201	0.078
0.000	0.050	0.080	0.202	0.079
0.000	0.051	0.079	0.201	0.079
0.000	0.051	0.079	0.201	0.079
Standard Deviations	0.0004	0.0005	0.0003	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


THOMAS J GRAHAM


 Signature and Printed Name

05/31/2022
 Date

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	80-001204	Panama City Police Department	5/31/2022	PN

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083	≤0.003 of Wet																																																																																																																																															
PANAMA CITY PD Intoxilyzer - Alcohol Analyzer Model 8000 05/31/2022 Software: 8100.27 SN 80-001204 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>08:58</td></tr><tr><td>Control Test</td><td>0.051</td><td>08:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:59</td></tr><tr><td>Control Test</td><td>0.051</td><td>09:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:01</td></tr><tr><td>Control Test</td><td>0.051</td><td>09:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:02</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0510</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	08:58	Control Test	0.051	08:59	Air Blank	0.000	08:59	Control Test	0.051	09:00	Air Blank	0.000	09:01	Control Test	0.051	09:01	Air Blank	0.000	09:02	Control Test Stats			Average	0.0510		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PANAMA CITY PD Intoxilyzer - Alcohol Analyzer Model 8000 05/31/2022 Software: 8100.27 SN 80-001204 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:04</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:05</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:05</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:06</td></tr><tr><td>Control Test</td><td>0.081</td><td>09:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:07</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	09:04	Control Test	0.080	09:05	Air Blank	0.000	09:05	Control Test	0.080	09:06	Air Blank	0.000	09:06	Control Test	0.081	09:07	Air Blank	0.000	09:07	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187		PANAMA CITY PD Intoxilyzer - Alcohol Analyzer Model 8000 05/31/2022 Software: 8100.27 SN 80-001204 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:10</td></tr><tr><td>Control Test</td><td>0.201</td><td>09:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:11</td></tr><tr><td>Control Test</td><td>0.201</td><td>09:11</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:12</td></tr><tr><td>Control Test</td><td>0.201</td><td>09:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:13</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2010</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	09:10	Control Test	0.201	09:10	Air Blank	0.000	09:11	Control Test	0.201	09:11	Air Blank	0.000	09:12	Control Test	0.201	09:13	Air Blank	0.000	09:13	Control Test Stats			Average	0.2010		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PANAMA CITY PD Intoxilyzer - Alcohol Analyzer Model 8000 05/31/2022 Software: 8100.27 SN 80-001204 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:15</td></tr><tr><td>Control Test</td><td>0.078</td><td>09:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:16</td></tr><tr><td>Control Test</td><td>0.079</td><td>09:16</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:17</td></tr><tr><td>Control Test</td><td>0.079</td><td>09:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:17</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	09:15	Control Test	0.078	09:15	Air Blank	0.000	09:16	Control Test	0.079	09:16	Air Blank	0.000	09:17	Control Test	0.079	09:17	Air Blank	0.000	09:17	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:58																																																																																																																																																	
Control Test	0.051	08:59																																																																																																																																																	
Air Blank	0.000	08:59																																																																																																																																																	
Control Test	0.051	09:00																																																																																																																																																	
Air Blank	0.000	09:01																																																																																																																																																	
Control Test	0.051	09:01																																																																																																																																																	
Air Blank	0.000	09:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0510																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:04																																																																																																																																																	
Control Test	0.080	09:05																																																																																																																																																	
Air Blank	0.000	09:05																																																																																																																																																	
Control Test	0.080	09:06																																																																																																																																																	
Air Blank	0.000	09:06																																																																																																																																																	
Control Test	0.081	09:07																																																																																																																																																	
Air Blank	0.000	09:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:10																																																																																																																																																	
Control Test	0.201	09:10																																																																																																																																																	
Air Blank	0.000	09:11																																																																																																																																																	
Control Test	0.201	09:11																																																																																																																																																	
Air Blank	0.000	09:12																																																																																																																																																	
Control Test	0.201	09:13																																																																																																																																																	
Air Blank	0.000	09:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2010																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:15																																																																																																																																																	
Control Test	0.078	09:15																																																																																																																																																	
Air Blank	0.000	09:16																																																																																																																																																	
Control Test	0.079	09:16																																																																																																																																																	
Air Blank	0.000	09:17																																																																																																																																																	
Control Test	0.079	09:17																																																																																																																																																	
Air Blank	0.000	09:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		

Comments: N/A



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001204, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001204	UNCERTAINTY* $\pm$
Owning Agency:	PANAMA CITY PD	0.050 g/ 210 L 0.004
Calibration Date:	05/31/2022	0.080 g/ 210 L 0.004
Calibration Time:	12:31	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

05/31/2022

Date

THOMAS J GRAHAM,
Department Inspector

Service • Integrity • Respect • Quality